



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124961

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/13/2023	15:50
2	CAL2	1	Water	NA	NA	20.7	7.00	04/13/2023	15:51
3	CAL3	1	Water	NA	NA	20.8	10.03	04/13/2023	15:53
4	ICV	1	Water	NA	NA	20.6	6.98	04/13/2023	16:00
5	CCV1	1	Water	NA	NA	20.7	2.01	04/13/2023	16:05
6	O2341-01	1	Solid	20.02	20	25.2	8.33	04/13/2023	16:10
7	O2341-01DUP	1	Solid	20.03	20	25.3	8.34	04/13/2023	16:11
8	O2341-02	1	Solid	20.04	20	25.2	8.44	04/13/2023	16:15
9	CCV2	1	Water	NA	NA	20.8	12.02	04/13/2023	16:20

2B2496)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O2341 WorkList ID : 169038 Department : Wet-Chemistry Date : 04-13-2023 15:32:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/13/2023	Solid	O2341-01	pH	Cool 4 deg C	CORN02	M11	PH1-BOT-1	04/13/2023	9045D
04/13/2023	Solid	O2341-02	pH	Cool 4 deg C	CORN02	M11	PH1-BOT-2	04/13/2023	9045D

Date/Time 04.13.23 15:40
Raw Sample Received by: JPC
Raw Sample Relinquished by: JPC

Date/Time 04.13.23 16:45
Raw Sample Received by: JPC
Raw Sample Relinquished by: JPC