

LB 124966


7470A

INSTRUMENT ID: CUL

Linear

 μ Abs.:

49271



A= 0.0000e+000

B= 2.0072e-004

C= -1.4899e-002

Rho= 0.9995268

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.004	0.004	93	0.000	93	0				0
0.2	0.200	0.200	-0.000	1069	0.0 %	1069					0
2.5	2.500	2.499	-0.001	12525	0.0 %	12525					0
5.0	5.000	4.897	-0.103	24472	0.0 %	24472					2
7.5	7.500	7.726	0.226	38565	0.0 %	38565					3
10.0	10.000	9.875	-0.125	49271	0.0 %	49271					1

LB124966 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Apr 2023 09:10:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	93	-	0.0000	7470A	PPB	14 Apr 2023 09:21:57	S	Std
0.2 - 1	50.2	1069	-	0.2000	7470A	PPB	14 Apr 2023 09:24:21	S	Std
2.5 - 1	52.5	12525	-	2.5000	7470A	PPB	14 Apr 2023 09:26:38	S	Std
5.0 - 1	55	24472	-	5.0000	7470A	PPB	14 Apr 2023 09:28:55	S	Std
7.5 - 1	57.5	38565	-	7.5000	7470A	PPB	14 Apr 2023 09:34:07	S	Std
10.0 - 1	51.0	49271	-	10.0000	7470A	PPB	14 Apr 2023 09:39:31	S	Std
ICV21 - 1	ICV21	20250	4.0496	-	7470A	PPB	14 Apr 2023 09:46:46	U	SMPL
ICB21 - 1	ICB21	-86	-0.0322	-	7470A	PPB	14 Apr 2023 09:49:01	U	SMPL
CCV65 - 1	CCV65	26355	5.2750	-	7470A	PPB	14 Apr 2023 09:51:16	U	SMPL
CCB65 - 1	CCB65	-105	-0.0360	-	7470A	PPB	14 Apr 2023 09:53:31	U	SMPL
CRA - 1	CRA	1054	0.1967	-	7470A	PPB	14 Apr 2023 09:55:48	U	SMPL
ChkStd - 1	ChkStd	33615	6.7322	-	7470A	PPB	14 Apr 2023 10:00:19	U	SMPL
PB152111BL - 1	PBW	-265	-0.0681	-	7470A	PPB	14 Apr 2023 10:02:35	U	SMPL
PB152111BS - 1	LCSW	19433	3.8856	-	7470A	PPB	14 Apr 2023 10:04:54	U	SMPL
HighStd - 1	HighStd	49181	9.8566	-	7470A	PPB	14 Apr 2023 10:07:25	U	SMPL
O2277-01 - 1	COMP-2	-80	-0.0310	-	7470A	PPB	14 Apr 2023 10:09:57	U	SMPL
O2277-02 - 1	COMP-1	-54	-0.0257	-	7470A	PPB	14 Apr 2023 10:12:14	U	SMPL
O2287-02 - 1	BS-1-2	130	0.0112	-	7470A	PPB	14 Apr 2023 10:14:31	U	SMPL
O2287-04 - 1	BS-3-4	78	0.0008	-	7470A	PPB	14 Apr 2023 10:16:48	U	SMPL
O2288-01 - 1	SOIL-CUTTINGS	85	0.0022	-	7470A	PPB	14 Apr 2023 10:19:04	U	SMPL
CCV66 - 1	CCV66	26635	5.3312	-	7470A	PPB	14 Apr 2023 10:21:20	U	SMPL
CCB66 - 1	CCB66	-84	-0.0318	-	7470A	PPB	14 Apr 2023 10:23:35	U	SMPL
O2292-02 - 1	IDW-SO-VB3-041123	24	-0.0101	-	7470A	PPB	14 Apr 2023 10:25:52	U	SMPL
O2292-04 - 1	IDW-SO-VB1-041123	68	-0.0012	-	7470A	PPB	14 Apr 2023 10:28:07	U	SMPL
O2292-06 - 1	IDW-SO-G1-041123	66	-0.0017	-	7470A	PPB	14 Apr 2023 10:30:23	U	SMPL
O2292-08 - 1	IDW-SO-G2-041123	70	-0.0008	-	7470A	PPB	14 Apr 2023 10:32:39	U	SMPL
O2292-10 - 1	IDW-SO-G3-041123	71	-0.0006	-	7470A	PPB	14 Apr 2023 10:34:55	U	SMPL
O2305-02 - 1	VNJ-232	188	0.0228	-	7470A	PPB	14 Apr 2023 10:37:12	U	SMPL
O2310-01 - 1	VNJ-228	-16	-0.0181	-	7470A	PPB	14 Apr 2023 10:39:30	U	SMPL
O2310-02 - 1	RBR-200059	81	0.0014	-	7470A	PPB	14 Apr 2023 10:41:47	U	SMPL
O2310-02DUP - 1	RBR-200059DUP	103	0.0058	-	7470A	PPB	14 Apr 2023 10:44:05	U	SMPL
O2310-02MS - 1	RBR-200059MS	18379	3.6741	-	7470A	PPB	14 Apr 2023 10:46:20	U	SMPL
CCV67 - 1	CCV67	26654	5.3350	-	7470A	PPB	14 Apr 2023 10:48:36	U	SMPL
CCB67 - 1	CCB67	-93	-0.0336	-	7470A	PPB	14 Apr 2023 10:50:52	U	SMPL
O2310-02MSD - 1	RBR-200059MSD	18167	3.6315	-	7470A	PPB	14 Apr 2023 10:53:09	U	SMPL
PB152112BL - 1	PBW	-54	-0.0257	-	7470A	PPB	14 Apr 2023 10:55:25	U	SMPL
PB152112BS - 1	LCSW	19577	3.9145	-	7470A	PPB	14 Apr 2023 11:00:18	U	SMPL
O2277-03 - 1	SOM-23-00002	63	-0.0023	-	7470A	PPB	14 Apr 2023 11:02:33	U	SMPL
O2277-03DUP - 1	SOM-23-00002DUP	133	0.0118	-	7470A	PPB	14 Apr 2023 11:04:49	U	SMPL
O2277-03MS - 1	SOM-23-00002MS	18655	3.7295	-	7470A	PPB	14 Apr 2023 11:07:06	U	SMPL
O2277-03MSD - 1	SOM-23-00002MSD	20482	4.0962	-	7470A	PPB	14 Apr 2023 11:09:23	U	SMPL
O2295-01 - 1	SEW-30994	2689	0.5248	-	7470A	PPB	14 Apr 2023 11:11:40	U	SMPL
CCV68 - 1	CCV68	25092	5.0215	-	7470A	PPB	14 Apr 2023 11:13:58	U	SMPL
CCB68 - 1	CCB68	-81	-0.0312	-	7470A	PPB	14 Apr 2023 11:16:14	U	SMPL
PB152069TB - 1	PB152069TB	79	0.0010	-	7470A	PPB	14 Apr 2023 11:18:30	U	SMPL
PB152110TB - 1	PB152110TB	72	-0.0004	-	7470A	PPB	14 Apr 2023 11:24:03	U	SMPL
O2310-02LX5 - 1		71	-0.0006	-	7470A	PPB	14 Apr 2023 11:26:19	U	SMPL
O2310-02A - 1		19619	3.9230	-	7470A	PPB	14 Apr 2023 11:28:34	U	SMPL
O2277-03LX5 - 1		-10	-0.0169	-	7470A	PPB	14 Apr 2023 11:30:50	U	SMPL
O2277-03A - 1		19239	3.8467	-	7470A	PPB	14 Apr 2023 11:33:06	U	SMPL
CCV69 - 1	CCV69	26303	5.2646	-	7470A	PPB	14 Apr 2023 11:35:23	U	SMPL
CCB69 - 1	CCB69	-63	-0.0275	-	7470A	PPB	14 Apr 2023 11:37:39	U	SMPL