

LB124979


SFAM01.1

INSTRUMENT ID: 1001

Linear

 μ Abs.:

54889



A= 0.0000e+000

B= 1.8253e-004 slope

C= 1.6698e-002 $y = mx + b$

Rho= 0.9995895

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p
0.00	0.000	0.049	0.049	176	0.000	176					7
0.05	0.050					438	0				11
0.20	0.200	0.222	0.022	1125	0.0 %	1125					2
2.50	2.500	2.542	0.042	13835	0.0 %	13835					5
5.00	5.000	4.767	-0.233	26028	0.0 %	26028					1
7.50	7.500	7.584	0.084	41461	0.0 %	41461					0
10.0	10.000	10.035	0.035	54889	0.0 %	54889					

LB124973

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	176	-		0 SFAM01.1	PPB	4/14/2023 14:59	Std
	0.2 S01	1125	-		0.2 SFAM01.1	PPB	4/14/2023 15:01	Std
	2.5 S02	13835	-		2.5 SFAM01.1	PPB	4/14/2023 15:03	Std
	5 S03	26028	-		5 SFAM01.1	PPB	4/14/2023 15:06	Std
	7.5 S04	41461	-		7.5 SFAM01.1	PPB	4/14/2023 15:11	Std
	10 S05	54889	-		10 SFAM01.1	PPB	4/14/2023 15:13	Std
ICV022	ICV022	20766	3.807 -		SFAM01.1	PPB	4/14/2023 15:16	SMPL
ICB022	ICB022	63	0.0282 -		SFAM01.1	PPB	4/14/2023 15:18	SMPL
CCV070	CCV070	25951	4.7534 -		SFAM01.1	PPB	4/14/2023 15:20	SMPL
CCB070	CCB070	56	0.0269 -		SFAM01.1	PPB	4/14/2023 15:22	SMPL
PB152136BL	PBS136	40	0.024 -		SFAM01.1	PPB	4/14/2023 15:25	SMPL
O2196-01	GBSW9	6292	1.1651 -		SFAM01.1	PPB	4/14/2023 15:27	SMPL
O2196-02	GBSW9D	5458	1.0129 -		SFAM01.1	PPB	4/14/2023 15:29	SMPL
O2196-03	GBSW9S	19139	3.5101 -		SFAM01.1	PPB	4/14/2023 15:31	SMPL
O2196-04	GBSX0	3963	0.74 -		SFAM01.1	PPB	4/14/2023 15:34	SMPL
O2196-05	GBSX1	4222	0.7873 -		SFAM01.1	PPB	4/14/2023 15:36	SMPL
O2196-06	GBSX2	7043	1.3022 -		SFAM01.1	PPB	4/14/2023 15:38	SMPL
O2196-07	GBSX3	1911	0.3655 -		SFAM01.1	PPB	4/14/2023 15:41	SMPL
O2196-08	GBSX4	626	0.131 -		SFAM01.1	PPB	4/14/2023 15:43	SMPL
O2196-09	GBSX5	430	0.0952 -		SFAM01.1	PPB	4/14/2023 15:45	SMPL
O2196-10	GBSX6	2640	0.4986 -		SFAM01.1	PPB	4/14/2023 15:47	SMPL
O2196-11	GBSX7	14835	2.7245 -		SFAM01.1	PPB	4/14/2023 15:50	SMPL
O2196-12	GBSX8	4776	0.8884 -		SFAM01.1	PPB	4/14/2023 15:52	SMPL
O2196-13	GBSX9	6632	1.2272 -		SFAM01.1	PPB	4/14/2023 15:54	SMPL
O2196-14	GBSY0	5252	0.9753 -		SFAM01.1	PPB	4/14/2023 15:56	SMPL
O2196-15	GBSY1	4393	0.8185 -		SFAM01.1	PPB	4/14/2023 15:59	SMPL
O2196-16	GBSY2	1653	0.3184 -		SFAM01.1	PPB	4/14/2023 16:01	SMPL
O2196-17	GBSY3	285	0.0687 -		SFAM01.1	PPB	4/14/2023 16:03	SMPL
O2196-18	GBSY4	444	0.0977 -		SFAM01.1	PPB	4/14/2023 16:05	SMPL
O2196-19	GBSY5	412	0.0919 -		SFAM01.1	PPB	4/14/2023 16:08	SMPL
CCV071	CCV071	25594	4.6883 -		SFAM01.1	PPB	4/14/2023 16:10	SMPL
CCB071	CCB071	29	0.022 -		SFAM01.1	PPB	4/14/2023 16:12	SMPL
O2196-20	GBSY6	1519	0.294 -		SFAM01.1	PPB	4/14/2023 16:15	SMPL
O2196-21	GBSY7	3351	0.6283 -		SFAM01.1	PPB	4/14/2023 16:17	SMPL
O2196-22	GBT08	6349	1.1756 -		SFAM01.1	PPB	4/14/2023 16:19	SMPL
O2196-23	GBT09	28756	5.2654 -		SFAM01.1	PPB	4/14/2023 16:21	SMPL
PB152137BL	PBS137	6	0.0178 -		SFAM01.1	PPB	4/14/2023 16:24	SMPL
O2197-01	GBSY8	3905	0.7295 -		SFAM01.1	PPB	4/14/2023 16:26	SMPL
O2197-02	GBSY9	4303	0.8021 -		SFAM01.1	PPB	4/14/2023 16:28	SMPL
O2197-03	GBSY9D	4712	0.8768 -		SFAM01.1	PPB	4/14/2023 16:30	SMPL
O2197-04	GBSY9S	19827	3.6356 -		SFAM01.1	PPB	4/14/2023 16:33	SMPL
O2197-05	GBSZ0	8024	1.4813 -		SFAM01.1	PPB	4/14/2023 16:35	SMPL
O2197-06	GBSZ1	4492	0.8366 -		SFAM01.1	PPB	4/14/2023 16:37	SMPL
O2197-07	GBSZ2	3168	0.5949 -		SFAM01.1	PPB	4/14/2023 16:40	SMPL
O2197-08	GBSZ3	3046	0.5727 -		SFAM01.1	PPB	4/14/2023 16:42	SMPL
O2197-09	GBSZ4	4508	0.8395 -		SFAM01.1	PPB	4/14/2023 16:44	SMPL

INSTRUMENT ID : CV1

O2197-10	GBSZ5	6607	1.2226 -	SFAM01.1	PPB	4/14/2023 16:46 SMPL
O2197-11	GBSZ6	702	0.1448 -	SFAM01.1	PPB	4/14/2023 16:49 SMPL
O2197-12	GBSZ7	1764	0.3387 -	SFAM01.1	PPB	4/14/2023 16:51 SMPL
O2197-13	GBSZ8	2596	0.4905 -	SFAM01.1	PPB	4/14/2023 16:53 SMPL
O2197-14	GBSZ9	3519	0.659 -	SFAM01.1	PPB	4/14/2023 16:55 SMPL
O2197-15	GBT00	2549	0.482 -	SFAM01.1	PPB	4/14/2023 16:58 SMPL
CCV072	CCV072	25668	4.7018 -	SFAM01.1	PPB	4/14/2023 17:00 SMPL
CCB072	CCB072	19	0.0202 -	SFAM01.1	PPB	4/14/2023 17:02 SMPL
O2197-16	GBT01	4471	0.8328 -	SFAM01.1	PPB	4/14/2023 17:05 SMPL
O2197-17	GBT02	5183	0.9627 -	SFAM01.1	PPB	4/14/2023 17:07 SMPL
O2197-18	GBT03	3618	0.6771 -	SFAM01.1	PPB	4/14/2023 17:09 SMPL
O2197-19	GBT04	3103	0.5831 -	SFAM01.1	PPB	4/14/2023 17:11 SMPL
O2197-20	GBT05	672	0.1394 -	SFAM01.1	PPB	4/14/2023 17:14 SMPL
O2197-21	GBT06	494	0.1069 -	SFAM01.1	PPB	4/14/2023 17:16 SMPL
O2197-22	GBT07	1527	0.2954 -	SFAM01.1	PPB	4/14/2023 17:18 SMPL
CCV073	CCV073	25697	4.7071 -	SFAM01.1	PPB	4/14/2023 17:28 SMPL
CCB073	CCB073	19	0.0202 -	SFAM01.1	PPB	4/14/2023 17:30 SMPL