



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124979

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	04/14/2023	08:40
2	O2295-01		1	14	100.00	>212.0	>212.0	04/14/2023	09:10
3	O2345-03		1	14	100.00	>212.0	>212.0	04/14/2023	09:40
4	O2345-03DUP		1	12	100.00	>212.0	>212.0	04/14/2023	10:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6124979

WorkList Name : fp-4-14

WorkList ID : 169054

Department : Wet-Chemistry

Date : 04-14-2023 07:57:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/19/2023	Water	Q2295-01E	Flash Point	Cool 4 deg C	PSEG03	M11	SEW-30994	04/12/2023	1010B
04/20/2023	Water	Q2345-03	Flash Point	Cool 4 deg C	PSEG03	M11	1599	04/13/2023	1010B

Date/Time 04/14/2023 08:30
 Raw Sample Received by: RMA CWSG
 Raw Sample Relinquished by: 786029

Date/Time 04/14/2023 10:45
 Raw Sample Received by: 786029
 Raw Sample Relinquished by: RMA CWSG