



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124982

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/16/2023	07:00
2	CAL2	1	Water	NA	NA	20.7	7.01	04/16/2023	07:01
3	CAL3	1	Water	NA	NA	20.1	10.03	04/16/2023	07:22
4	ICV	1	Water	NA	NA	20.7	6.98	04/16/2023	07:25
5	CCV1	1	Water	NA	NA	20.7	2.01	04/16/2023	07:30
6	O2345-03	1	Water	NA	NA	22.3	8.50	04/16/2023	07:32
7	O2353-01	1	Water	NA	NA	22.1	6.04	04/16/2023	07:38
8	O2353-02	1	Water	NA	NA	21.2	8.64	04/16/2023	07:42
9	O2353-03	1	Water	NA	NA	22.2	5.39	04/16/2023	07:48
10	O2353-04	1	Water	NA	NA	22.1	8.82	04/16/2023	07:45
11	O2353-05	1	Water	NA	NA	22.1	7.74	04/16/2023	07:50
12	O2353-05DUP	1	Water	NA	NA	22.2	7.75	04/16/2023	07:51
13	CCV2	1	Water	NA	NA	20.7	12.02	04/16/2023	08:00

VB1224982

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 02353

WorkList ID : 169074

Department : Wet-Chemistry

Date : 04-16-2023 06:35:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/13/2023	Water	O2345-03	pH C	Cool 4 deg C	PSEG03	M11	1599	04/13/2023	9040C
04/14/2023	Water	O2353-01	pH C	Cool 4 deg C	PSEG03	M11	BUR-1253	04/14/2023	9040C
04/14/2023	Water	O2353-02	pH C	Cool 4 deg C	PSEG03	M11	BUR-1260	04/14/2023	9040C
04/14/2023	Water	O2353-03	pH C	Cool 4 deg C	PSEG03	M11	BUR-1261	04/14/2023	9040C
04/14/2023	Water	O2353-04	pH C	Cool 4 deg C	PSEG03	M11	BUR-1262	04/14/2023	9040C
04/14/2023	Water	O2353-05	pH C	Cool 4 deg C	PSEG03	M11	BUR-1263	04/14/2023	9040C

Date/Time 04-16-23 6:40

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 04-16-23 10:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]