



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB124983

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/16/2023	07:00
2	CAL2	1	Water	NA	NA	20.7	7.01	04/16/2023	07:01
3	CAL3	1	Water	NA	NA	20.7	10.03	04/16/2023	07:03
4	ICV	1	Water	NA	NA	20.7	6.98	04/16/2023	07:05
5	CCV1	1	Water	NA	NA	20.7	2.02	04/16/2023	08:30
6	O2365-04	1	Water	NA	NA	22.1	8.07	04/16/2023	08:35
7	O2365-05	1	Water	NA	NA	22.1	6.97	04/16/2023	08:40
8	O2365-05DUP	1	Water	NA	NA	22.2	6.98	04/16/2023	08:45
9	CCV2	1	Water	NA	NA	20.7	12.02	04/16/2023	08:50

WB 124983

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 02365 WorkList ID : 169083 Department : Wet-Chemistry Date : 04-16-2023 08:20:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/14/2023	Water	O2365-04	Corrosivity	Cool 4 deg C	PSEG03	M11	ELA-23-00004	04/14/2023	9040C
04/14/2023	Water	O2365-05	Corrosivity	Cool 4 deg C	PSEG03	M11	ELA-23-00012	04/14/2023	9040C

Date/Time 04.16.23 08:25
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: JD (WC)

Date/Time 04.16.23 11:00
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: JD (WC)