

LB124984


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

55642



A= 0.0000e+000

B= 1.8041e-004 slope

C= -3.1437e-004 $\rightarrow$ intercept

Rho= 0.9999228

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.057	0.057	319	0.000	319					1
0.05	0.050					438	0				7
0.20	0.200	0.214	0.014	1189	0.0 %	1189					-3
2.50	2.500	2.416	-0.084	13391	0.0 %	13391					0
5.00	5.000	4.999	-0.001	27710	0.0 %	27710					0
7.50	7.500	7.476	-0.024	41443	0.0 %	41443					0
10.0	10.000	10.038	0.038	55642	0.0 %	55642					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	319	-		0 SFAM01.1	PPB	4/16/2023 12:40	Std
	0.2 S01	1189	-		0.2 SFAM01.1	PPB	4/16/2023 12:42	Std
	2.5 S02	13391	-		2.5 SFAM01.1	PPB	4/16/2023 12:44	Std
	5 S03	27710	-		5 SFAM01.1	PPB	4/16/2023 12:47	Std
	7.5 S04	41443	-		7.5 SFAM01.1	PPB	4/16/2023 12:49	Std
	10 S05	55642	-		10 SFAM01.1	PPB	4/16/2023 12:51	Std
ICV023	ICV023	22495	4.058 -		SFAM01.1	PPB	4/16/2023 12:57	SMPL
ICB023	ICB023	235	0.0421 -		SFAM01.1	PPB	4/16/2023 12:59	SMPL
CCV074	CCV074	27705	4.9979 -		SFAM01.1	PPB	4/16/2023 13:01	SMPL
CCB074	CCB074	218	0.039 -		SFAM01.1	PPB	4/16/2023 13:03	SMPL
PB152138BL	PBS138	12	0.0019 -		SFAM01.1	PPB	4/16/2023 13:06	SMPL
O2198-01	GBSC7	736	0.1325 -		SFAM01.1	PPB	4/16/2023 13:12	SMPL
O2198-02	GBSC7D	629	0.1132 -		SFAM01.1	PPB	4/16/2023 13:15	SMPL
O2198-03	GBSC7S	12424	2.2411 -		SFAM01.1	PPB	4/16/2023 13:17	SMPL
O2198-04	GBSC8	651	0.1171 -		SFAM01.1	PPB	4/16/2023 13:19	SMPL
O2198-05	GBSC9	431	0.0774 -		SFAM01.1	PPB	4/16/2023 13:22	SMPL
O2198-06	GBSD0	365	0.0655 -		SFAM01.1	PPB	4/16/2023 13:24	SMPL
O2198-07	GBSD1	401	0.072 -		SFAM01.1	PPB	4/16/2023 13:26	SMPL
O2198-08	GBSD2	598	0.1076 -		SFAM01.1	PPB	4/16/2023 13:28	SMPL
O2198-09	GBSD3	323	0.058 -		SFAM01.1	PPB	4/16/2023 13:31	SMPL
O2198-10	GBSD4	520	0.0935 -		SFAM01.1	PPB	4/16/2023 13:33	SMPL
O2198-11	GBSD5	365	0.0655 -		SFAM01.1	PPB	4/16/2023 13:35	SMPL
O2198-12	GBSD6	314	0.0563 -		SFAM01.1	PPB	4/16/2023 13:37	SMPL
O2198-13	GBSD7	1712	0.3085 -		SFAM01.1	PPB	4/16/2023 13:40	SMPL
O2198-14	GBSD8	593	0.1067 -		SFAM01.1	PPB	4/16/2023 13:42	SMPL
O2198-15	GBSD9	412	0.074 -		SFAM01.1	PPB	4/16/2023 13:44	SMPL
O2198-16	GBSE0	413	0.0742 -		SFAM01.1	PPB	4/16/2023 13:46	SMPL
O2198-17	GBSE1	380	0.0682 -		SFAM01.1	PPB	4/16/2023 13:49	SMPL
O2198-18	GBSE2	1571	0.2831 -		SFAM01.1	PPB	4/16/2023 13:51	SMPL
O2198-19	GBSE3	3049	0.5497 -		SFAM01.1	PPB	4/16/2023 13:53	SMPL
CCV075	CCV075	27198	4.9064 -		SFAM01.1	PPB	4/16/2023 13:56	SMPL
CCB075	CCB075	269	0.0482 -		SFAM01.1	PPB	4/16/2023 13:58	SMPL
O2198-20	GBSE4	6647	1.1989 -		SFAM01.1	PPB	4/16/2023 14:00	SMPL
O2198-21	GBSE5	43427	7.8343 -		SFAM01.1	PPB	4/16/2023 14:02	SMPL
O2198-22	GBSL3	257	0.0461 -		SFAM01.1	PPB	4/16/2023 14:05	SMPL
O2198-23	GBSL7	26310	4.7462 -		SFAM01.1	PPB	4/16/2023 14:07	SMPL
PB152139BL	PBS139	-41	-0.0077 -		SFAM01.1	PPB	4/16/2023 14:09	SMPL
O2199-01	GBSE6	12405	2.2376 -		SFAM01.1	PPB	4/16/2023 14:12	SMPL
O2199-02	GBSE7	11011	1.9862 -		SFAM01.1	PPB	4/16/2023 14:14	SMPL
O2199-03	GBSE7D	10461	1.8869 -		SFAM01.1	PPB	4/16/2023 14:16	SMPL
O2199-04	GBSE7S	22788	4.1108 -		SFAM01.1	PPB	4/16/2023 14:18	SMPL
O2199-05	GBSE8	12317	2.2218 -		SFAM01.1	PPB	4/16/2023 14:21	SMPL
O2199-06	GBSE9	9006	1.6244 -		SFAM01.1	PPB	4/16/2023 14:23	SMPL
O2199-07	GBSF0	2526	0.4554 -		SFAM01.1	PPB	4/16/2023 14:25	SMPL
O2199-08	GBSF1	4029	0.7265 -		SFAM01.1	PPB	4/16/2023 14:27	SMPL
O2199-09	GBSF2	7459	1.3453 -		SFAM01.1	PPB	4/16/2023 14:30	SMPL

LB124984
INSTRUMENT ID : CV1

O2199-10	GBSF3	14253	2.571 -	SFAM01.1	PPB	4/16/2023 14:32	SMPL
O2199-11	GBSF4	3595	0.6483 -	SFAM01.1	PPB	4/16/2023 14:34	SMPL
O2199-12	GBSF5	8153	1.4706 -	SFAM01.1	PPB	4/16/2023 14:36	SMPL
O2199-13	GBSF6	10030	1.8092 -	SFAM01.1	PPB	4/16/2023 14:39	SMPL
O2199-14	GBSF7	3737	0.6739 -	SFAM01.1	PPB	4/16/2023 14:41	SMPL
O2199-15	GBSF8	2252	0.406 -	SFAM01.1	PPB	4/16/2023 14:43	SMPL
CCV076	CCV076	26998	4.8703 -	SFAM01.1	PPB	4/16/2023 14:46	SMPL
CCB076	CCB076	224	0.0401 -	SFAM01.1	PPB	4/16/2023 14:48	SMPL
O2199-16	GBSF9	5428	0.9789 -	SFAM01.1	PPB	4/16/2023 14:50	SMPL
O2199-17	GBSG0	4860	0.8765 -	SFAM01.1	PPB	4/16/2023 14:52	SMPL
O2199-18	GBSG1	19648	3.5443 -	SFAM01.1	PPB	4/16/2023 14:55	SMPL
O2199-19	GBSG2	10028	1.8088 -	SFAM01.1	PPB	4/16/2023 14:57	SMPL
O2199-20	GBSG3	5442	0.9815 -	SFAM01.1	PPB	4/16/2023 14:59	SMPL
O2199-21	GBSG4	5291	0.9542 -	SFAM01.1	PPB	4/16/2023 15:01	SMPL
O2199-22	GBSL4	929	0.1673 -	SFAM01.1	PPB	4/16/2023 15:04	SMPL
PB152140BL	PBS140	206	0.0368 -	SFAM01.1	PPB	4/16/2023 15:06	SMPL
O2200-01	GBSG5	736	0.1325 -	SFAM01.1	PPB	4/16/2023 15:08	SMPL
O2200-02	GBSG6	2444	0.4406 -	SFAM01.1	PPB	4/16/2023 15:11	SMPL
O2200-03	GBSG7	855	0.1539 -	SFAM01.1	PPB	4/16/2023 15:13	SMPL
O2200-04	GBSG7D	797	0.1435 -	SFAM01.1	PPB	4/16/2023 15:15	SMPL
O2200-05	GBSG7S	13775	2.4848 -	SFAM01.1	PPB	4/16/2023 15:20	SMPL
O2200-06	GBSG8	722	0.1299 -	SFAM01.1	PPB	4/16/2023 15:22	SMPL
O2200-07	GBSG9	660	0.1188 -	SFAM01.1	PPB	4/16/2023 15:24	SMPL
O2200-08	GBSH0	548	0.0985 -	SFAM01.1	PPB	4/16/2023 15:27	SMPL
O2200-09	GBSH1	904	0.1628 -	SFAM01.1	PPB	4/16/2023 15:29	SMPL
O2200-10	GBSH2	681	0.1225 -	SFAM01.1	PPB	4/16/2023 15:31	SMPL
O2200-11	GBSH3	2972	0.5359 -	SFAM01.1	PPB	4/16/2023 15:33	SMPL
O2200-12	GBSH4	616	0.1108 -	SFAM01.1	PPB	4/16/2023 15:36	SMPL
CCV077	CCV077	27192	4.9053 -	SFAM01.1	PPB	4/16/2023 15:38	SMPL
CCB077	CCB077	251	0.045 -	SFAM01.1	PPB	4/16/2023 15:40	SMPL
O2200-13	GBSH5	610	0.1097 -	SFAM01.1	PPB	4/16/2023 15:43	SMPL
O2200-14	GBSH6	557	0.1002 -	SFAM01.1	PPB	4/16/2023 15:45	SMPL
O2200-15	GBSH7	569	0.1023 -	SFAM01.1	PPB	4/16/2023 15:47	SMPL
O2200-16	GBSH8	1953	0.352 -	SFAM01.1	PPB	4/16/2023 15:49	SMPL
O2200-17	GBSH9	2356	0.4247 -	SFAM01.1	PPB	4/16/2023 15:52	SMPL
O2200-18	GBSJ0	4523	0.8157 -	SFAM01.1	PPB	4/16/2023 15:54	SMPL
O2200-19	GBSJ1	12099	2.1824 -	SFAM01.1	PPB	4/16/2023 15:56	SMPL
O2200-20	GBSJ2	5575	1.0055 -	SFAM01.1	PPB	4/16/2023 15:58	SMPL
O2200-21	GBSJ3	723	0.1301 -	SFAM01.1	PPB	4/16/2023 16:01	SMPL
O2200-22	GBSL5	761	0.137 -	SFAM01.1	PPB	4/16/2023 16:03	SMPL
PB152141BL	PBS141	240	0.043 -	SFAM01.1	PPB	4/16/2023 16:05	SMPL
O2201-01	GBSJ4	5146	0.9281 -	SFAM01.1	PPB	4/16/2023 16:08	SMPL
O2201-02	GBSJ5	10360	1.8687 -	SFAM01.1	PPB	4/16/2023 16:10	SMPL
O2201-03	GBSJ6	12381	2.2333 -	SFAM01.1	PPB	4/16/2023 16:12	SMPL
O2201-04	GBSJ7	1441	0.2597 -	SFAM01.1	PPB	4/16/2023 16:20	SMPL
O2201-05	GBSJ7D	1357	0.2445 -	SFAM01.1	PPB	4/16/2023 16:22	SMPL
O2201-06	GBSJ7S	14581	2.6302 -	SFAM01.1	PPB	4/16/2023 16:24	SMPL

LB124984
INSTRUMENT ID : CV1

O2201-07	GBSJ8	675	0.1215 -	SFAM01.1	PPB	4/16/2023 16:26 SMPL
O2201-08	GBSJ9	593	0.1067 -	SFAM01.1	PPB	4/16/2023 16:29 SMPL
O2201-09	GBSK0	3889	0.7013 -	SFAM01.1	PPB	4/16/2023 16:31 SMPL
CCV078	CCV078	27372	4.9378 -	SFAM01.1	PPB	4/16/2023 16:33 SMPL
CCB078	CCB078	239	0.0428 -	SFAM01.1	PPB	4/16/2023 16:36 SMPL
O2201-10	GBSK1	1390	0.2505 -	SFAM01.1	PPB	4/16/2023 16:38 SMPL
O2201-11	GBSK2	580	0.1043 -	SFAM01.1	PPB	4/16/2023 16:40 SMPL
O2201-12	GBSK3	3694	0.6661 -	SFAM01.1	PPB	4/16/2023 16:42 SMPL
O2201-13	GBSK4	3580	0.6455 -	SFAM01.1	PPB	4/16/2023 16:45 SMPL
O2201-14	GBSK5	3789	0.6833 -	SFAM01.1	PPB	4/16/2023 16:47 SMPL
O2201-15	GBSK6	2601	0.4689 -	SFAM01.1	PPB	4/16/2023 16:49 SMPL
O2201-16	GBSK7	850	0.153 -	SFAM01.1	PPB	4/16/2023 16:52 SMPL
O2201-17	GBSK8	671	0.1207 -	SFAM01.1	PPB	4/16/2023 16:54 SMPL
O2201-18	GBSK9	415	0.0746 -	SFAM01.1	PPB	4/16/2023 16:56 SMPL
O2201-19	GBSL0	2619	0.4722 -	SFAM01.1	PPB	4/16/2023 16:58 SMPL
O2201-20	GBSL1	1566	0.2822 -	SFAM01.1	PPB	4/16/2023 17:01 SMPL
O2201-21	GBSL2	526	0.0946 -	SFAM01.1	PPB	4/16/2023 17:03 SMPL
O2201-22	GBSL6	697	0.1254 -	SFAM01.1	PPB	4/16/2023 17:05 SMPL
CCV079	CCV079	27395	4.942 -	SFAM01.1	PPB	4/16/2023 17:14 SMPL
CCB079	CCB079	230	0.0412 -	SFAM01.1	PPB	4/16/2023 17:17 SMPL