


LB124985

7471B

Linear

μ Abs.:

53245



A= 0.0000e+000

B= 1.8368e-004

C= -8.3738e-003

Rho= 0.9990417

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0D
0.0	0.000	0.007	0.007	83	0.000	83					
0.2	0.200	0.207	0.007	1171	0.0 %	1171					
2.5	2.500	2.352	-0.148	12852	0.0 %	12852					
5.0	5.000	5.090	0.090	27754	0.0 %	27754					
7.5	7.500	7.773	0.273	42363	0.0 %	42363					
10.0	10.000	9.772	-0.228	53245	0.0 %	53245					

LB124985 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Apr 2023 07:47:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	83	-	0.00007471B	PPB		17 Apr 2023 08:00:50	S	Std
0.2 - 1	50.2	1171	-	0.20007471B	PPB		17 Apr 2023 08:03:07	S	Std
2.5 - 1	52.5	12852	-	2.50007471B	PPB		17 Apr 2023 08:05:25	S	Std
5.0 - 1	55	27754	-	5.00007471B	PPB		17 Apr 2023 08:10:15	S	Std
7.5 - 1	57.5	42363	-	7.50007471B	PPB		17 Apr 2023 08:15:23	S	Std
10.0 - 1	510	53245	-	10.00007471B	PPB		17 Apr 2023 08:20:20	S	Std
ICV24 - 1	ICV24	23869	4.3759	-7471B	PPB		17 Apr 2023 08:23:13	U	SMPL
ICB24 - 1	ICB24	-152	-0.0363	-7471B	PPB		17 Apr 2023 08:25:28	U	SMPL
CCV80 - 1	CCV80	28523	5.2308	-7471B	PPB		17 Apr 2023 08:27:45	U	SMPL
CCB80 - 1	CCB80	-121	-0.0306	-7471B	PPB		17 Apr 2023 08:30:00	U	SMPL
CRA - 1	CRA	1084	0.1907	-7471B	PPB		17 Apr 2023 08:32:18	U	SMPL
HighStd - 1	HighStd	54212	9.9493	-7471B	PPB		17 Apr 2023 08:34:33	U	SMPL
ChkStd - 1	ChkStd	39116	7.1765	-7471B	PPB		17 Apr 2023 08:36:48	U	SMPL
PB152127BL - 1	PBS	-295	-0.0626	-7471B	PPB		17 Apr 2023 08:39:06	U	SMPL
PB152127BS - 1	LCSS	23064	4.2280	-7471B	PPB		17 Apr 2023 08:41:24	U	SMPL
O2301-02 - 1	WASTE	5948	1.0842	-7471B	PPB		17 Apr 2023 08:43:41	U	SMPL
O2341-01 - 1	PH1-BOT-1	905	0.1579	-7471B	PPB		17 Apr 2023 08:45:59	U	SMPL
O2341-01DUP - 1	PH1-BOT-1DUP	761	0.1314	-7471B	PPB		17 Apr 2023 08:48:18	U	SMPL
O2341-01MS - 1	PH1-BOT-1MS	21546	3.9492	-7471B	PPB		17 Apr 2023 08:54:04	U	SMPL
O2341-01MSD - 1	PH1-BOT-1MSD	20599	3.7753	-7471B	PPB		17 Apr 2023 08:56:19	U	SMPL
CCV81 - 1	CCV81	28594	5.2438	-7471B	PPB		17 Apr 2023 08:58:36	U	SMPL
CCB81 - 1	CCB81	-100	-0.0267	-7471B	PPB		17 Apr 2023 09:00:53	U	SMPL
O2341-02 - 1	PH1-BOT-2	658	0.1125	-7471B	PPB		17 Apr 2023 09:03:10	U	SMPL
O2345-01 - 1	1595	928	0.1621	-7471B	PPB		17 Apr 2023 09:05:25	U	SMPL
O2345-02 - 1	1611	398	0.0647	-7471B	PPB		17 Apr 2023 09:07:42	U	SMPL
O2346-01 - 1	31925	257	0.0388	-7471B	PPB		17 Apr 2023 09:09:58	U	SMPL
O2346-02 - 1	31501	15113	2.7676	-7471B	PPB		17 Apr 2023 09:12:14	U	SMPL
O2346-03 - 1	32123	3706	0.6723	-7471B	PPB		17 Apr 2023 09:14:30	U	SMPL
O2347-05 - 1	20230331-COMP-9	787	0.1362	-7471B	PPB		17 Apr 2023 09:16:48	U	SMPL
O2347-06 - 1	20230332-COMP-10	1190	0.2102	-7471B	PPB		17 Apr 2023 09:19:05	U	SMPL
O2347-07 - 1	20230333-COMP-11	3951	0.7173	-7471B	PPB		17 Apr 2023 09:21:23	U	SMPL
O2347-08 - 1	20230337-COMP-15	31631	5.8016	-7471B	PPB		17 Apr 2023 09:23:39	U	SMPL
CCV82 - 1	CCV82	28692	5.2618	-7471B	PPB		17 Apr 2023 09:25:55	U	SMPL
CCB82 - 1	CCB82	-61	-0.0196	-7471B	PPB		17 Apr 2023 09:28:12	U	SMPL
O2347-09 - 1	20230338-COMP-16	38029	6.9768	-7471B	PPB		17 Apr 2023 09:30:30	U	SMPL
O2349-01 - 1	NB-07-041323	5232	0.9526	-7471B	PPB		17 Apr 2023 09:32:45	U	SMPL
O2341-01LX5 - 1		156	0.0203	-7471B	PPB		17 Apr 2023 09:40:55	U	SMPL
O2341-01A - 1		23772	4.3581	-7471B	PPB		17 Apr 2023 09:43:10	U	SMPL
CCV83 - 1	CCV83	28820	5.2853	-7471B	PPB		17 Apr 2023 09:48:03	U	SMPL
CCB83 - 1	CCB83	-94	-0.0256	-7471B	PPB		17 Apr 2023 09:50:19	U	SMPL