



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 4/18/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 17:00

In Date: 04/17/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:25

Out Date: 04/18/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB124988

| Lab ID   | Client SampleID   | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| O2371-01 | SB-26-9.0-9.5     | 1      | 1.14           | 8.59         | 9.73                    | 8.99                      | 91.4    |          |
| O2371-02 | SB-26-COMP        | 2      | 1.12           | 8.56         | 9.68                    | 9.03                      | 92.4    |          |
| O2371-03 | SB-23-5.0-5.5     | 3      | 1.12           | 8.54         | 9.66                    | 9.17                      | 94.3    |          |
| O2371-04 | SB-23-COMP        | 4      | 1.16           | 8.65         | 9.81                    | 9.28                      | 93.9    |          |
| O2371-05 | SB-22-8.5-9.0     | 5      | 1.13           | 8.56         | 9.69                    | 8.57                      | 86.9    |          |
| O2371-06 | SB-22-COMP        | 6      | 1.13           | 8.50         | 9.63                    | 9.21                      | 95.1    |          |
| O2371-07 | SB-21-8.5-9.0     | 7      | 1.19           | 8.53         | 9.72                    | 8.63                      | 87.2    |          |
| O2371-08 | SB-21-COMP        | 8      | 1.15           | 8.59         | 9.74                    | 9.12                      | 92.8    |          |
| O2371-09 | SB-20-8.0-8.5     | 9      | 1.1            | 8.75         | 9.85                    | 8.53                      | 84.9    |          |
| O2371-10 | SB-20-COMP        | 10     | 1.19           | 8.69         | 9.88                    | 9.21                      | 92.3    |          |
| O2371-11 | SB-19-8.0-8.5     | 11     | 1.19           | 8.42         | 9.61                    | 8.85                      | 91.0    |          |
| O2371-12 | SB-19-COMP        | 12     | 1.17           | 8.47         | 9.64                    | 8.77                      | 89.7    |          |
| O2371-13 | SB-18-9.5-10.0    | 13     | 1.19           | 8.52         | 9.71                    | 9.33                      | 95.5    |          |
| O2371-14 | SB-18-COMP        | 14     | 1.18           | 8.56         | 9.74                    | 9.35                      | 95.4    |          |
| O2371-15 | SB-17-8.0-8.5     | 15     | 1.15           | 8.80         | 9.95                    | 9.00                      | 89.2    |          |
| O2371-16 | SB-17-COMP        | 16     | 1.19           | 8.79         | 9.98                    | 9.33                      | 92.6    |          |
| O2371-17 | SB-16-9.0-9.5     | 17     | 1.17           | 8.44         | 9.61                    | 9.16                      | 94.7    |          |
| O2371-18 | SB-16-COMP        | 18     | 1.13           | 8.50         | 9.63                    | 9.00                      | 92.6    |          |
| O2384-01 | ROLL-OFF-8PT-COMP | 19     | 1.14           | 8.53         | 9.67                    | 8.78                      | 89.6    |          |
| O2384-03 | ROLL-OFF-5PT-COMP | 20     | 1.15           | 8.50         | 9.65                    | 8.42                      | 85.5    |          |
| O2385-01 | BU-3-041723       | 21     | 1.19           | 8.36         | 9.55                    | 8.2                       | 83.9    |          |
| O2388-01 | CC-1              | 22     | 1.11           | 8.51         | 9.62                    | 8.21                      | 83.4    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$