

LB125002

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

4/17/2023 12:02

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.988	0.0	0.101	
ICB1	-0.001	0.0	0.010	
CCV1	0.991	0.0	0.102	
CCB1	-0.001	0.0	0.010	
PB152159BL	0.003	0.0	0.010	
PB152159BS	0.986	0.0	0.101	
O2213-10	0.029	0.0	0.013	
O2213-11	0.056	0.0	0.015	
CCV2	0.992	0.0	0.102	
CCB2	-0.002	0.0	0.010	

N	10
Mean	0.404
SD	0.5037
CV%	124.63

Aquakem v. 7.2AQ1

Results from time period:

Mon Apr 17 10:31:07 2023

Mon Apr 17 12:02:30 2023

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0.0008	mg/l	4/17/2023 10:52:21	
0.05PPM	A	Phenolics-	P	0.0524	mg/l	4/17/2023 10:52:22	
0.1PPM	A	Phenolics-	P	0.1081	mg/l	4/17/2023 10:52:23	
0.25PPM	A	Phenolics-	P	0.2527	mg/l	4/17/2023 10:52:24	
0.50PPM	A	Phenolics-	P	0.4932	mg/l	4/17/2023 10:52:25	
1.0PPM	A	Phenolics-	P	0.9838	mg/l	4/17/2023 10:52:26	
2.0PPM	A	Phenolics-	P	2.009	mg/l	4/17/2023 10:52:27	
ICV1	S	Phenolics-	P	0.9876	mg/l	4/17/2023 11:55:38	
ICB1	S	Phenolics-	P	-0.0012	mg/l	4/17/2023 11:55:40	
CCV1	S	Phenolics-	P	0.991	mg/l	4/17/2023 11:55:42	
CCB1	S	Phenolics-	P	-0.0005	mg/l	4/17/2023 11:55:44	
PB152159BL	S	Phenolics-	P	0.0033	mg/l	4/17/2023 11:55:45	
PB152159BS	S	Phenolics-	P	0.9862	mg/l	4/17/2023 11:55:48	
O2213-10	S	Phenolics-	P	0.0291	mg/l	4/17/2023 12:02:23	
O2213-11	S	Phenolics-	P	0.0565	mg/l	4/17/2023 12:02:25	
CCV2	S	Phenolics-	P	0.9915	mg/l	4/17/2023 12:02:27	
CCB2	S	Phenolics-	P	-0.0018	mg/l	4/17/2023 12:02:30	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : _____ Instrument ID : Konelab

4/17/2023 10:52


Test Phenolics-

Accepted 4/17/2023 10:52

Factor 10.8
Bias 0.010

Coeff. of det. 0.999851

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.010	0.0008	0.0000	-
2	Phenol-2	0.015	0.0524	0.0500	4.8
3	Phenol-2	0.020	0.1081	0.1000	8.1
4	Phenol-2	0.033	0.2527	0.2500	1.1
5	Phenol-2	0.056	0.4932	0.5000	-1.4
6	Phenol-2	0.101	0.9838	1.0000	-1.6
7	Phenol-2	0.196	2.0090	2.0000	0.9

12
4/17/23