



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125004

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/17/2023	14:20
2	CAL2	1	Water	NA	NA	20.7	7.00	04/17/2023	14:21
3	CAL3	1	Water	NA	NA	20.7	10.02	04/17/2023	14:22
4	ICV	1	Water	NA	NA	20.8	6.99	04/17/2023	14:25
5	CCV1	1	Water	NA	NA	20.7	2.01	04/17/2023	14:30
6	O2379-01	1	Water	NA	NA	20.9	5.08	04/17/2023	14:33
7	O2379-01DUP	1	Water	NA	NA	20.8	5.09	04/17/2023	14:34
8	CCV2	1	Water	NA	NA	20.7	12.02	04/17/2023	14:35

VB 125004

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 02379 WorkList ID : 169113 Department : Wet-Chemistry Date : 04-17-2023 14:03:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/14/2023	Water	02379-01	pH		Cool 4 deg C	STER10	M11	CELL-A-POOL-WATER(DE-IONIZED)	04/14/2023 9040C

Date/Time 04-14-23 14:15
Raw Sample Received by: JR (QC)
Raw Sample Relinquished by: JD (CSM)

Date/Time 04-14-23 16:30
Raw Sample Received by: JD (CSM)
Raw Sample Relinquished by: JR (QC)