



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125005

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/17/2023	14:20
2	CAL2	1	Water	NA	NA	20.7	7.00	04/17/2023	14:21
3	CAL3	1	Water	NA	NA	20.7	10.02	04/17/2023	14:22
4	ICV	1	Water	NA	NA	20.8	6.99	04/17/2023	14:25
5	CCV1	1	Water	NA	NA	20.8	2.02	04/17/2023	14:45
6	O2385-01	1	Solid	20.02	20	21.4	7.39	04/17/2023	14:50
7	O2385-01DUP	1	Solid	20.03	20	21.5	7.40	04/17/2023	14:51
8	CCV2	1	Water	NA	NA	20.7	12.02	04/17/2023	15:00

✓ 125005

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 02385 S

WorkList ID : 169114

Department : Wet-Chemistry

Date : 04-17-2023 14:04:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/17/2023	Solid	02385-01	pH		Cool 4 deg C	PSEG05 M11	BU-3-041723	04/17/2023	9045D

Date/Time 04-17-23 14:10
Raw Sample Received by: 1P (WC)
Raw Sample Relinquished by: 50 (SM)

Date/Time 04-17-23 16:00
Raw Sample Received by: 50 (SM)
Raw Sample Relinquished by: 1P (WC)