

LB125008


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56177



A= 0.0000e+000

B= 1.8305e-004 slope

C= -6.3985e-002 intercept

Rho= 0.9990443

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	YOD
0.00	0.000	0.041	0.041	575	0.000	575					
0.05	0.050					438	0				
0.20	0.200	0.217	0.017	1537	0.0 %	1537					
2.50	2.500	2.567	0.067	14372	0.0 %	14372					
5.00	5.000	4.973	-0.027	27520	0.0 %	27520					
7.50	7.500	7.182	-0.318	39587	0.0 %	39587					
10.0	10.000	10.219	0.219	56177	0.0 %	56177					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	575	-		0 SFAM01.1	PPB	4/17/2023 15:57	Std
	0.2 S1	1537	-		0.2 SFAM01.1	PPB	4/17/2023 16:01	Std
	2.5 S02	14372	-		2.5 SFAM01.1	PPB	4/17/2023 16:04	Std
	5 S03	27520	-		5 SFAM01.1	PPB	4/17/2023 16:06	Std
	7.5 S04	39587	-		7.5 SFAM01.1	PPB	4/17/2023 16:11	Std
	10 S05	56177	-		10 SFAM01.1	PPB	4/17/2023 16:15	Std
ICV027	ICV027	22597	4.0723 -		SFAM01.1	PPB	4/17/2023 16:18	SMPL
ICB027	ICB027	461	0.0204 -		SFAM01.1	PPB	4/17/2023 16:20	SMPL
CCV092	CCV092	27657	4.9985 -		SFAM01.1	PPB	4/17/2023 16:22	SMPL
CCB092	CCB092	442	0.0169 -		SFAM01.1	PPB	4/17/2023 16:24	SMPL
PB152170BL	PBS170	57	-0.0536 -		SFAM01.1	PPB	4/17/2023 16:27	SMPL
O2204-01	GBSQ6	678	0.0601 -		SFAM01.1	PPB	4/17/2023 16:29	SMPL
O2204-02	GBSQ7	1987	0.2997 -		SFAM01.1	PPB	4/17/2023 16:31	SMPL
O2204-03	GBSQ8	756	0.0744 -		SFAM01.1	PPB	4/17/2023 16:33	SMPL
O2204-04	GBSQ8D	615	0.0486 -		SFAM01.1	PPB	4/17/2023 16:36	SMPL
O2204-05	GBSQ8S	15443	2.7628 -		SFAM01.1	PPB	4/17/2023 16:38	SMPL
O2204-06	GBSQ9	450	0.0184 -		SFAM01.1	PPB	4/17/2023 16:40	SMPL
O2204-07	GBSR0	421	0.0131 -		SFAM01.1	PPB	4/17/2023 16:42	SMPL
O2204-08	GBSR1	1380	0.1886 -		SFAM01.1	PPB	4/17/2023 16:45	SMPL
O2204-09	GBSR2	378	0.0052 -		SFAM01.1	PPB	4/17/2023 16:47	SMPL
O2204-10	GBSR3	501	0.0277 -		SFAM01.1	PPB	4/17/2023 16:49	SMPL
O2204-11	CCB087	441	0.0167 -		SFAM01.1	PPB	4/17/2023 16:52	SMPL
O2204-12	GBSR5	559	0.0383 -		SFAM01.1	PPB	4/17/2023 16:54	SMPL
O2204-13	GBSR6	552	0.0371 -		SFAM01.1	PPB	4/17/2023 16:56	SMPL
O2204-14	GBSR7	244	-0.0193 -		SFAM01.1	PPB	4/17/2023 16:58	SMPL
O2204-15	GBSR8	1793	0.2642 -		SFAM01.1	PPB	4/17/2023 17:01	SMPL
O2204-16	GBSR9	3160	0.5144 -		SFAM01.1	PPB	4/17/2023 17:03	SMPL
O2204-17	GBSS0	2331	0.3627 -		SFAM01.1	PPB	4/17/2023 17:05	SMPL
O2204-18	GBSS1	3819	0.6351 -		SFAM01.1	PPB	4/17/2023 17:07	SMPL
O2204-19	GBSS2	5242	0.8955 -		SFAM01.1	PPB	4/17/2023 17:10	SMPL
CCV093	CCV093	27215	4.9176 -		SFAM01.1	PPB	4/17/2023 17:12	SMPL
CCB093	CCB093	385	0.0065 -		SFAM01.1	PPB	4/17/2023 17:14	SMPL
O2204-20	GBSS3	2103	0.321 -		SFAM01.1	PPB	4/17/2023 17:16	SMPL
O2204-21	GBSS4	8328	1.4604 -		SFAM01.1	PPB	4/17/2023 17:19	SMPL
O2204-22	GBSW6	3863	0.6431 -		SFAM01.1	PPB	4/17/2023 17:21	SMPL
PB152172BL	PBS172	158	-0.0351 -		SFAM01.1	PPB	4/17/2023 17:23	SMPL
O2205-01	GBSS5	5462	0.9358 -		SFAM01.1	PPB	4/17/2023 17:26	SMPL
O2205-02	GBSS6	4399	0.7412 -		SFAM01.1	PPB	4/17/2023 17:28	SMPL
O2205-03	GBSS7	3951	0.6592 -		SFAM01.1	PPB	4/17/2023 17:30	SMPL
O2205-04	GBSS8	2872	0.4617 -		SFAM01.1	PPB	4/17/2023 17:32	SMPL
O2205-05	GBSS8D	2781	0.4451 -		SFAM01.1	PPB	4/17/2023 17:35	SMPL
O2205-06	GBSS8S	16436	2.9446 -		SFAM01.1	PPB	4/17/2023 17:37	SMPL
O2205-07	GBSS9	3045	0.4934 -		SFAM01.1	PPB	4/17/2023 17:39	SMPL
O2205-08	GBST0	3058	0.4958 -		SFAM01.1	PPB	4/17/2023 17:41	SMPL
O2205-09	GBST1	3754	0.6232 -		SFAM01.1	PPB	4/17/2023 17:44	SMPL
O2205-10	GBST2	1824	0.2699 -		SFAM01.1	PPB	4/17/2023 17:46	SMPL

LB125008

INSTRUMENT ID : CV1

O2205-11	GBST3	358	0.0015 -	SFAM01.1	PPB	4/17/2023 17:48 SMPL
O2205-12	GBST4	413	0.0116 -	SFAM01.1	PPB	4/17/2023 17:51 SMPL
O2205-13	GBST5	1386	0.1897 -	SFAM01.1	PPB	4/17/2023 17:53 SMPL
O2205-14	GBST6	7349	1.2812 -	SFAM01.1	PPB	4/17/2023 17:55 SMPL
O2205-15	GBST7	5155	0.8796 -	SFAM01.1	PPB	4/17/2023 17:57 SMPL
O2205-16	GBST8	2605	0.4128 -	SFAM01.1	PPB	4/17/2023 18:00 SMPL
CCV094	CCV094	27334	4.9394 -	SFAM01.1	PPB	4/17/2023 18:02 SMPL
CCB094	CCB094	412	0.0114 -	SFAM01.1	PPB	4/17/2023 18:04 SMPL
O2205-17	GBST9	2677	0.426 -	SFAM01.1	PPB	4/17/2023 18:06 SMPL
O2205-18	GBSW0	2942	0.4745 -	SFAM01.1	PPB	4/17/2023 18:09 SMPL
O2205-19	GBSW1	548	0.0363 -	SFAM01.1	PPB	4/17/2023 18:11 SMPL
O2205-20	GBSW2	452	0.0188 -	SFAM01.1	PPB	4/17/2023 18:13 SMPL
O2205-21	GBSW3	2026	0.3069 -	SFAM01.1	PPB	4/17/2023 18:16 SMPL
O2205-22	GBSW7	4575	0.7734 -	SFAM01.1	PPB	4/17/2023 18:18 SMPL
CCV095	CCV095	27334	4.9394 -	SFAM01.1	PPB	4/17/2023 18:20 SMPL
CCB095	CCB095	384	0.0063 -	SFAM01.1	PPB	4/17/2023 18:22 SMPL