



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 4/19/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 04/18/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:06
Out Date: 04/19/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125016

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2382-01	EYXD8	1	1.13	8.56	9.69	7.82	78.2	
O2382-02	EYXD9	2	1.13	8.76	9.89	7.93	77.6	
O2382-03	EYXE0	3	1.15	8.39	9.54	7.8	79.3	
O2382-04	EYXE1	4	1.13	8.68	9.81	8.07	80.0	
O2382-05	EYXE2	5	1.15	8.81	9.96	8.04	78.2	
O2382-06	EYXE3	6	1.13	8.57	9.7	7.96	79.7	
O2382-07	EYXE4	7	1.15	8.82	9.97	8.38	82.0	
O2382-08	EYXE5	8	1.1	8.62	9.72	7.83	78.1	
O2382-09	EYXE6	9	1.15	8.77	9.92	8.24	80.8	
O2382-10	EYXE7	10	1.15	8.84	9.99	7.93	76.7	
O2382-11	EYXE8	11	1.15	8.79	9.94	8.05	78.5	
O2382-12	EYXE8D	12	1.15	8.79	9.94	8.05	78.5	
O2382-13	EYXE8S	13	1.15	8.79	9.94	8.05	78.5	
O2382-14	EYXF0	14	1.19	8.49	9.68	7.83	78.2	
O2382-15	EYXF1	15	1.15	8.40	9.55	7.66	77.5	
O2382-16	EYXF2	16	1.18	8.52	9.7	8.21	82.5	
O2382-17	EYXF3	17	1.19	8.60	9.79	7.76	76.4	
O2382-18	EYXF4	18	1.18	8.59	9.77	7.81	77.2	
O2382-19	EYXF5	19	1.15	8.81	9.96	8.34	81.6	
O2382-20	EYXF6	20	1.16	8.82	9.98	8.39	82.0	
O2382-21	EYXF7	21	1.1	8.65	9.75	8.08	80.7	
O2382-22	EYXF8	22	1.16	8.54	9.7	7.88	78.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$