

LB125023


7471B

INSTRUMENT ID: EV1

Linear

 μ Abs.:

60353



A= 0.0000e+000

B= 1.6947e-004

C= -1.1436e-002

Rho= 0.9987108

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.056	0.056	396	0.000	396					1
0.2	0.200	0.212	0.012	1321	0.0 %	1321					6
2.5	2.500	2.496	-0.004	14797	0.0 %	14797					0
5.0	5.000	5.105	0.105	30191	0.0 %	30191					2
7.5	7.500	7.114	-0.386	42043	0.0 %	42043					5
10.0	10.000	10.217	0.217	60353	0.0 %	60353					2

LB125023 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Apr 2023 11:45:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	396	-	0.0000	7471B	PPB	18 Apr 2023 11:53:14	S	Std
0.2 - 1	502	1321	-	0.2000	7471B	PPB	18 Apr 2023 11:55:32	S	Std
2.5 - 1	525	14797	-	2.5000	7471B	PPB	18 Apr 2023 11:57:49	S	Std
5.0 - 1	55	30191	-	5.0000	7471B	PPB	18 Apr 2023 12:00:06	S	Std
7.5 - 1	575	42043	-	7.5000	7471B	PPB	18 Apr 2023 12:02:23	S	Std
10.0 - 1	50	60353	-	10.0000	7471B	PPB	18 Apr 2023 12:10:02	S	Std
ICV28 - 1	ICV28	25163	4.2530	-	7471B	PPB	18 Apr 2023 12:12:56	U	SMPL
ICB28 - 1	ICB28	499	0.0731	-	7471B	PPB	18 Apr 2023 12:15:11	U	SMPL
CCV96 - 1	CCV96	30460	5.1507	-	7471B	PPB	18 Apr 2023 12:17:29	U	SMPL
CCB96 - 1	CCB96	525	0.0775	-	7471B	PPB	18 Apr 2023 12:19:44	U	SMPL
CRA - 1	CRA	1193	0.1907	-	7471B	PPB	18 Apr 2023 12:22:01	U	SMPL
HighStd - 1	HighStd	60490	10.2400	-	7471B	PPB	18 Apr 2023 12:24:16	U	SMPL
ChkStd - 1	ChkStd	39521	6.6863	-	7471B	PPB	18 Apr 2023 12:26:31	U	SMPL
PB152180BL - 1	PBS	343	0.0467	-	7471B	PPB	18 Apr 2023 12:28:50	U	SMPL
PB152180BS - 1	LCSS	22281	3.7646	-	7471B	PPB	18 Apr 2023 12:33:41	U	SMPL
O2348-01 - 1	TR-05-041323	6907	1.1591	-	7471B	PPB	18 Apr 2023 12:35:57	U	SMPL
O2351-01 - 1	TRE-1691	1498	0.2424	-	7471B	PPB	18 Apr 2023 12:38:15	U	SMPL
O2357-01 - 1	P2-S1-230413	17312	2.9225	-	7471B	PPB	18 Apr 2023 12:40:34	U	SMPL
O2357-02 - 1	P2-S2-230413	12669	2.1356	-	7471B	PPB	18 Apr 2023 12:42:51	U	SMPL
O2362-01 - 1	HW-STONE-1-0414	598	0.0899	-	7471B	PPB	18 Apr 2023 12:45:07	U	SMPL
CCV97 - 1	CCV97	30736	5.1975	-	7471B	PPB	18 Apr 2023 12:47:23	U	SMPL
CCB97 - 1	CCB97	537	0.0796	-	7471B	PPB	18 Apr 2023 12:49:39	U	SMPL
O2362-03 - 1	HW-STONE-2-0414	824	0.1282	-	7471B	PPB	18 Apr 2023 12:51:56	U	SMPL
O2366-06 - 1	20230330-COMP-8	1394	0.2248	-	7471B	PPB	18 Apr 2023 12:54:11	U	SMPL
O2366-07 - 1	20230335-COMP-13	5846	0.9793	-	7471B	PPB	18 Apr 2023 12:56:27	U	SMPL
O2366-08 - 1	20230339-COMP-17	38692	6.5458	-	7471B	PPB	18 Apr 2023 12:58:43	U	SMPL
O2366-09 - 1	20230341-COMP-19	59786	10.1207	-	7471B	PPB	18 Apr 2023 13:00:59	U	SMPL
O2366-10 - 1	20230342-COMP-20	64556	10.9291	-	7471B	PPB	18 Apr 2023 13:03:18	U	SMPL
O2366-11 - 1	20230348-COMP-26	42865	7.2530	-	7471B	PPB	18 Apr 2023 13:05:37	U	SMPL
O2369-01 - 1	SU-04-041423	1400	0.2258	-	7471B	PPB	18 Apr 2023 13:07:56	U	SMPL
O2370-01 - 1	HR-04-041423	8297	1.3947	-	7471B	PPB	18 Apr 2023 13:10:16	U	SMPL
O2384-01 - 1	ROLL-OFF-8PT-COMP	45379	7.6791	-	7471B	PPB	18 Apr 2023 13:12:32	U	SMPL
CCV98 - 1	CCV98	31250	5.2846	-	7471B	PPB	18 Apr 2023 13:14:48	U	SMPL
CCB98 - 1	CCB98	565	0.0843	-	7471B	PPB	18 Apr 2023 13:17:06	U	SMPL
O2385-01 - 1	BU-3-041723	1861	0.3040	-	7471B	PPB	18 Apr 2023 13:19:24	U	SMPL
O2388-01 - 1	CC-1	2096	0.3438	-	7471B	PPB	18 Apr 2023 13:21:39	U	SMPL
O2388-01DUP - 1	CC-1DUP	2071	0.3395	-	7471B	PPB	18 Apr 2023 13:23:55	U	SMPL
O2388-01MS - 1	CC-1MS	24235	4.0957	-	7471B	PPB	18 Apr 2023 13:26:11	U	SMPL
O2388-01MSD - 1	CC-1MSD	23409	3.9558	-	7471B	PPB	18 Apr 2023 13:28:27	U	SMPL
O2388-01LX5 - 1		420	0.0597	-	7471B	PPB	18 Apr 2023 13:33:46	U	SMPL
O2388-01A - 1		25950	4.3864	-	7471B	PPB	18 Apr 2023 13:36:02	U	SMPL
O2366-09DLX2 - 1		31363	5.3038	-	7471B	PPB	18 Apr 2023 13:38:20	U	SMPL
O2366-10DLX2 - 1		33513	5.6681	-	7471B	PPB	18 Apr 2023 13:40:38	U	SMPL
CCV99 - 1	CCV99	31102	5.2595	-	7471B	PPB	18 Apr 2023 13:42:56	U	SMPL
CCB99 - 1	CCB99	578	0.0865	-	7471B	PPB	18 Apr 2023 13:45:15	U	SMPL