

LB125026


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52122



A= 0.0000e+000

B= 1.9100e-004 slope

C= -4.1951e-002 $\rightarrow$ y-intercept

Rho= 0.9997879

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.010	-0.010	168	0.000	168					
0.05	0.050					438	0				-16
0.20	0.200	0.169	-0.031	1103	0.0 %	1103					-2
2.50	2.500	2.449	-0.051	13041	0.0 %	13041					3
5.00	5.000	5.150	0.150	27185	0.0 %	27185					0
7.50	7.500	7.529	0.029	39638	0.0 %	39638					-1
10.0	10.000	9.913	-0.087	52122	0.0 %	52122					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	168	-		0 SFAM01.1	PPB	4/18/2023 15:07	Std
	0.2 S01	1103	-		0.2 SFAM01.1	PPB	4/18/2023 15:10	Std
	2.5 S02	13041	-		2.5 SFAM01.1	PPB	4/18/2023 15:12	Std
	5 S03	27185	-		5 SFAM01.1	PPB	4/18/2023 15:14	Std
	7.5 S04	39638	-		7.5 SFAM01.1	PPB	4/18/2023 15:17	Std
	10 S05	52122	-		10 SFAM01.1	PPB	4/18/2023 15:19	Std
ICV030	ICV030	23069	4.3642 -		SFAM01.1	PPB	4/18/2023 15:25	SMPL
ICB030	ICB030	57	-0.0311 -		SFAM01.1	PPB	4/18/2023 15:27	SMPL
CCV001	CCV001	27213	5.1557 -		SFAM01.1	PPB	4/18/2023 15:29	SMPL
CCB001	CCB001	46	-0.0332 -		SFAM01.1	PPB	4/18/2023 15:32	SMPL
PB152200BL	PBS200	151	-0.0131 -		SFAM01.1	PPB	4/18/2023 15:34	SMPL
O2202-01	GBSL8	876	0.1254 -		SFAM01.1	PPB	4/18/2023 15:36	SMPL
O2202-02	GBSL8D	593	0.0713 -		SFAM01.1	PPB	4/18/2023 15:38	SMPL
O2202-03	GBSL8S	13412	2.5197 -		SFAM01.1	PPB	4/18/2023 15:41	SMPL
O2202-04	GBSL9	2321	0.4014 -		SFAM01.1	PPB	4/18/2023 15:43	SMPL
O2202-05	GBSM0	5162	0.944 -		SFAM01.1	PPB	4/18/2023 15:45	SMPL
O2202-06	GBSM1	540	0.0612 -		SFAM01.1	PPB	4/18/2023 15:47	SMPL
O2202-07	GBSM2	3270	0.5826 -		SFAM01.1	PPB	4/18/2023 15:50	SMPL
O2202-08	GBSM3	3071	0.5446 -		SFAM01.1	PPB	4/18/2023 15:52	SMPL
O2202-09	GBSM4	2136	0.366 -		SFAM01.1	PPB	4/18/2023 15:54	SMPL
O2202-10	GBSM5	663	0.0847 -		SFAM01.1	PPB	4/18/2023 15:57	SMPL
O2202-11	GBSM6	963	0.142 -		SFAM01.1	PPB	4/18/2023 15:59	SMPL
O2202-12	GBSM7	559	0.0648 -		SFAM01.1	PPB	4/18/2023 16:01	SMPL
O2202-13	GBSM8	540	0.0612 -		SFAM01.1	PPB	4/18/2023 16:03	SMPL
O2202-14	GBSM9	410	0.0364 -		SFAM01.1	PPB	4/18/2023 16:06	SMPL
O2202-15	GBSN0	688	0.0895 -		SFAM01.1	PPB	4/18/2023 16:08	SMPL
O2202-16	GBSN1	961	0.1416 -		SFAM01.1	PPB	4/18/2023 16:10	SMPL
O2202-17	GBSN2	1621	0.2677 -		SFAM01.1	PPB	4/18/2023 16:12	SMPL
O2202-18	GBSN3	535	0.0602 -		SFAM01.1	PPB	4/18/2023 16:15	SMPL
O2202-19	GBSN4	427	0.0396 -		SFAM01.1	PPB	4/18/2023 16:17	SMPL
CCV002	CCV002	27582	5.2261 -		SFAM01.1	PPB	4/18/2023 16:19	SMPL
CCB002	CCB002	67	-0.0292 -		SFAM01.1	PPB	4/18/2023 16:22	SMPL
O2202-20	GBSN5	444	0.0429 -		SFAM01.1	PPB	4/18/2023 16:24	SMPL
O2202-21	GBSN6	293	0.014 -		SFAM01.1	PPB	4/18/2023 16:26	SMPL
O2202-22	GBSW4	3401	0.6076 -		SFAM01.1	PPB	4/18/2023 16:28	SMPL
PB152201BL	PBS201	89	-0.025 -		SFAM01.1	PPB	4/18/2023 16:33	SMPL
O2203-01	GBSN7	359	0.0266 -		SFAM01.1	PPB	4/18/2023 16:35	SMPL
O2203-02	GBSN8	3627	0.6508 -		SFAM01.1	PPB	4/18/2023 16:37	SMPL
O2203-03	GBSN8D	4165	0.7536 -		SFAM01.1	PPB	4/18/2023 16:40	SMPL
O2203-04	GBSN8S	16251	3.0619 -		SFAM01.1	PPB	4/18/2023 16:42	SMPL
O2203-05	GBSN9	5563	1.0206 -		SFAM01.1	PPB	4/18/2023 16:44	SMPL
O2203-06	GBSP0	1670	0.277 -		SFAM01.1	PPB	4/18/2023 16:47	SMPL
O2203-07	GBSP1	688	0.0895 -		SFAM01.1	PPB	4/18/2023 16:49	SMPL
O2203-08	GBSP2	758	0.1028 -		SFAM01.1	PPB	4/18/2023 16:51	SMPL
O2203-09	GBSP3	2583	0.4514 -		SFAM01.1	PPB	4/18/2023 16:53	SMPL
O2203-10	GBSP4	6142	1.1312 -		SFAM01.1	PPB	4/18/2023 16:56	SMPL

LB125026

INSTRUMENT ID : CV1

O2203-11	GBSP5	4447	0.8074 -	SFAM01.1	PPB	4/18/2023 16:58 SMPL
O2203-12	GBSP6	14096	2.6503 -	SFAM01.1	PPB	4/18/2023 17:00 SMPL
O2203-13	GBSP7	5159	0.9434 -	SFAM01.1	PPB	4/18/2023 17:02 SMPL
O2203-14	GBSP8	5834	1.0723 -	SFAM01.1	PPB	4/18/2023 17:05 SMPL
O2203-15	GBSP9	3569	0.6397 -	SFAM01.1	PPB	4/18/2023 17:07 SMPL
CCV003	CCV003	27361	5.1839 -	SFAM01.1	PPB	4/18/2023 17:09 SMPL
CCB003	CCB003	67	-0.0292 -	SFAM01.1	PPB	4/18/2023 17:12 SMPL
O2203-16	GBSQ0	5176	0.9467 -	SFAM01.1	PPB	4/18/2023 17:14 SMPL
O2203-17	GBSQ1	460	0.0459 -	SFAM01.1	PPB	4/18/2023 17:16 SMPL
O2203-18	GBSQ2	369	0.0285 -	SFAM01.1	PPB	4/18/2023 17:18 SMPL
O2203-19	GBSQ3	530	0.0593 -	SFAM01.1	PPB	4/18/2023 17:21 SMPL
O2203-20	GBSQ4	291	0.0136 -	SFAM01.1	PPB	4/18/2023 17:23 SMPL
O2203-21	GBSQ5	741	0.0996 -	SFAM01.1	PPB	4/18/2023 17:25 SMPL
O2203-22	GBSW5	4727	0.8609 -	SFAM01.1	PPB	4/18/2023 17:27 SMPL
O2202-23	GBSW8	26388	4.9981 -	SFAM01.1	PPB	4/18/2023 17:34 SMPL
CCV004	CCV004	26965	5.1083 -	SFAM01.1	PPB	4/18/2023 17:41 SMPL
CCB004	CCB004	73	-0.028 -	SFAM01.1	PPB	4/18/2023 17:43 SMPL