



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125029

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/18/2023	16:10
2	CAL2	1	Water	NA	NA	20.7	7.00	04/18/2023	16:11
3	CAL3	1	Water	NA	NA	20.7	10.03	04/18/2023	16:14
4	ICV	1	Water	NA	NA	20.6	6.98	04/18/2023	16:15
5	CCV1	1	Water	NA	NA	20.8	2.01	04/18/2023	16:20
6	O2403-01	1	Water	NA	NA	20.4	12.14	04/18/2023	16:22
7	O2403-01DUP	1	Water	NA	NA	20.5	12.15	04/18/2023	16:23
8	O2403-02	1	Water	NA	NA	20.7	11.79	04/18/2023	16:25
9	O2403-03	1	Water	NA	NA	20.9	11.99	04/18/2023	16:29
10	O2403-04	1	Water	NA	NA	20.8	12.03	04/18/2023	16:30
11	O2403-05	1	Water	NA	NA	20.9	12.09	04/18/2023	16:32
12	O2403-06	1	Water	NA	NA	20.8	12.06	04/18/2023	16:35
13	O2403-07	1	Water	NA	NA	20.4	12.09	04/18/2023	16:40
14	O2403-08	1	Water	NA	NA	20.1	12.06	04/18/2023	16:42
15	O2403-09	1	Water	NA	NA	20.8	11.93	04/18/2023	16:45
16	CCV2	1	Water	NA	NA	20.7	12.02	04/18/2023	16:47
17	O2403-10	1	Water	NA	NA	20.1	12.05	04/18/2023	16:50
18	O2403-11	1	Water	NA	NA	20.1	12.09	04/18/2023	17:00
19	O2403-12	1	Water	NA	NA	20.3	12.01	04/18/2023	17:02
20	O2403-13	1	Water	NA	NA	20.6	12.04	04/18/2023	17:05

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	O2403-14	1	Water	NA	NA	20.5	12.10	04/18/2023	17:10
22	O2403-15	1	Water	NA	NA	20.4	12.02	04/18/2023	17:15
23	O2403-16	1	Water	NA	NA	20.5	12.07	04/18/2023	17:20
24	O2403-17	1	Water	NA	NA	20.6	12.04	04/18/2023	17:25
25	O2403-18	1	Water	NA	NA	20.8	11.92	04/18/2023	17:30
26	O2403-18DUP	1	Water	NA	NA	20.9	11.93	04/18/2023	17:31
27	CCV3	1	Water	NA	NA	20.7	2.01	04/18/2023	17:33