

LB125035


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44994



A= 0.0000e+000

B= 2.2464e-004

C= 2.6841e-002

Rho= 0.9997201

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.043	0.043	73	0.000	73	0				1
0.2	0.200	0.229	0.029	902	0.0 %	902					15
2.5	2.500	2.496	-0.004	10991	0.0 %	10991					0
5.0	5.000	4.945	-0.055	21894	0.0 %	21894					-1
7.5	7.500	7.352	-0.148	32608	0.0 %	32608					-2
10.0	10.000	10.134	0.134	44994	0.0 %	44994					1

LB125035 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Apr 2023 08:30:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	73	-	0.0000	7470A	PPB	19 Apr 2023 08:40:28	S	Std
0.2 - 1	50.2	902	-	0.2000	7470A	PPB	19 Apr 2023 08:42:46	S	Std
2.5 - 1	52.5	10991	-	2.5000	7470A	PPB	19 Apr 2023 08:45:03	S	Std
5.0 - 1	55	21894	-	5.0000	7470A	PPB	19 Apr 2023 08:47:20	S	Std
7.5 - 1	57.5	32608	-	7.5000	7470A	PPB	19 Apr 2023 08:52:16	S	Std
10.0 - 1	59.8	44994	-	10.0000	7470A	PPB	19 Apr 2023 08:54:32	S	Std
ICV31 - 1	ICV31	18158	4.1059	-	7470A	PPB	19 Apr 2023 08:59:32	U	SMPL
ICB31 - 1	ICB31	-141	-0.0048	-	7470A	PPB	19 Apr 2023 09:01:49	U	SMPL
CCV05 - 1	CCV05	20795	4.6983	-	7470A	PPB	19 Apr 2023 09:06:39	U	SMPL
CCB05 - 1	CCB05	-137	-0.0039	-	7470A	PPB	19 Apr 2023 09:08:54	U	SMPL
CRA - 1	CRA	844	0.2164	-	7470A	PPB	19 Apr 2023 09:11:10	U	SMPL
HighStd - 1	HighStd	45102	10.1586	-	7470A	PPB	19 Apr 2023 09:13:25	U	SMPL
ChkStd - 1	ChkStd	30712	6.9260	-	7470A	PPB	19 Apr 2023 09:15:40	U	SMPL
PB152193BL - 1	PBW	-282	-0.0365	-	7470A	PPB	19 Apr 2023 09:17:59	U	SMPL
O2365-01 - 1	COMP-1	-100	0.0044	-	7470A	PPB	19 Apr 2023 09:22:34	U	SMPL
O2365-02 - 1	COMP-2	42	0.0363	-	7470A	PPB	19 Apr 2023 09:24:50	U	SMPL
O2365-03 - 1	COMP-3	10	0.0291	-	7470A	PPB	19 Apr 2023 09:27:07	U	SMPL
O2371-02 - 1	SB-26-COMP	63	0.0410	-	7470A	PPB	19 Apr 2023 09:29:25	U	SMPL
O2371-04 - 1	SB-23-COMP	42	0.0363	-	7470A	PPB	19 Apr 2023 09:31:40	U	SMPL
PB152193BS - 1	LCSW	17030	3.8525	-	7470A	PPB	19 Apr 2023 09:35:05	U	SMPL
CCV06 - 1	CCV06	20808	4.7012	-	7470A	PPB	19 Apr 2023 09:37:42	U	SMPL
CCB06 - 1	CCB06	-97	0.0051	-	7470A	PPB	19 Apr 2023 09:39:57	U	SMPL
O2371-06 - 1	SB-22-COMP	71	0.0428	-	7470A	PPB	19 Apr 2023 09:42:14	U	SMPL
O2371-08 - 1	SB-21-COMP	50	0.0381	-	7470A	PPB	19 Apr 2023 09:44:30	U	SMPL
O2371-10 - 1	SB-20-COMP	37	0.0352	-	7470A	PPB	19 Apr 2023 09:46:45	U	SMPL
O2371-12 - 1	SB-19-COMP	62	0.0408	-	7470A	PPB	19 Apr 2023 09:49:01	U	SMPL
O2371-14 - 1	SB-18-COMP	64	0.0412	-	7470A	PPB	19 Apr 2023 09:51:18	U	SMPL
O2371-16 - 1	SB-17-COMP	97	0.0486	-	7470A	PPB	19 Apr 2023 09:53:34	U	SMPL
O2371-18 - 1	SB-16-COMP	53	0.0387	-	7470A	PPB	19 Apr 2023 09:55:51	U	SMPL
O2384-02 - 1	ROLL-OFF-8PT-COMP	56	0.0394	-	7470A	PPB	19 Apr 2023 09:58:08	U	SMPL
O2386-04 - 1	SPA-1WC-23-04	980	0.2470	-	7470A	PPB	19 Apr 2023 10:00:25	U	SMPL
O2386-08 - 1	SPA-2WC-23-04	818	0.2106	-	7470A	PPB	19 Apr 2023 10:02:41	U	SMPL
CCV07 - 1	CCV07	20325	4.5927	-	7470A	PPB	19 Apr 2023 10:04:58	U	SMPL
CCB07 - 1	CCB07	-51	0.0154	-	7470A	PPB	19 Apr 2023 10:07:14	U	SMPL
O2386-12 - 1	SPA-3WC-23-04	5339	1.2262	-	7470A	PPB	19 Apr 2023 10:09:31	U	SMPL
O2386-16 - 1	SPA-4WC-23-04	4701	1.0829	-	7470A	PPB	19 Apr 2023 10:11:47	U	SMPL
O2386-20 - 1	SPA-5WC-23-04	1580	0.3818	-	7470A	PPB	19 Apr 2023 10:14:04	U	SMPL
O2386-20DUP - 1	SPA-5WC-23-04DUP	1776	0.4258	-	7470A	PPB	19 Apr 2023 10:18:35	U	SMPL
O2386-20MSD - 1	SPA-5WC-23-04MSD	17753	4.0149	-	7470A	PPB	19 Apr 2023 10:23:08	U	SMPL
O2386-20MS - 1	SPA-5WC-23-04MS	17876	4.0425	-	7470A	PPB	19 Apr 2023 10:25:43	U	SMPL
PB152194BL - 1	PBW	19	0.0311	-	7470A	PPB	19 Apr 2023 10:28:38	U	SMPL
PB152194BS - 1	LCSW	18500	4.1827	-	7470A	PPB	19 Apr 2023 10:33:52	U	SMPL
O2365-04 - 1	ELA-23-00004	-48	0.0161	-	7470A	PPB	19 Apr 2023 10:36:09	U	SMPL
O2365-05 - 1	ELA-23-00012	45	0.0369	-	7470A	PPB	19 Apr 2023 10:38:25	U	SMPL
CCV08 - 1	CCV08	21010	4.7466	-	7470A	PPB	19 Apr 2023 10:40:41	U	SMPL
CCB08 - 1	CCB08	-48	0.0161	-	7470A	PPB	19 Apr 2023 10:42:57	U	SMPL
O2365-05DUP - 1	ELA-23-00012DUP	28	0.0331	-	7470A	PPB	19 Apr 2023 10:45:14	U	SMPL
O2365-05MS - 1	ELA-23-00012MS	16643	3.7655	-	7470A	PPB	19 Apr 2023 10:47:29	U	SMPL
O2365-05MSD - 1	ELA-23-00012MSD	14529	3.2907	-	7470A	PPB	19 Apr 2023 10:49:45	U	SMPL
PB152196BL - 1	PBW	-76	0.0098	-	7470A	PPB	19 Apr 2023 10:52:02	U	SMPL
PB152196BS - 1	LCSW	21152	4.7785	-	7470A	PPB	19 Apr 2023 10:56:57	U	SMPL
O2353-01 - 1	BUR-1253	-95	0.0055	-	7470A	PPB	19 Apr 2023 10:59:13	U	SMPL
O2353-02 - 1	BUR-1260	109	0.0513	-	7470A	PPB	19 Apr 2023 11:01:30	U	SMPL
O2353-03 - 1	BUR-1261	2295	0.5424	-	7470A	PPB	19 Apr 2023 11:03:47	U	SMPL
O2353-04 - 1	BUR-1262	37	0.0352	-	7470A	PPB	19 Apr 2023 11:06:04	U	SMPL
O2353-05 - 1	BUR-1263	151	0.0608	-	7470A	PPB	19 Apr 2023 11:09:28	U	SMPL
CCV09 - 1	CCV09	21002	4.7448	-	7470A	PPB	19 Apr 2023 11:20:11	U	SMPL
CCB09 - 1	CCB09	-8	0.0250	-	7470A	PPB	19 Apr 2023 11:25:16	U	SMPL
O2353-05DUP - 1	BUR-1263DUP	118	0.0533	-	7470A	PPB	19 Apr 2023 11:27:32	U	SMPL
O2353-05MS - 1	BUR-1263MS	8714	1.9844	-	7470A	PPB	19 Apr 2023 11:31:57	U	SMPL
O2353-05MSD - 1	BUR-1263MSD	8509	1.9383	-	7470A	PPB	19 Apr 2023 11:34:13	U	SMPL
PB152149TB - 1	PB152149TB	40	0.0358	-	7470A	PPB	19 Apr 2023 11:39:39	U	SMPL
PB152150TB - 1	PB152150TB	-5	0.0257	-	7470A	PPB	19 Apr 2023 11:41:55	U	SMPL
CCV10 - 1	CCV10	21751	4.9130	-	7470A	PPB	19 Apr 2023 11:44:11	U	SMPL
CCB10 - 1	CCB10	-267	-0.0331	-	7470A	PPB	19 Apr 2023 11:46:28	U	SMPL
O2386-20LX5 - 1		102	0.0498	-	7470A	PPB	19 Apr 2023 11:48:48	U	SMPL
O2386-20A - 1		17448	3.9464	-	7470A	PPB	19 Apr 2023 11:51:05	U	SMPL
O2365-05LX5 - 1		-521	-0.0902	-	7470A	PPB	19 Apr 2023 11:53:21	U	SMPL
O2365-05A - 1		19136	4.3256	-	7470A	PPB	19 Apr 2023 11:55:41	U	SMPL
O2353-05LX5 - 1		-345	-0.0507	-	7470A	PPB	19 Apr 2023 11:57:57	U	SMPL
O2353-05A - 1		17949	4.0589	-	7470A	PPB	19 Apr 2023 12:00:16	U	SMPL
CCV11 - 1	CCV11	22154	5.0035	-	7470A	PPB	19 Apr 2023 12:02:33	U	SMPL
CCB11 - 1	CCB11	-330	-0.0473	-	7470A	PPB	19 Apr 2023 12:04:51	U	SMPL