

LB125053


7471B

INSTRUMENT ID: CY 1

Linear

 μ Abs.:

57431



A= 0.0000e+000

B= 1.7162e-004

C= -1.4670e-002

Rho= 0.9996070

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.0	0.000	-0.008	-0.008	37	0.000	37					1
0.2	0.200	0.177	-0.023	1119	0.0 %	1119					-1/2
2.5	2.500	2.451	-0.049	14364	0.0 %	14364					-2
5.0	5.000	5.057	0.057	29554	0.0 %	29554					1
7.5	7.500	7.681	0.181	44842	0.0 %	44842					2
10.0	10.000	9.842	-0.158	57431	0.0 %	57431					-2

LB125053 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Apr 2023 10:05:43

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	37	-	0.0000	7471B	PPB	20 Apr 2023 10:19:42	S	Std
0.2 - 1	51.2	1119	-	0.2000	7471B	PPB	20 Apr 2023 10:22:00	S	Std
2.5 - 1	52.5	14364	-	2.5000	7471B	PPB	20 Apr 2023 10:24:18	S	Std
5.0 - 1	55	29554	-	5.0000	7471B	PPB	20 Apr 2023 10:26:36	S	Std
7.5 - 1	57.5	44842	-	7.5000	7471B	PPB	20 Apr 2023 10:28:52	S	Std
10.0 - 1	510	57431	-	10.0000	7471B	PPB	20 Apr 2023 10:31:10	S	Std
ICV33 - 1	ICV33	24223	4.1425	-	7471B	PPB	20 Apr 2023 10:36:05	U	SMPL
ICB33 - 1	ICB33	-130	-0.0370	-	7471B	PPB	20 Apr 2023 10:38:22	U	SMPL
CCV15 - 1	CCV15	27977	4.7868	-	7471B	PPB	20 Apr 2023 10:40:39	U	SMPL
CCB15 - 1	CCB15	-167	-0.0433	-	7471B	PPB	20 Apr 2023 10:42:54	U	SMPL
CRA - 1	CRA	1182	0.1882	-	7471B	PPB	20 Apr 2023 10:45:11	U	SMPL
HighStd - 1	HighStd	57163	9.7958	-	7471B	PPB	20 Apr 2023 10:47:26	U	SMPL
ChkStd - 1	ChkStd	40907	7.0059	-	7471B	PPB	20 Apr 2023 10:49:42	U	SMPL
PB152234BL - 1	PBS	-171	-0.0440	-	7471B	PPB	20 Apr 2023 10:52:00	U	SMPL
PB152234BS - 1	LCSS	19876	3.3965	-	7471B	PPB	20 Apr 2023 10:54:19	U	SMPL
O2270-01 - 1	SB-17-0-2-20230410	8070	1.3703	-	7471B	PPB	20 Apr 2023 10:56:35	U	SMPL
O2270-02 - 1	SB-17-7-9-20230410	1092	0.1727	-	7471B	PPB	20 Apr 2023 10:58:54	U	SMPL
O2270-07 - 1	SB-25-0-2-20230411	58708	10.0608	-	7471B	PPB	20 Apr 2023 11:01:14	U	SMPL
O2270-08 - 1	SB-25-7.5-9.5-20230411	10722	1.8255	-	7471B	PPB	20 Apr 2023 11:03:32	U	SMPL
O2270-10 - 1	SB-27-0-2-20230411	17430	2.9767	-	7471B	PPB	20 Apr 2023 11:05:51	U	SMPL
CCV16 - 1	CCV16	28505	4.8774	-	7471B	PPB	20 Apr 2023 11:11:09	U	SMPL
CCB16 - 1	CCB16	-81	-0.0286	-	7471B	PPB	20 Apr 2023 11:13:24	U	SMPL
O2270-11 - 1	SB-27-8-10-20230411	10128	1.7235	-	7471B	PPB	20 Apr 2023 11:15:41	U	SMPL
O2270-13 - 1	SB-24-0-2-20230411	35126	6.0137	-	7471B	PPB	20 Apr 2023 11:17:57	U	SMPL
O2270-14 - 1	SB-24-5-7-20230411	9343	1.5888	-	7471B	PPB	20 Apr 2023 11:20:15	U	SMPL
O2270-15 - 1	SB-19-0-2-20230411	16265	2.7768	-	7471B	PPB	20 Apr 2023 11:22:33	U	SMPL
O2270-16 - 1	SB-19-5-7-20230411	13173	2.2461	-	7471B	PPB	20 Apr 2023 11:24:50	U	SMPL
O2270-18 - 1	SB-31-0-2-20230412	16350	2.7914	-	7471B	PPB	20 Apr 2023 11:27:08	U	SMPL
O2270-20 - 1	SB-30-0-2-20230412	12730	2.1701	-	7471B	PPB	20 Apr 2023 11:29:27	U	SMPL
O2270-22 - 1	SB-15-0-2-20230412	24837	4.2479	-	7471B	PPB	20 Apr 2023 11:31:45	U	SMPL
O2270-23 - 1	SB-15-5-7-20230412	5108	0.8620	-	7471B	PPB	20 Apr 2023 11:34:04	U	SMPL
O2270-25 - 1	SB-16-0-2-20230413	55894	9.5780	-	7471B	PPB	20 Apr 2023 11:36:21	U	SMPL
CCV17 - 1	CCV17	29346	5.0218	-	7471B	PPB	20 Apr 2023 11:38:37	U	SMPL
CCB17 - 1	CCB17	-63	-0.0255	-	7471B	PPB	20 Apr 2023 11:40:56	U	SMPL
O2270-25DUP - 1	SB-16-0-2-20230413DUP	57358	9.8292	-	7471B	PPB	20 Apr 2023 11:43:14	U	SMPL
O2270-25MS - 1	SB-16-0-2-20230413MS	82837	14.2020	-	7471B	PPB	20 Apr 2023 11:45:30	U	SMPL
O2270-25MSD - 1	SB-16-0-2-20230413MSD	83006	14.2310	-	7471B	PPB	20 Apr 2023 11:47:48	U	SMPL
O2416-01 - 1	NATIONAL-GRID-COMP	10458	1.7802	-	7471B	PPB	20 Apr 2023 11:50:06	U	SMPL
O2426-01 - 1	EO-3-041923	5102	0.8609	-	7471B	PPB	20 Apr 2023 11:52:25	U	SMPL
PB152235BL - 1	PBS	223	0.0236	-	7471B	PPB	20 Apr 2023 11:54:42	U	SMPL
PB152235BS - 1	LCSS	22519	3.8501	-	7471B	PPB	20 Apr 2023 11:56:59	U	SMPL
CCV18 - 1	CCV18	27479	4.7013	-	7471B	PPB	20 Apr 2023 12:22:51	U	SMPL
CCB18 - 1	CCB18	-423	-0.0873	-	7471B	PPB	20 Apr 2023 12:50:09	U	SMPL
O2393-01 - 1	VNU-229	7928	1.3460	-	7471B	PPB	20 Apr 2023 12:53:17	U	SMPL
O2404-01 - 1	OK-2-041823	11947	2.0357	-	7471B	PPB	20 Apr 2023 12:55:37	U	SMPL
O2407-01 - 1	NORTH-ST-BAY-AVE-INTERSECTION	4019	0.6751	-	7471B	PPB	20 Apr 2023 12:57:56	U	SMPL
O2407-03 - 1	BEACH-BLVD-TEST-PIT	4727	0.7966	-	7471B	PPB	20 Apr 2023 13:00:15	U	SMPL
O2407-03DUP - 1	BEACH-BLVD-TEST-PITDUP	5559	0.9394	-	7471B	PPB	20 Apr 2023 13:02:32	U	SMPL
O2407-03MS - 1	BEACH-BLVD-TEST-PITMS	30274	5.1810	-	7471B	PPB	20 Apr 2023 13:04:49	U	SMPL
O2407-03MSD - 1	BEACH-BLVD-TEST-PITMSD	26440	4.5230	-	7471B	PPB	20 Apr 2023 13:07:07	U	SMPL
O2270-07DLX2 - 1		30464	5.2136	-	7471B	PPB	20 Apr 2023 13:19:57	U	SMPL
CCV19 - 1	CCV19	29750	5.0911	-	7471B	PPB	20 Apr 2023 13:22:16	U	SMPL
CCB19 - 1	CCB19	-226	-0.0535	-	7471B	PPB	20 Apr 2023 13:27:23	U	SMPL
O2270-25LX5 - 1		10970	1.8680	-	7471B	PPB	20 Apr 2023 13:29:39	U	SMPL
O2270-25A - 1		84967	14.5676	-	7471B	PPB	20 Apr 2023 13:31:55	U	SMPL
O2407-03A - 1		29715	5.0851	-	7471B	PPB	20 Apr 2023 13:36:32	U	SMPL
O2270-25DLX2 - 1		28981	4.9591	-	7471B	PPB	20 Apr 2023 13:38:49	U	SMPL
O2270-25DUPDLX2 - 1		28546	4.8845	-	7471B	PPB	20 Apr 2023 13:41:07	U	SMPL
O2270-25MSDLX2 - 1		41239	7.0629	-	7471B	PPB	20 Apr 2023 13:43:27	U	SMPL
O2270-25MSDDLX2 - 1		42907	7.3491	-	7471B	PPB	20 Apr 2023 13:45:47	U	SMPL
O2270-25LDLX10 - 1		5040	0.8503	-	7471B	PPB	20 Apr 2023 13:48:06	U	SMPL
O2270-25ADLX2 - 1		42154	7.2199	-	7471B	PPB	20 Apr 2023 13:50:25	U	SMPL
O2407-02LX5 - 1		873	0.1352	-	7471B	PPB	20 Apr 2023 13:57:51	U	SMPL
CCV21 - 1	CCV21	30791	5.2698	-	7471B	PPB	20 Apr 2023 14:00:07	U	SMPL
CCB21 - 1	CCB21	-188	-0.0469	-	7471B	PPB	20 Apr 2023 14:02:24	U	SMPL