



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125065

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	04/20/2023	11:30
2	O2403-01		1	14	100.00	>212.0	>212.0	04/20/2023	12:00
3	O2403-01DUP		1	13	100.00	>212.0	>212.0	04/20/2023	12:30
4	O2403-07		1	14	100.00	>212.0	>212.0	04/20/2023	13:00
5	O2403-13		1	12	100.00	>212.0	>212.0	04/20/2023	13:30
6	O2415-01		1	12	100.00	>212.0	>212.0	04/20/2023	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-4-19

WorkList ID : 169184

Department : Wet-Chemistry

Date : 04-19-2023 08:22:52

4825065

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/25/2023	Water	O2403-01 L	Flash Point	Cool 4 deg C	PSEG03	M11	COMP-1	04/18/2023	1010B
04/25/2023	Water	O2403-07 L	Flash Point	Cool 4 deg C	PSEG03	M11	COMP-2	04/18/2023	1010B
04/25/2023	Water	O2403-13 L	Flash Point	Cool 4 deg C	PSEG03	M11	COMP-3	04/18/2023	1010B
04/24/2023	Water	O2415-01 T	Flash Point	Cool 4 deg C	SOUN03	M11	SAMPLE-1	04/19/2023	1010B

Date/Time 04/20/2023 11:26
 Raw Sample Received by: Ray Cury
 Raw Sample Relinquished by: 70100C1

Date/Time 04/20/2023 14:30
 Raw Sample Received by: 70100C1
 Raw Sample Relinquished by: Ray Cury