

LB125071


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55612



A= 0.0000e+000

B= 1.8306e-004 slope

C= -4.6602e-002 y-intercept

Rho= 0.9990313

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RD
0.00	0.000	-0.000	-0.000	254	0.000	254					
0.05	0.050					438	0				
0.20	0.200	0.198	-0.002	1335	0.0 %	1335					
2.50	2.500	2.509	0.009	13962	0.0 %	13962					
5.00	5.000	5.185	0.185	28580	0.0 %	28580					
7.50	7.500	7.174	-0.326	39447	0.0 %	39447					
10.0	10.000	10.134	0.134	55612	0.0 %	55612					

1110441

LB125071

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	254	-		0 SFAM01.1	PPB	4/20/2023 16:05	Std
	0.2 S01	1335	-		0.2 SFAM01.1	PPB	4/20/2023 16:12	Std
	2.5 S02	13962	-		2.5 SFAM01.1	PPB	4/20/2023 16:25	Std
	5 S03	28580	-		5 SFAM01.1	PPB	4/20/2023 16:27	Std
	7.5 S04	39447	-		7.5 SFAM01.1	PPB	4/20/2023 16:32	Std
	10 S05	55612	-		10 SFAM01.1	PPB	4/20/2023 16:35	Std
ICV034	ICV034	21531	3.8948	-	SFAM01.1	PPB	4/20/2023 16:37	SMPL
ICB034	ICB034	453	0.0363	-	SFAM01.1	PPB	4/20/2023 16:39	SMPL
CCV021	CCV021	28291	5.1322	-	SFAM01.1	PPB	4/20/2023 16:42	SMPL
CCB021	CCB021	473	0.04	-	SFAM01.1	PPB	4/20/2023 16:44	SMPL
O2387-06DLX100		314838	HIGH	-	SFAM01.1	PPB	4/20/2023 17:07	
O2387-07DLX2		33778	6.1367	-	SFAM01.1	PPB	4/20/2023 17:12	SMPL
O2387-09DL2X25		20617	3.7275	-	SFAM01.1	PPB	4/20/2023 17:14	SMPL
O2387-10DL2X50		16217	2.922	-	SFAM01.1	PPB	4/20/2023 17:16	SMPL
O2387-11DL2X50		16031	2.888	-	SFAM01.1	PPB	4/20/2023 17:18	SMPL
O2387-12DL2X50		16459	2.9663	-	SFAM01.1	PPB	4/20/2023 17:21	SMPL
O2387-06DL2X1000		35637	6.477	-	SFAM01.1	PPB	4/20/2023 17:23	SMPL
CCV022	CCV022	27516	4.9904	-	SFAM01.1	PPB	4/20/2023 17:25	SMPL
CCB022	CCB022	413	0.029	-	SFAM01.1	PPB	4/20/2023 17:28	SMPL