



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB125076

Slope : 98.6

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/20/2023	14:21
2	CAL2	1	Water	NA	NA	20.7	7.00	04/20/2023	14:22
3	CAL3	1	Water	NA	NA	20.7	10.02	04/20/2023	14:24
4	ICV	1	Water	NA	NA	20.8	6.99	04/20/2023	14:25
5	CCV1	1	Water	NA	NA	20.7	2.01	04/20/2023	14:30
6	O2423-04	1	Solid	20.02	20	21.4	10.44	04/20/2023	14:35
7	O2423-08	1	Solid	20.01	20	21.1	10.79	04/20/2023	14:37
8	O2423-12	1	Solid	20.03	20	20.2	8.15	04/20/2023	14:40
9	O2423-16	1	Solid	20.05	20	20.7	7.87	04/20/2023	14:45
10	O2425-01	1	Solid	20.02	20	22.1	7.24	04/20/2023	14:50
11	O2425-02	1	Solid	20.03	20	21.0	6.78	04/20/2023	15:00
12	O2425-02DUP	1	Solid	20.04	20	21.1	6.79	04/20/2023	15:01
13	CCV2	1	Water	NA	NA	20.7	12.02	04/20/2023	15:05

VB125076

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 02432			WorkList ID : 169252		Department : Wet-Chemistry		Date : 04-20-2023 14:13:52		
Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/18/2023	Solid	O2423-04	Corrosivity	Cool 4 deg C	ENVO01	M11	SPD-1WC-23-04-18	04/18/2023	9045D
04/18/2023	Solid	O2423-08	Corrosivity	Cool 4 deg C	ENVO01	M11	SPD-2WC-23-04-18	04/18/2023	9045D
04/18/2023	Solid	O2423-12	Corrosivity	Cool 4 deg C	ENVO01	M11	SPE-1WC-23-04-18	04/18/2023	9045D
04/18/2023	Solid	O2423-16	Corrosivity	Cool 4 deg C	ENVO01	M11	SPE-2WC-23-04-18	04/18/2023	9045D
04/19/2023	Solid	O2425-01	Corrosivity	Cool 4 deg C	PSEG03	M11	3002-3010	04/19/2023	9045D
04/19/2023	Solid	O2425-02	Corrosivity	Cool 4 deg C	PSEG03	M11	2989-2999	04/19/2023	9045D

Date/Time 04/20/23 16:20
 Raw Sample Received by: J.C. (sm)
 Raw Sample Relinquished by: J.C. (sm)

Date/Time 04/20/23 16:00
 Raw Sample Received by: J.C. (sm)
 Raw Sample Relinquished by: J.C. (sm)