



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125085

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/21/2023	09:25
2	CAL2	1	Water	NA	NA	20.7	7.00	04/21/2023	09:26
3	CAL3	1	Water	NA	NA	20.7	10.03	04/21/2023	09:29
4	ICV	1	Water	NA	NA	20.8	6.99	04/21/2023	09:30
5	CCV1	1	Water	NA	NA	20.6	2.01	04/21/2023	09:33
6	O2428-01	1	Water	NA	NA	20.9	6.20	04/21/2023	09:35
7	O2428-01DUP	1	Water	NA	NA	20.8	6.21	04/21/2023	09:36
8	O2431-03	1	Water	NA	NA	20.6	7.01	04/21/2023	09:40
9	CCV2	1	Water	NA	NA	20.7	12.02	04/21/2023	09:42

UB125085

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o2428

WorkList ID : 169277

Department : Wet-Chemistry

Date : 04-21-2023 09:14:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/20/2023	Water	O2428-01	pH	+	Cool 4 deg C	PSEG03 M11	NWB-1911	04/20/2023	9040C
04/20/2023	Water	O2431-03	pH	+	Cool 4 deg C	PSEG03 M11	AUD-1972-1975	04/20/2023	9040C

Date/Time04-21-2309:20

Raw Sample Received by: JPC

Raw Sample Relinquished by: J.C. (Son)

Date/Time04-21-2311:00

Raw Sample Received by: J.C. (Son)

Raw Sample Relinquished by: J.C. (Son)