

LB125086


7471B

INSTRUMENT ID:

Linear

 μ Abs.:

64827



A= 0.0000e+000

B= 1.5319e-004

C= 3.1799e-002

Rho= 0.9999095

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.027	0.027	-30	0.000	-30					1
0.2	0.200	0.216	0.016	1201	0.0 %	1201					8
2.5	2.500	2.430	-0.070	15658	0.0 %	15658					-3
5.0	5.000	4.980	-0.020	32301	0.0 %	32301					0
7.5	7.500	7.584	0.084	49304	0.0 %	49304					1
10.0	10.000	9.962	-0.038	64827	0.0 %	64827					0

LB125086 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Apr 2023 12:49:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-30	-	0.0000	7471B	PPB	21 Apr 2023 13:26:40	S	Std
0.2 - 1	50.2	1201	-	0.2000	7471B	PPB	21 Apr 2023 13:28:58	S	Std
2.5 - 1	52.5	15658	-	2.5000	7471B	PPB	21 Apr 2023 13:31:15	S	Std
5.0 - 1	55	32301	-	5.0000	7471B	PPB	21 Apr 2023 13:33:33	S	Std
7.5 - 1	57.5	49304	-	7.5000	7471B	PPB	21 Apr 2023 13:35:50	S	Std
10.0 - 1	510	64827	-	10.0000	7471B	PPB	21 Apr 2023 13:38:08	S	Std
ICV35 - 1	ICV35	27633	4.2648	-	7471B	PPB	21 Apr 2023 13:40:54	U	SMPL
ICB35 - 1	ICB35	-273	-0.0100	-	7471B	PPB	21 Apr 2023 13:43:10	U	SMPL
CCV23 - 1	CCV23	32971	5.0825	-	7471B	PPB	21 Apr 2023 13:45:27	U	SMPL
CCB23 - 1	CCB23	-224	-0.0025	-	7471B	PPB	21 Apr 2023 13:47:42	U	SMPL
CRA - 1	CRA	1276	0.2273	-	7471B	PPB	21 Apr 2023 13:50:00	U	SMPL
HighStd - 1	HighStd	64590	9.9260	-	7471B	PPB	21 Apr 2023 13:52:15	U	SMPL
ChkStd - 1	ChkStd	45896	7.0624	-	7471B	PPB	21 Apr 2023 13:54:31	U	SMPL
PB152313BL - 1	PBS	-321	-0.0174	-	7471B	PPB	21 Apr 2023 13:56:50	U	SMPL
PB152313BS - 1	LCSS	23378	3.6130	-	7471B	PPB	21 Apr 2023 13:59:08	U	SMPL
O2444-01 - 1	WASTE	2525	0.4186	-	7471B	PPB	21 Apr 2023 14:06:17	U	SMPL
O2444-01DUP - 1	WASTEDUP	2342	0.3906	-	7471B	PPB	21 Apr 2023 14:08:34	U	SMPL
O2444-01MS - 1	WASTEMS	23668	3.6574	-	7471B	PPB	21 Apr 2023 14:10:51	U	SMPL
O2444-01MSD - 1	WASTEMSD	22946	3.5468	-	7471B	PPB	21 Apr 2023 14:13:08	U	SMPL
CCV24 - 1	CCV24	31319	4.8294	-	7471B	PPB	21 Apr 2023 14:15:26	U	SMPL
CCB24 - 1	CCB24	-72	0.0208	-	7471B	PPB	21 Apr 2023 14:17:43	U	SMPL
O2444-01LX5 - 1		456	0.1017	-	7471B	PPB	21 Apr 2023 14:20:01	U	SMPL
O2444-01A - 1		27534	4.2496	-	7471B	PPB	21 Apr 2023 14:22:16	U	SMPL
CCV25 - 1	CCV25	32539	5.0163	-	7471B	PPB	21 Apr 2023 14:24:31	U	SMPL
CCB25 - 1	CCB25	-36	0.0263	-	7471B	PPB	21 Apr 2023 14:26:49	U	SMPL