



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 4/24/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 14:10

In Date: 04/21/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:10

Out Date: 04/22/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB125087

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2435-01	EYXH6	1	1.14	8.61	9.75	8.83	89.3	
O2435-02	EYXH7	2	1.19	8.61	9.8	8.14	80.7	
O2435-03	EYXH8	3	1.11	8.76	9.87	8.27	81.7	
O2435-04	EYXH9	4	1.15	8.79	9.94	8.42	82.7	
O2435-05	EYXJ0	5	1.15	8.81	9.96	8.01	77.9	
O2435-06	EYXJ1	6	1.19	8.75	9.94	9.19	91.4	
O2435-07	EYXJ2	7	1.17	8.75	9.92	8.43	83.0	
O2435-08	EYXJ3	8	1.18	8.32	9.5	7.62	77.4	
O2435-09	EYXJ4	9	1.15	8.81	9.96	9.22	91.6	
O2435-10	EYXJ5	10	1.19	8.50	9.69	8.47	85.6	
O2435-11	EYXJ6	11	1.12	8.46	9.58	7.89	80.0	
O2435-12	EYXJ7	12	1.15	8.81	9.96	8.22	80.2	
O2435-13	EYXJ7D	13	1.15	8.81	9.96	8.22	80.2	
O2435-14	EYXJ7S	14	1.15	8.81	9.96	8.22	80.2	
O2435-15	EYXJ8	15	1.11	8.55	9.66	7.98	80.4	
O2435-16	EYXJ9	16	1.14	8.83	9.97	8.54	83.8	
O2435-17	EYXK0	17	1.17	8.46	9.63	7.84	78.8	
O2435-18	EYXK1	18	1.12	8.44	9.56	8.56	88.2	
O2435-19	EYXK2	19	1.11	8.40	9.51	8.55	88.6	
O2435-20	EYXK6	20	1.18	8.72	9.9	8.48	83.7	
O2435-21	EYXK7	21	1.18	8.78	9.96	7.98	77.4	
O2435-22	EYXK8	22	1.17	8.48	9.65	9.14	94.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$