

LB125090


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51248



A= 0.0000e+000

B= 1.9431e-004 slope

C= -4.0023e-002 y-intercept

Rho= 0.9996537

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.033	-0.033	36	0.000	36					1
0.05	0.050					438	0				2
0.20	0.200	0.197	-0.003	1221	0.0 %	1221					3
2.50	2.500	2.421	-0.079	12665	0.0 %	12665					4
5.00	5.000	5.206	0.206	26998	0.0 %	26998					0
7.50	7.500	7.491	-0.009	38758	0.0 %	38758					1
10.0	10.000	9.918	-0.082	51248	0.0 %	51248					1

LB125090

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	36	-		0 SFAM01.1	PPB	4/21/2023 14:49	Std
	0.2 S01	1221	-		0.2 SFAM01.1	PPB	4/21/2023 14:51	Std
	2.5 S02	12665	-		2.5 SFAM01.1	PPB	4/21/2023 14:54	Std
	5 S03	26998	-		5 SFAM01.1	PPB	4/21/2023 14:56	Std
	7.5 S04	38758	-		7.5 SFAM01.1	PPB	4/21/2023 15:00	Std
	10 S05	51248	-		10 SFAM01.1	PPB	4/21/2023 15:24	Std
ICV036	ICV036	21813	4.1984 -		SFAM01.1	PPB	4/21/2023 15:27	SMPL
ICB036	ICB036	-197	-0.0783 -		SFAM01.1	PPB	4/21/2023 15:29	SMPL
CCV026	CCV026	26151	5.0414 -		SFAM01.1	PPB	4/21/2023 15:32	SMPL
CCB026	CCB026	-196	-0.0781 -		SFAM01.1	PPB	4/21/2023 15:34	SMPL
PB152315BL	PBW315	-14	-0.0427 -		SFAM01.1	PPB	4/21/2023 15:36	SMPL
O2228-07	X3F58	19425	3.7344 -		SFAM01.1	PPB	4/21/2023 15:38	SMPL
O2228-08	X3F58D	19334	3.7168 -		SFAM01.1	PPB	4/21/2023 15:43	SMPL
O2228-09	X3F58S	23696	4.5643 -		SFAM01.1	PPB	4/21/2023 15:52	SMPL
CCV027	CCV027	26187	5.0484 -		SFAM01.1	PPB	4/21/2023 15:55	SMPL
CCB027	CCB027	-138	-0.0668 -		SFAM01.1	PPB	4/21/2023 15:57	SMPL