

Analysis Method: SM2510 BANALYST: IwonaParameter: ConductanceSUPERVISOR REVIEW BY: SohilRun Number: LB125093Meter ID: WC pH Meter-1Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP101736

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB125093BL		1	20.2	0.83	0.91	04/20/2023	15:30
2	LB125093BS	1413	1	20.5	1301.00	1423.34	04/20/2023	15:34
3	O2413-01		1	19.7	2.81	3.13	04/20/2023	15:38
4	O2413-01DUP		1	19.7	2.79	3.10	04/20/2023	15:42

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

WORKLIST(Hardcopy Internal Chain)

LB1250P3

WorkList Name : CONDUCTANCE-042023

WorkList ID : 169247

Department : Wet-Chemistry

Date : 04-20-2023 12:25:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/25/2023	Water	O2413-01	Conductance	Cool 4 deg C	STER17	M11	JS9600	03/28/2023	SM2510 B

Date/Time 04/20/23 15:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 04/20/23 16:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]