

SFAM01.1


LB125096

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54685



A= 0.0000e+000

B= 1.8403e-004 slope

C= 2.9051e-002 intercept

Rho= 0.9998574

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.00	0.000	0.047	0.047	96	0.000	96					0/6)
0.05	0.050					438					=
0.20	0.200	0.220	0.020	1040	0.0 %	1040	0				10
2.50	2.500	2.461	-0.039	13214	0.0 %	13214					-2
5.00	5.000	4.980	-0.020	26906	0.0 %	26906					0
7.50	7.500	7.399	-0.101	40048	0.0 %	40048					-1
10.0	10.000	10.093	0.093	54685	0.0 %	54685					1

LB125096

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	96	-		0 SFAM01.1	PPB	4/21/2023 16:19	Std
	0.2 S01	1040	-		0.2 SFAM01.1	PPB	4/21/2023 16:21	Std
	2.5 S02	13214	-		2.5 SFAM01.1	PPB	4/21/2023 16:23	Std
	5 S03	26906	-		5 SFAM01.1	PPB	4/21/2023 16:26	Std
	7.5 S04	40048	-		7.5 SFAM01.1	PPB	4/21/2023 16:28	Std
	10 S05	54685	-		10 SFAM01.1	PPB	4/21/2023 16:32	Std
ICV037	ICV037	23201	4.2987 -		SFAM01.1	PPB	4/21/2023 16:35	SMPL
ICB037	ICB037	-67	0.0167 -		SFAM01.1	PPB	4/21/2023 16:37	SMPL
CCV028	CCV028	27013	5.0002 -		SFAM01.1	PPB	4/21/2023 16:39	SMPL
CCB028	CCB028	-61	0.0178 -		SFAM01.1	PPB	4/21/2023 16:41	SMPL
PB152270BL	PBS270	53	0.0388 -		SFAM01.1	PPB	4/21/2023 16:44	SMPL
O2228-19	X3F62	20874	3.8704 -		SFAM01.1	PPB	4/21/2023 16:48	SMPL
O2228-20	X3F62D	21845	4.0491 -		SFAM01.1	PPB	4/21/2023 16:50	SMPL
O2228-21	X3F62S	34212	6.325 -		SFAM01.1	PPB	4/21/2023 16:57	SMPL
CCV029	CCV029	25674	4.7538 -		SFAM01.1	PPB	4/21/2023 17:00	SMPL
CCB029	CCB029	-1	0.0289 -		SFAM01.1	PPB	4/21/2023 17:02	SMPL
CCB029	CCB029	-1	0.0289 -		SFAM01.1	PPB	4/21/2023 17:02	SMPL