



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125101

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/21/2023	17:00
2	CAL2	1	Water	NA	NA	20.7	7.01	04/21/2023	17:01
3	CAL3	1	Water	NA	NA	20.7	10.03	04/21/2023	17:02
4	ICV	1	Water	NA	NA	20.7	6.99	04/21/2023	17:05
5	CCV1	1	Water	NA	NA	20.7	2.01	04/21/2023	17:10
6	O2469-01	1	Water	NA	NA	20.1	6.78	04/21/2023	18:00
7	O2469-01DUP	1	Water	NA	NA	20.2	6.79	04/21/2023	18:01
8	O2469-02	1	Water	NA	NA	20.2	6.69	04/21/2023	18:05
9	O2469-03	1	Water	NA	NA	20.2	7.02	04/21/2023	18:14
10	CCV2	1	Water	NA	NA	20.8	12.01	04/21/2023	18:15

VB 125101

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 02469 WorkList ID : 169298 Department : Wet-Chemistry Date : 04-21-2023 16:48:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/20/2023	Water	O2469-01	pH	Cool 4 deg C	LIRO01	M11	TWP-02	04/20/2023	9040C
04/19/2023	Water	O2469-02	pH	Cool 4 deg C	LIRO01	M11	TWP-15	04/19/2023	9040C
04/20/2023	Water	O2469-03	pH	Cool 4 deg C	LIRO01	M11	TWP-24	04/20/2023	9040C

Date/Time 04-21-23 16:50
Raw Sample Received by: JD (CSC)
Raw Sample Relinquished by: JD (CSC)

Date/Time 04-21-23 19:00
Raw Sample Received by: JD (CSC)
Raw Sample Relinquished by: JD (CSC)