

LB125106


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5604E



A= 0.0000e+000

B= 1.8006e-004

C= -1.2375e-002

Rho= 0.9997021

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.0	0.000	0.031	0.031	240	0.000	240					1
0.2	0.200	0.197	-0.003	1165	0.0 %	1165					-2
2.5	2.500	2.454	-0.046	13697	0.0 %	13697					-2
5.0	5.000	5.106	0.106	28426	0.0 %	28426					2
7.5	7.500	7.332	-0.168	40791	0.0 %	40791					-2
10.0	10.000	10.079	0.079	56046	0.0 %	56046					1

LB125106 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Apr 2023 08:53:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.2	240	-	0.0000	7471B	PPB	22 Apr 2023 09:17:37	S	Std
0.2 - 1	50.2	1165	-	0.2000	7471B	PPB	22 Apr 2023 09:19:55	S	Std
2.5 - 1	50.2	13697	-	2.5000	7471B	PPB	22 Apr 2023 09:22:12	S	Std
5.0 - 1	50.2	28426	-	5.0000	7471B	PPB	22 Apr 2023 09:24:29	S	Std
7.5 - 1	50.2	40791	-	7.5000	7471B	PPB	22 Apr 2023 09:26:47	S	Std
10.0 - 1	50.2	56046	-	10.0000	7471B	PPB	22 Apr 2023 09:29:05	S	Std
ICV38 - 1	ICV38	22253	3.9945	-	7471B	PPB	22 Apr 2023 09:31:56	U	SMPL
ICB38 - 1	ICB38	-98	-0.0300	-	7471B	PPB	22 Apr 2023 09:34:11	U	SMPL
CCV30 - 1	CCV30	28636	5.1438	-	7471B	PPB	22 Apr 2023 09:36:28	U	SMPL
CCB30 - 1	CCB30	-97	-0.0298	-	7471B	PPB	22 Apr 2023 09:38:43	U	SMPL
CRA - 1	CRA	1177	0.1996	-	7471B	PPB	22 Apr 2023 09:41:01	U	SMPL
HighStd - 1	HighStd	56227	10.1119	-	7471B	PPB	22 Apr 2023 09:43:17	U	SMPL
ChkStd - 1	ChkStd	38499	6.9198	-	7471B	PPB	22 Apr 2023 09:45:32	U	SMPL
PB152274BL - 1	PBS	28	-0.0073	-	7471B	PPB	22 Apr 2023 09:47:51	U	SMPL
PB152274BS - 1	LCSS	22580	4.0534	-	7471B	PPB	22 Apr 2023 09:50:10	U	SMPL
O2386-03DLX25 - 1	SPA-1WC-23-04	518689	HIGH	-	7471B	PPB	22 Apr 2023 09:52:26	X	
O2386-07DLX25 - 1	SPA-2WC-23-04	552677	HIGH	-	7471B	PPB	22 Apr 2023 09:54:18	X	
O2386-11DLX10 - 1	SPA-3WC-23-04	433	0.0656	-	7471B	PPB	22 Apr 2023 10:00:32	U	SMPL
O2386-15DLX25 - 1	SPA-4WC-23-04	841664	HIGH	-	7471B	PPB	22 Apr 2023 10:02:49	X	
O2386-19DLX25 - 1	SPA-5WC-23-04	874354	HIGH	-	7471B	PPB	22 Apr 2023 10:04:37	X	
CCV31 - 1	CCV31	26827	4.8181	-	7471B	PPB	22 Apr 2023 10:07:02	U	SMPL
CCB31 - 1	CCB31	147	0.0141	-	7471B	PPB	22 Apr 2023 10:13:28	U	SMPL
O2389-03DLX25 - 1	SPB-1WC-23-04-17	688626	HIGH	-	7471B	PPB	22 Apr 2023 10:15:43	X	
O2389-07DLX25 - 1	SPB-2WC-23-04-17	874388	HIGH	-	7471B	PPB	22 Apr 2023 10:17:32	X	
O2389-11DLX25 - 1	SPB-3WC-23-04-17	613317	HIGH	-	7471B	PPB	22 Apr 2023 10:19:46	X	
O2389-15DLX25 - 1	SPB-4WC-23-04-17	549096	HIGH	-	7471B	PPB	22 Apr 2023 10:22:16	X	
O2389-19DLX25 - 1	SPB-5WC-23-04-17	1100648	HIGH	-	7471B	PPB	22 Apr 2023 10:24:40	X	
O2389-23DLX25 - 1	SPB-6WC-23-04-17	497386	HIGH	-	7471B	PPB	22 Apr 2023 10:26:58	X	
O2389-27DLX25 - 1	SPB-7WC-23-04-17	983548	HIGH	-	7471B	PPB	22 Apr 2023 10:29:52	X	
O2389-31DLX25 - 1	SPB-8WC-23-04-17	980656	HIGH	-	7471B	PPB	22 Apr 2023 10:32:14	X	
O2390-03DLX25 - 1	SPC-1WC-23-04-17	575433	HIGH	-	7471B	PPB	22 Apr 2023 10:34:49	X	
O2390-07DLX25 - 1	SPC-2WC-23-04-17	382844	HIGH	-	7471B	PPB	22 Apr 2023 10:46:05	X	
CCV32 - 1	CCV32	26164	4.6987	-	7471B	PPB	22 Apr 2023 10:48:01	U	SMPL
CCB32 - 1	CCB32	-41	-0.0198	-	7471B	PPB	22 Apr 2023 10:53:20	U	SMPL
O2390-11DLX25 - 1	SPC-3WC-23-04-17	465844	83.8679	-	7471B	PPB	22 Apr 2023 10:55:35	U	SMPL
O2472-01 - 1	FP-1	2968	0.5220	-	7471B	PPB	22 Apr 2023 10:57:51	U	SMPL
O2472-01DUP - 1	FP-1DUP	3601	0.6360	-	7471B	PPB	22 Apr 2023 11:03:34	U	SMPL
O2472-01MS - 1	FP-1MS	25304	4.5439	-	7471B	PPB	22 Apr 2023 11:05:49	U	SMPL
O2472-01MSD - 1	FP-1MSD	25454	4.5709	-	7471B	PPB	22 Apr 2023 11:08:06	U	SMPL
O2472-01LX5 - 1		501	0.0778	-	7471B	PPB	22 Apr 2023 11:10:24	U	SMPL
O2472-01A - 1		25498	4.5788	-	7471B	PPB	22 Apr 2023 11:12:43	U	SMPL
O2386-03DL2X2500 - 1		5673	1.0091	-	7471B	PPB	22 Apr 2023 11:15:00	U	SMPL
O2386-07DL2X2500 - 1		6940	1.2372	-	7471B	PPB	22 Apr 2023 11:17:19	U	SMPL
O2386-15DL2X2500 - 1		12984	2.3255	-	7471B	PPB	22 Apr 2023 11:19:36	U	SMPL
CCV33 - 1	CCV33	28524	5.1237	-	7471B	PPB	22 Apr 2023 11:21:54	U	SMPL
CCB33 - 1	CCB33	53	-0.0028	-	7471B	PPB	22 Apr 2023 11:24:11	U	SMPL
O2386-19DL2X2500 - 1		12079	2.1626	-	7471B	PPB	22 Apr 2023 11:26:28	U	SMPL
O2389-03DL2X2500 - 1		8358	1.4926	-	7471B	PPB	22 Apr 2023 11:28:44	U	SMPL
O2389-07DL2X2500 - 1		17854	3.2024	-	7471B	PPB	22 Apr 2023 11:31:01	U	SMPL
O2389-11DL2X2500 - 1		17582	3.1535	-	7471B	PPB	22 Apr 2023 11:33:19	U	SMPL
O2389-15DL2X2500 - 1		8464	1.5117	-	7471B	PPB	22 Apr 2023 11:35:37	U	SMPL
O2389-19DL2X2500 - 1		28296	5.0826	-	7471B	PPB	22 Apr 2023 11:37:56	U	SMPL
O2389-23DL2X2500 - 1		6997	1.2475	-	7471B	PPB	22 Apr 2023 11:40:13	U	SMPL
O2389-27DL2X2500 - 1		18172	3.2597	-	7471B	PPB	22 Apr 2023 11:42:33	U	SMPL
O2389-31DL2X2500 - 1		20373	3.6560	-	7471B	PPB	22 Apr 2023 11:44:52	U	SMPL
O2390-03DL2X2500 - 1		9744	1.7421	-	7471B	PPB	22 Apr 2023 11:47:10	U	SMPL
CCV34 - 1	CCV34	28803	5.1739	-	7471B	PPB	22 Apr 2023 11:49:29	U	SMPL
CCB34 - 1	CCB34	-3	-0.0129	-	7471B	PPB	22 Apr 2023 11:51:46	U	SMPL
O2390-07DL2X250 - 1		34818	6.2570	-	7471B	PPB	22 Apr 2023 11:56:14	U	SMPL
O2390-11DL2X250 - 1		49354	8.8744	-	7471B	PPB	22 Apr 2023 11:58:30	U	SMPL
CCV35 - 1	CCV35	27909	5.0129	-	7471B	PPB	22 Apr 2023 12:04:46	U	SMPL
CCB35 - 1	CCB35	14	-0.0099	-	7471B	PPB	22 Apr 2023 12:07:01	U	SMPL