



PERCENT SOLID

Supervisor: Iwona

Analyst: JIGNESH

Date: 4/24/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 14:25

In Date: 04/22/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:15

Out Date: 04/23/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB125108

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2438-01	EYXL38	1	1.18	8.65	9.83	8.36	83.0	
O2438-02	EYXL39	2	1.15	8.82	9.97	8.32	81.3	
O2438-03	EYXL40	3	1.14	8.48	9.62	8.04	81.4	
O2438-04	EYXL41	4	1.14	8.65	9.79	7.75	76.4	
O2438-05	EYXL42	5	1.11	8.46	9.57	8.35	85.6	
O2438-06	EYXL43	6	1.19	8.35	9.54	7.85	79.8	
O2438-07	EYXL44	7	1.18	8.42	9.6	8.05	81.6	
O2438-08	EYXL45	8	1.18	8.47	9.65	7.86	78.9	
O2438-09	EYXL46	9	1.19	8.50	9.69	8.14	81.8	
O2438-10	EYXL47	10	1.18	8.50	9.68	7.89	78.9	
O2438-11	EYXL48	11	1.14	8.81	9.95	8.22	80.4	
O2438-12	EYXL49	12	1.13	8.39	9.52	8.42	86.9	
O2438-13	EYXL50	13	1.19	8.45	9.64	8.18	82.7	
O2438-14	EYXL50D	14	1.19	8.45	9.64	8.18	82.7	
O2438-15	EYXL50S	15	1.19	8.45	9.64	8.18	82.7	
O2438-16	EYXL51	16	1.18	8.80	9.98	8.88	87.5	
O2438-17	EYXL52	17	1.18	8.39	9.57	7.6	76.5	
O2438-18	EYXL53	18	1.19	8.78	9.97	8.02	77.8	
O2438-19	EYXL54	19	1.15	8.78	9.93	8.2	80.3	
O2438-20	EYXL55	20	1.19	8.66	9.85	8.14	80.3	
O2438-21	EYXL56	21	1.1	8.59	9.69	7.85	78.6	
O2438-22	EYXL57	22	1.17	8.78	9.95	8.06	78.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$