

LB125110


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47428



A= 0.0000e+000

B= 2.1362e-004

C= 4.2452e-002

Rho= 0.9993748

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.048	0.048	26	0.000	26	0				20
0.2	0.200	0.240	0.040	923	0.0 %	923					-2
2.5	2.500	2.461	-0.039	11321	0.0 %	11321					-3
5.0	5.000	5.033	0.033	23364	0.0 %	23364					2
7.5	7.500	7.244	-0.256	33714	0.0 %	33714					
10.0	10.000	10.174	0.174	47428	0.0 %	47428					

LB125110 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Apr 2023 14:18:41

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7470A	PPB	22 Apr 2023 14:21:15	S	Std
0.2 - 1	50.2	923	-	0.2000	7470A	PPB	22 Apr 2023 14:23:33	S	Std
2.5 - 1	52.5	11321	-	2.5000	7470A	PPB	22 Apr 2023 14:25:50	S	Std
5.0 - 1	55	23364	-	5.0000	7470A	PPB	22 Apr 2023 14:28:07	S	Std
7.5 - 1	57.5	33714	-	7.5000	7470A	PPB	22 Apr 2023 14:30:24	S	Std
10.0 - 1	510	47428	-	10.0000	7470A	PPB	22 Apr 2023 14:38:10	S	Std
ICV39 - 1	ICV39	19183	4.1403	-	7470A	PPB	22 Apr 2023 14:57:43	U	SMPL
ICB39 - 1	ICB39	-123	0.0162	-	7470A	PPB	22 Apr 2023 14:59:58	U	SMPL
CCV36 - 1	CCV36	23335	5.0272	-	7470A	PPB	22 Apr 2023 15:02:15	U	SMPL
CCB36 - 1	CCB36	-183	0.0034	-	7470A	PPB	22 Apr 2023 15:04:33	U	SMPL
CRA - 1	CRA	871	0.2285	-	7470A	PPB	22 Apr 2023 15:06:50	U	SMPL
HighStd - 1	HighStd	46341	9.9417	-	7470A	PPB	22 Apr 2023 15:09:05	U	SMPL
ChkStd - 1	ChkStd	30853	6.6332	-	7470A	PPB	22 Apr 2023 15:11:21	U	SMPL
PB152300BL - 1	PBW	-153	0.0098	-	7470A	PPB	22 Apr 2023 15:13:40	U	SMPL
PB152300BS - 1	LCSW	16351	3.5353	-	7470A	PPB	22 Apr 2023 15:15:58	U	SMPL
O2389-04 - 1	SPB-1WC-23-04-17	17978	3.8829	-	7470A	PPB	22 Apr 2023 15:18:15	U	SMPL
O2389-08 - 1	SPB-2WC-23-04-17	16120	3.4860	-	7470A	PPB	22 Apr 2023 15:20:32	U	SMPL
O2389-12 - 1	SPB-3WC-23-04-17	20384	4.3968	-	7470A	PPB	22 Apr 2023 15:22:51	U	SMPL
O2389-16 - 1	SPB-4WC-23-04-17	15355	3.3225	-	7470A	PPB	22 Apr 2023 15:25:09	U	SMPL
O2389-20 - 1	SPB-5WC-23-04-17	42904	9.2075	-	7470A	PPB	22 Apr 2023 15:27:27	U	SMPL
CCV37 - 1	CCV37	23152	4.9881	-	7470A	PPB	22 Apr 2023 15:29:42	U	SMPL
CCB37 - 1	CCB37	-222	-0.0050	-	7470A	PPB	22 Apr 2023 15:32:00	U	SMPL
O2389-24 - 1	SPB-6WC-23-04-17	36674	7.8766	-	7470A	PPB	22 Apr 2023 15:34:17	U	SMPL
O2389-28 - 1	SPB-7WC-23-04-17	100836	21.5827	-	7470A	PPB	22 Apr 2023 15:36:32	U	SMPL
O2389-32 - 1	SPB-8WC-23-04-17	34868	7.4908	-	7470A	PPB	22 Apr 2023 15:38:51	U	SMPL
O2390-04 - 1	SPC-1WC-23-04-17	17322	3.7427	-	7470A	PPB	22 Apr 2023 15:41:11	U	SMPL
O2390-08 - 1	SPC-2WC-23-04-17	14666	3.1754	-	7470A	PPB	22 Apr 2023 15:43:29	U	SMPL
O2390-12 - 1	SPC-3WC-23-04-17	5455	1.2077	-	7470A	PPB	22 Apr 2023 15:45:47	U	SMPL
O2406-02 - 1	SB-12-COMP	-15	0.0392	-	7470A	PPB	22 Apr 2023 15:48:05	U	SMPL
O2406-04 - 1	SB-13-COMP	27	0.0482	-	7470A	PPB	22 Apr 2023 15:50:22	U	SMPL
O2406-06 - 1	SB-14-COMP	99	0.0636	-	7470A	PPB	22 Apr 2023 15:52:39	U	SMPL
O2406-08 - 1	SB-15-COMP	75	0.0585	-	7470A	PPB	22 Apr 2023 15:54:55	U	SMPL
CCV38 - 1	CCV38	21838	4.7074	-	7470A	PPB	22 Apr 2023 16:02:48	U	SMPL
CCB38 - 1	CCB38	-170	0.0061	-	7470A	PPB	22 Apr 2023 16:05:03	U	SMPL
O2407-02 - 1	NORTH-ST-BAY-AVE-INTERSECTION	6	0.0437	-	7470A	PPB	22 Apr 2023 16:07:21	U	SMPL
O2425-01 - 1	3002-3010	136	0.0715	-	7470A	PPB	22 Apr 2023 16:09:36	U	SMPL
O2425-02 - 1	2989-2999	97	0.0632	-	7470A	PPB	22 Apr 2023 16:11:52	U	SMPL
O2425-02DUP - 1	2989-2999DUP	142	0.0728	-	7470A	PPB	22 Apr 2023 16:14:08	U	SMPL
O2425-02MS - 1	2989-2999MS	17948	3.8764	-	7470A	PPB	22 Apr 2023 16:16:23	U	SMPL
O2425-02MSD - 1	2989-2999MSD	17021	3.6784	-	7470A	PPB	22 Apr 2023 16:18:40	U	SMPL
PB152222TB - 1	PB152222TB	-156	0.0091	-	7470A	PPB	22 Apr 2023 16:20:57	U	SMPL
O2425-02LX5 - 1		37	0.0504	-	7470A	PPB	22 Apr 2023 16:23:16	U	SMPL
O2425-02A - 1		18308	3.9533	-	7470A	PPB	22 Apr 2023 16:25:33	U	SMPL
O2389-28DLX5 - 1		22269	4.7995	-	7470A	PPB	22 Apr 2023 16:27:49	U	SMPL
CCV39 - 1	CCV39	23737	5.1131	-	7470A	PPB	22 Apr 2023 16:30:06	U	SMPL
CCB39 - 1	CCB39	-134	0.0138	-	7470A	PPB	22 Apr 2023 16:32:24	U	SMPL