

7471B

INSTRUMENT ID: CV2

Linear

$$\mu \text{ Abs.} :$$

6595E

A= 0.0000e+000

B= 1.5171e-004

C= 5.6811e-002

Rho= 0.9995513

Accept=Accepted

Max. Conc:— 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%rel)
0.0	0.000	0.081	0.081	159	0.000	159					-
0.2	0.200	0.218	0.018	1064	0.0 %	1064					9
2.5	2.500	2.332	-0.168	14995	0.0 %	14995					-7
5.0	5.000	5.133	0.133	33461	0.0 %	33461					3
7.5	7.500	7.372	-0.128	48219	0.0 %	48219					-2
10.0	10.000	10.064	0.064	65958	0.0 %	65958					1

LB125117 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Apr 2023 10:27:52

Sample ID	Extended ID	μ Abs.	Conc.	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	159	-	0.0000	7471B	PPB	24 Apr 2023 11:10:41	S	Std
0.2 - 1	50-2	1064	-	0.2000	7471B	PPB	24 Apr 2023 11:13:07	S	Std
2.5 - 1	52-5	14995	-	2.5000	7471B	PPB	24 Apr 2023 11:15:24	S	Std
5.0 - 1	55	33461	-	5.0000	7471B	PPB	24 Apr 2023 11:17:42	S	Std
7.5 - 1	57-5	48219	-	7.5000	7471B	PPB	24 Apr 2023 11:19:58	S	Std
10.0 - 1	510	65958	-	10.0000	7471B	PPB	24 Apr 2023 11:22:17	S	Std
ICV40 - 1	ICV40	26589	4.0907	-7471B	PPB	PPB	24 Apr 2023 11:25:07	U	SMPL
ICB40 - 1	ICB40	-151	0.0339	-7471B	PPB	PPB	24 Apr 2023 11:27:23	U	SMPL
CCV40 - 1	CCV40	30533	4.6891	-7471B	PPB	PPB	24 Apr 2023 11:29:41	U	SMPL
CCB40 - 1	CCB40	-56	0.0483	-7471B	PPB	PPB	24 Apr 2023 11:31:56	U	SMPL
CRA - 1	CRA	1046	0.2155	-7471B	PPB	PPB	24 Apr 2023 11:34:13	U	SMPL
HighStd - 1	HighStd	65444	9.9855	-7471B	PPB	PPB	24 Apr 2023 11:36:28	U	SMPL
ChkStd - 1	ChkStd	45277	6.9259	-7471B	PPB	PPB	24 Apr 2023 11:38:43	U	SMPL
PB152339BL - 1	PB152339BL	-237	0.0209	-7471B	PPB	PPB	24 Apr 2023 11:41:04	U	SMPL
PB152339BS - 1	PB152339BS	25334	3.9003	-7471B	PPB	PPB	24 Apr 2023 11:43:23	U	SMPL
O2423-03 - 1	SPD-1WC-23-04-18	45244	6.9209	-7471B	PPB	PPB	24 Apr 2023 11:48:16	U	SMPL
O2423-07 - 1	SPD-2WC-23-04-18	102681	15.6349	-7471B	PPB	PPB	24 Apr 2023 11:50:33	U	SMPL
O2423-11 - 1	SPE-1WC-23-04-18	952536	HIGH	-7471B	PPB	PPB	24 Apr 2023 11:52:52	X	
O2423-15 - 1	SPE-2WC-23-04-18	620833	HIGH	-7471B	PPB	PPB	24 Apr 2023 11:54:46	X	
O2427-05 - 1	20230343-COMP-21	7389	1.1778	-7471B	PPB	PPB	24 Apr 2023 11:57:06	U	SMPL
CCV41 - 1	CCV41	32094	4.9259	-7471B	PPB	PPB	24 Apr 2023 11:59:55	U	SMPL
CCB41 - 1	CCB41	-272	0.0155	-7471B	PPB	PPB	24 Apr 2023 12:02:13	U	SMPL
O2427-06 - 1	20230344-COMP-22	38757	5.9368	-7471B	PPB	PPB	24 Apr 2023 12:04:30	U	SMPL
O2427-07 - 1	20230347-COMP-25	62363	9.5181	-7471B	PPB	PPB	24 Apr 2023 12:06:46	U	SMPL
O2427-08 - 1	20230349-COMP-27	29391	4.5158	-7471B	PPB	PPB	24 Apr 2023 12:09:04	U	SMPL
O2427-09 - 1	20230350-COMP-28	53268	8.1383	-7471B	PPB	PPB	24 Apr 2023 12:11:23	U	SMPL
O2427-10 - 1	20230351-COMP-29	53366	8.1531	-7471B	PPB	PPB	24 Apr 2023 12:13:42	U	SMPL
O2427-11 - 1	20230352-COMP-30	45156	6.9076	-7471B	PPB	PPB	24 Apr 2023 12:16:01	U	SMPL
O2427-12 - 1	20230353-COMP-31	38621	5.9161	-7471B	PPB	PPB	24 Apr 2023 12:18:21	U	SMPL
O2427-12DUP - 1	20230353-COMP-31DUP	41465	6.3476	-7471B	PPB	PPB	24 Apr 2023 12:20:41	U	SMPL
O2427-13 - 1	O2427-12MS	61010	9.3128	-7471B	PPB	PPB	24 Apr 2023 12:23:00	U	SMPL
O2427-14 - 1	O2427-12MSD	60369	9.2156	-7471B	PPB	PPB	24 Apr 2023 12:29:06	U	SMPL
CCV42 - 1	CCV42	32618	5.0054	-7471B	PPB	PPB	24 Apr 2023 12:31:22	U	SMPL
CCB42 - 1	CCB42	-128	0.0374	-7471B	PPB	PPB	24 Apr 2023 12:33:41	U	SMPL
PB152340BL - 1	PB152340BL	51	0.0645	-7471B	PPB	PPB	24 Apr 2023 12:35:59	U	SMPL
PB152340BS - 1	PB152340BS	25784	3.9686	-7471B	PPB	PPB	24 Apr 2023 12:38:15	U	SMPL
O2431-01 - 1	AUD-1943	2067	0.3704	-7471B	PPB	PPB	24 Apr 2023 12:40:31	U	SMPL
O2445-01 - 1	TRE-1715	1038	0.2143	-7471B	PPB	PPB	24 Apr 2023 12:42:48	U	SMPL
O2445-02 - 1	TRE-1730-1731	871	0.1890	-7471B	PPB	PPB	24 Apr 2023 12:45:05	U	SMPL
O2446-01 - 1	OR-03-042023	2517	0.4387	-7471B	PPB	PPB	24 Apr 2023 12:47:22	U	SMPL
O2448-02 - 1	20230324-COMP-2	23112	3.5632	-7471B	PPB	PPB	24 Apr 2023 12:49:39	U	SMPL
O2448-02DUP - 1	20230324-COMP-2DUP	23771	3.6632	-7471B	PPB	PPB	24 Apr 2023 12:51:56	U	SMPL
O2448-03 - 1	O2448-02MS	49171	7.5167	-7471B	PPB	PPB	24 Apr 2023 12:59:18	U	SMPL
O2448-04 - 1	O2448-02MSD	49053	7.4988	-7471B	PPB	PPB	24 Apr 2023 13:01:34	U	SMPL
CCV43 - 1	CCV43	32647	5.0098	-7471B	PPB	PPB	24 Apr 2023 13:03:53	U	SMPL
CCB43 - 1	CCB43	-172	0.0307	-7471B	PPB	PPB	24 Apr 2023 13:06:11	U	SMPL
O2448-05 - 1	20230336-COMP-14	14714	2.2891	-7471B	PPB	PPB	24 Apr 2023 13:08:28	U	SMPL
O2448-06 - 1	20230340-COMP-18	41578	6.3647	-7471B	PPB	PPB	24 Apr 2023 13:10:44	U	SMPL
O2448-07 - 1	20230345-COMP-23	57797	8.8254	-7471B	PPB	PPB	24 Apr 2023 13:13:02	U	SMPL
O2448-08 - 1	20230346-COMP-24	78245	11.9276	-7471B	PPB	PPB	24 Apr 2023 13:15:20	U	SMPL
O2448-09 - 1	20230354-COMP-32	34203	5.2459	-7471B	PPB	PPB	24 Apr 2023 13:17:39	U	SMPL
O2464-01 - 1	MH-91	4218	0.6967	-7471B	PPB	PPB	24 Apr 2023 13:20:01	U	SMPL
O2464-03 - 1	MH-92	3760	0.6273	-7471B	PPB	PPB	24 Apr 2023 13:22:20	U	SMPL
O2467-02 - 1	279	37172	5.6963	-7471B	PPB	PPB	24 Apr 2023 13:24:37	U	SMPL
O2386-11 - 1		5248	0.8530	-7471B	PPB	PPB	24 Apr 2023 13:26:54	U	SMPL
CCV44 - 1	CCV44	32472	4.9832	-7471B	PPB	PPB	24 Apr 2023 13:29:13	U	SMPL
CCB44 - 1	CCB44	-121	0.0385	-7471B	PPB	PPB	24 Apr 2023 13:31:28	U	SMPL
O2427-12LX5 - 1		7623	1.2133	-7471B	PPB	PPB	24 Apr 2023 13:33:46	U	SMPL
O2427-12A - 1		61616	9.4048	-7471B	PPB	PPB	24 Apr 2023 13:36:02	U	SMPL
O2448-02LX5 - 1		3687	0.6162	-7471B	PPB	PPB	24 Apr 2023 13:38:17	U	SMPL
O2448-02A - 1		49775	7.6083	-7471B	PPB	PPB	24 Apr 2023 13:40:37	U	SMPL
O2423-07DLX5 - 1		20189	3.1198	-7471B	PPB	PPB	24 Apr 2023 13:45:39	U	SMPL
O2423-11DLX10 - 1		162438	24.7008	-7471B	PPB	PPB	24 Apr 2023 13:47:55	U	SMPL
O2423-15DLX10 - 1		125420	19.0847	-7471B	PPB	PPB	24 Apr 2023 13:50:13	U	SMPL
O2423-11DL2X50 - 1		31871	4.8921	-7471B	PPB	PPB	24 Apr 2023 13:56:10	U	SMPL
O2423-15DL2X50 - 1		27104	4.1688	-7471B	PPB	PPB	24 Apr 2023 13:58:29	U	SMPL
CCV45 - 1	CCV45	33185	5.0914	-7471B	PPB	PPB	24 Apr 2023 14:00:49	U	SMPL
CCB45 - 1	CCB45	-207	0.0254	-7471B	PPB	PPB	24 Apr 2023 14:03:07	U	SMPL