

Analysis Method: SM2510 BANALYST: IwonaParameter: ConductanceSUPERVISOR REVIEW BY: SohilRun Number: LB125120Meter ID: WC pH Meter-1Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP101736

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB125120BL		1	20.8	0.84	0.91	04/20/2023	15:25
2	LB125120BS	1413	1	20.7	1292.00	1407.61	04/20/2023	15:28
3	O2413-01		1	19.7	2.81	3.13	04/20/2023	15:34
4	O2413-01DUP		1	19.8	2.82	3.13	04/20/2023	15:38

Note : Conductivity result is corrected to 25°C

Result = <u> </u> (Reading) 1 + 0.0191 (Temperature - 25)

WORKLIST(Hardcopy Internal Chain)

LB125120

WorkList Name : CONDUCTANCE-042023 WorkList ID : 169247 Department : Wet-Chemistry Date : 04-20-2023 12:25:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/25/2023	Water	O2413-01	Conductance	Cool 4 deg C	STER17	M11	JS9600	03/28/2023	SM2510 B

Date/Time 04/20/23 15:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 04/20/23 16:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]