



PERCENT SOLID

Supervisor: Sohil

Analyst: JIGNESH

Date: 4/26/2023

OVENTEMP IN Celsius(°C): 107

Time IN: 16:10

In Date: 04/24/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:25

Out Date: 04/25/2023

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB125129

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2447-01	EW905	1	1.14	8.77	9.91	9.06	90.3	
O2447-02	EW907	2	1.14	8.48	9.62	8.25	83.8	
O2447-03	EW908	3	0.01	0.01	0.02	0.02	100.0	
O2447-04	EW909	4	1.12	8.49	9.61	8.45	86.3	
O2447-05	EW910	5	1.13	8.75	9.88	6.83	65.1	
O2447-06	EW911	6	1.15	8.75	9.9	8.57	84.8	
O2447-07	EW912	7	1.13	8.65	9.78	8.2	81.7	
O2447-08	EW913	8	1.13	8.71	9.84	8.21	81.3	
O2447-09	EW915	9	1.18	8.64	9.82	8.49	84.6	
O2447-10	EW916	10	1.17	8.82	9.99	8.64	84.7	
O2447-11	EW917	11	1.13	8.64	9.77	8.5	85.3	
O2447-12	VHBLK001	12	1.00	1.00	2.00	2.00	100.0	VHBLK
O2447-13	EW933	13	1.15	8.60	9.75	8.18	81.7	
O2447-14	EW934	14	1.19	8.46	9.65	8.49	86.3	
O2447-15	EW935	15	1.16	8.77	9.93	8.71	86.1	
O2447-16	EW936	16	1.17	8.55	9.72	8.6	86.9	
O2447-17	EW937	17	1.13	8.57	9.7	7.65	76.1	
O2447-18	EW937MS	18	1.13	8.57	9.7	7.65	76.1	
O2447-19	EW937MSD	19	1.13	8.57	9.7	7.65	76.1	
O2447-20	EW938	20	1.14	8.83	9.97	8.4	82.2	
O2447-21	EW939	21	1.13	8.45	9.58	8.1	82.5	
O2447-22	EW940	22	1.1	8.67	9.77	8.14	81.2	
O2447-23	EW941	23	1.12	8.73	9.85	8.06	79.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$