



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125135

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	04/24/2023	13:50
2	O2428-01		1	14	100.00	>212.0	>212.0	04/24/2023	14:20
3	O2428-01DUP		1	13	100.00	>212.0	>212.0	04/24/2023	14:50
4	O2431-02		1	12	100.00	>212.0	>212.0	04/24/2023	15:20
5	O2431-03		1	14	100.00	>212.0	>212.0	04/24/2023	15:50
6	O2468-01		1	14	100.00	>212.0	>212.0	04/24/2023	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6125135

WorkList Name : fp-4-24

WorkList ID : 169336

Department : Wet-Chemistry

Date : 04-24-2023 08:16:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/25/2023	Water	02428-01	Flash Point	Cool 4 deg C	PSEG03	M11	NWB-1911	04/20/2023	1010B
04/25/2023	Water	02431-02	Flash Point	Cool 4 deg C	PSEG03	M11	AUD-1979	04/20/2023	1010B
04/25/2023	Water	02431-03	Flash Point	Cool 4 deg C	PSEG03	M11	AUD-1972-1975	04/20/2023	1010B
04/28/2023	Water	02468-01	Flash Point	Cool 4 deg C	LIRO01		SUMP-PUMP-GW-SAMPLE	04/21/2023	1010B

Date/Time 04/24/2023 13:40
 Raw Sample Received by: RM avg
 Raw Sample Relinquished by: JH (COG)

Date/Time 04/24/2023 16:55
 Raw Sample Received by: JH (COG)
 Raw Sample Relinquished by: RM (avg)