



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 4/25/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 04/24/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:29
Out Date: 04/25/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125136

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2480-01	MH2849	1	1.13	8.53	9.66	9.16	94.1	
O2480-02	MH2849D	2	1.13	8.53	9.66	9.16	94.1	
O2480-03	MH2849S	3	1.13	8.53	9.66	9.16	94.1	
O2480-04	MH2850	4	1.13	8.65	9.78	9.02	91.2	
O2480-05	MH2851	5	1.15	8.82	9.97	9.54	95.1	
O2480-06	MH2852	6	1.13	8.69	9.82	9.42	95.4	
O2480-07	MH2853	7	1.13	8.39	9.52	9.11	95.1	
O2480-08	MH2854	8	1.11	8.57	9.68	9.17	94.0	
O2480-09	MH2855	9	1.19	8.51	9.7	9.06	92.5	
O2480-10	MH2856	10	1.1	8.56	9.66	9.24	95.1	
O2480-11	MH2857	11	1.19	8.56	9.75	9.14	92.9	
O2480-12	MH2858	12	1.16	8.81	9.97	9.54	95.1	
O2480-13	MH2859	13	1.17	8.46	9.63	9.04	93.0	
O2480-14	MH2860	14	1.15	8.75	9.9	9.47	95.1	
O2480-15	MH2861	15	1.16	8.61	9.77	9.04	91.5	
O2480-16	MH2862	16	1.11	8.44	9.55	8.99	93.4	
O2480-17	MH2863	17	1.13	8.64	9.77	9.12	92.5	
O2480-18	MH2864	18	1.15	8.82	9.97	9.17	90.9	
O2480-19	MH2865	19	1.13	8.56	9.69	8.66	88.0	
O2480-20	MH2866	20	1.11	8.88	9.99	8.04	78.0	
O2480-21	MH2867	21	1.13	8.60	9.73	8.23	82.6	
O2480-22	MH2868	22	1.1	8.77	9.87	8.98	89.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$