


LB 125146

245.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.3627e-004

C= -5.4851e-003

Rho= 0.9999435

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.016	0.016	89	0.000	89					
0.05	0.050	0.050	0.000	235	0.0 %	235	0				
0.2	0.200	0.214	0.014	931	0.0 %	931					
2.5	2.500	2.436	-0.064	10335	0.0 %	10335					
5.0	5.000	5.062	0.062	21446	0.0 %	21446					
7.5	7.500	7.453	-0.047	31566	0.0 %	31566					
10.0	10.000	10.019	0.019	42428	0.0 %	42428					

0.1-1.0

LB125140 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 25 Apr 2023 10:17:01

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	89	-	0.0000	245.1	PPB	25 Apr 2023 10:23:05	S	Std
0.05 - 1	50.75	235	-	0.0500	245.1	PPB	25 Apr 2023 10:25:23	S	Std
0.2 - 1	50.2	931	-	0.2000	245.1	PPB	25 Apr 2023 10:27:40	S	Std
2.5 - 1	52.5	10335	-	2.5000	245.1	PPB	25 Apr 2023 10:29:57	S	Std
5.0 - 1	55	21446	-	5.0000	245.1	PPB	25 Apr 2023 10:32:14	S	Std
7.5 - 1	57.5	31566	-	7.5000	245.1	PPB	25 Apr 2023 10:34:31	S	Std
10.0 - 1	510	42428	-	10.0000	245.1	PPB	25 Apr 2023 10:45:35	S	Std
ICV42 - 1	ICV42	16465	3.8848	-245.1		PPB	25 Apr 2023 10:50:51	U	SMPL
ICB42 - 1	ICB42	-190	-0.0504	-245.1		PPB	25 Apr 2023 10:53:06	U	SMPL
CCV51 - 1	CCV51	21245	5.0142	-245.1		PPB	25 Apr 2023 10:55:23	U	SMPL
CCB51 - 1	CCB51	-168	-0.0452	-245.1		PPB	25 Apr 2023 10:57:39	U	SMPL
CRA - 1	CRA	186	0.0385	-245.1		PPB	25 Apr 2023 11:06:27	U	SMPL
HighStd - 1	HighStd	42445	10.0232	-245.1		PPB	25 Apr 2023 11:08:42	U	SMPL
ChkStd - 1	ChkStd	29546	6.9755	-245.1		PPB	25 Apr 2023 11:10:57	U	SMPL
PB152388BL - 1	PBW	-185	-0.0492	-245.1		PPB	25 Apr 2023 11:13:16	U	SMPL
PB152388BS - 1	LCSW	15504	3.6577	-245.1		PPB	25 Apr 2023 11:18:19	U	SMPL
O2415-01 - 1	SAMPLE-1	169	0.0344	-245.1		PPB	25 Apr 2023 11:20:35	U	SMPL
O2415-01DUP - 1	SAMPLE-1DUP	238	0.0507	-245.1		PPB	25 Apr 2023 11:22:53	U	SMPL
O2415-01MS - 1	SAMPLE-1MS	17099	4.0346	-245.1		PPB	25 Apr 2023 11:27:52	U	SMPL
O2415-01MSD - 1	SAMPLE-1MSD	16989	4.0086	-245.1		PPB	25 Apr 2023 11:30:09	U	SMPL
CCV52 - 1	CCV52	21032	4.9638	-245.1		PPB	25 Apr 2023 11:32:26	U	SMPL
CCB52 - 1	CCB52	-165	-0.0445	-245.1		PPB	25 Apr 2023 11:34:42	U	SMPL
O2415-01LX5 - 1		13	-0.0024	-245.1		PPB	25 Apr 2023 11:36:59	U	SMPL
O2415-01A - 1		17623	4.1584	-245.1		PPB	25 Apr 2023 11:39:14	U	SMPL
CCV53 - 1	CCV53	20642	4.8717	-245.1		PPB	25 Apr 2023 11:41:30	U	SMPL
CCB53 - 1	CCB53	-155	-0.0421	-245.1		PPB	25 Apr 2023 11:43:47	U	SMPL