

LB125149


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

46024



A= 0.0000e+000

B= 2.1400e-004

C= 6.5339e-002

Rho= 0.9981719

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/w)
0.0	0.000	0.101	0.101	166	0.000	166					
0.2	0.200	0.240	0.040	817	0.0 %	817					
2.5	2.500	2.509	0.009	11421	0.0 %	11421					20
5.0	5.000	4.595	-0.405	21167	0.0 %	21167					0
7.5	7.500	7.840	0.340	36328	0.0 %	36328					-8
10.0	10.000	9.915	-0.085	46024	0.0 %	46024					5

71

LB125149 INSTRUMENT ID :: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Apr 2023 13:31:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	166	-	0.0000	7470A	PPB	25 Apr 2023 13:36:40	S	Std
0.2 - 1	50.2	817	-	0.2000	7470A	PPB	25 Apr 2023 13:38:58	S	Std
2.5 - 1	52.5	11421	-	2.5000	7470A	PPB	25 Apr 2023 13:41:16	S	Std
5.0 - 1	55	21167	-	5.0000	7470A	PPB	25 Apr 2023 13:43:33	S	Std
7.5 - 1	57.5	36328	-	7.5000	7470A	PPB	25 Apr 2023 13:45:50	S	Std
10.0 - 1	510	46024	-	10.0000	7470A	PPB	25 Apr 2023 13:55:57	S	Std
ICV43 - 1	ICV43	19838	4.3107	-	7470A	PPB	25 Apr 2023 14:04:04	U	SMPL
ICB43 - 1	ICB43	-173	0.0283	-	7470A	PPB	25 Apr 2023 14:06:19	U	SMPL
CCV54 - 1	CCV54	21689	4.7069	-	7470A	PPB	25 Apr 2023 14:08:36	U	SMPL
CCB54 - 1	CCB54	-145	0.0343	-	7470A	PPB	25 Apr 2023 14:10:52	U	SMPL
CRA - 1	CRA	609	0.1957	-	7470A	PPB	25 Apr 2023 14:13:08	U	SMPL
HighStd - 1	HighStd	45045	9.7051	-	7470A	PPB	25 Apr 2023 14:15:24	U	SMPL
ChkStd - 1	ChkStd	29757	6.4335	-	7470A	PPB	25 Apr 2023 14:20:47	U	SMPL
PB152397BL - 1	PBW	-141	0.0352	-	7470A	PPB	25 Apr 2023 14:23:03	U	SMPL
PB152397BS - 1	LC5W	15271	3.3334	-	7470A	PPB	25 Apr 2023 14:25:20	U	SMPL
O2423-18 - 1	EB-1	-116	0.0405	-	7470A	PPB	25 Apr 2023 14:27:37	U	SMPL
PB152346TB - 1	PB152346TB	30	0.0718	-	7470A	PPB	25 Apr 2023 14:29:55	U	SMPL
CCV55 - 1	CCV55	22258	4.8286	-	7470A	PPB	25 Apr 2023 14:32:12	U	SMPL
CCB55 - 1	CCB55	-134	0.0367	-	7470A	PPB	25 Apr 2023 14:34:29	U	SMPL