

7471B


LB125152

INSTRUMENT ID: CV2

Linear

 $\mu$  Abs.:

58313



A= 0.0000e+000

B= 1.7280e-004

C= -6.2434e-002

Rho= 0.9997773

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | -0.042 | -0.042 | 121   | 0.000      | 121   |       |       |       |       |
| 0.2    | 0.200  | 0.149  | -0.051 | 1222  | 0.0 %      | 1222  |       |       |       |       |
| 2.5    | 2.500  | 2.666  | 0.166  | 15789 | 0.0 %      | 15789 |       |       |       |       |
| 5.0    | 5.000  | 4.971  | -0.029 | 29129 | 0.0 %      | 29129 |       |       |       |       |
| 7.5    | 7.500  | 7.441  | -0.059 | 43423 | 0.0 %      | 43423 |       |       |       |       |
| 10.0   | 10.000 | 10.014 | 0.014  | 58313 | 0.0 %      | 58313 |       |       |       |       |

0/10  
1  
-26  
7  
11  
1  
0

## LB125152 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Apr 2023 15:57:54

| Sample ID        | Extended ID      | $\mu$ Abs. | Conc.   | Std Conc | Method | Units                | Date                 | Type | Type |
|------------------|------------------|------------|---------|----------|--------|----------------------|----------------------|------|------|
| 0.0 - 1          | S0               | 121        | -       | 0.0000   | 7471B  | PPB                  | 25 Apr 2023 16:01:26 | S    | Std  |
| 0.2 - 1          | S0.2             | 1222       | -       | 0.2000   | 7471B  | PPB                  | 25 Apr 2023 16:03:44 | S    | Std  |
| 2.5 - 1          | S2.5             | 15789      | -       | 2.5000   | 7471B  | PPB                  | 25 Apr 2023 16:06:02 | S    | Std  |
| 5.0 - 1          | S5               | 29129      | -       | 5.0000   | 7471B  | PPB                  | 25 Apr 2023 16:08:19 | S    | Std  |
| 7.5 - 1          | S7.5             | 43423      | -       | 7.5000   | 7471B  | PPB                  | 25 Apr 2023 16:10:37 | S    | Std  |
| 10.0 - 1         | S10              | 58313      | -       | 10.0000  | 7471B  | PPB                  | 25 Apr 2023 16:12:55 | S    | Std  |
| ICV44 - 1        | ICV44            | 25647      | 4.3695  | -7471B   | PPB    | 25 Apr 2023 16:16:38 | U                    | SMPL |      |
| ICB44 - 1        | ICB44            | 43         | -0.0550 | -7471B   | PPB    | 25 Apr 2023 16:21:26 | U                    | SMPL |      |
| CCV56 - 1        | CCV56            | 28994      | 4.9479  | -7471B   | PPB    | 25 Apr 2023 16:23:41 | U                    | SMPL |      |
| CCB56 - 1        | CCB56            | -194       | -0.0960 | -7471B   | PPB    | 25 Apr 2023 16:25:56 | U                    | SMPL |      |
| CRA - 1          | CRA              | 1193       | 0.1437  | -7471B   | PPB    | 25 Apr 2023 16:28:13 | U                    | SMPL |      |
| HighStd - 1      | HighStd          | 58319      | 10.0154 | -7471B   | PPB    | 25 Apr 2023 16:30:28 | U                    | SMPL |      |
| ChkStd - 1       | ChkStd           | 40686      | 6.9683  | -7471B   | PPB    | 25 Apr 2023 16:32:44 | U                    | SMPL |      |
| PB152378BL - 1   | PBS              | 14         | -0.0600 | -7471B   | PPB    | 25 Apr 2023 16:37:35 | U                    | SMPL |      |
| PB152378BS - 1   | LCSS             | 20991      | 3.5649  | -7471B   | PPB    | 25 Apr 2023 16:39:51 | U                    | SMPL |      |
| O2476-01 - 1     | M-041923RB01     | 46738      | 8.0141  | -7471B   | PPB    | 25 Apr 2023 16:42:07 | U                    | SMPL |      |
| O2477-01 - 1     | E-H4(ELEV-22)    | 1093       | 0.1264  | -7471B   | PPB    | 25 Apr 2023 16:54:53 | U                    | SMPL |      |
| O2477-01DUP - 1  | E-H4(ELEV-22)DUP | 798        | 0.0755  | -7471B   | PPB    | 25 Apr 2023 16:57:10 | U                    | SMPL |      |
| O2477-01MS - 1   | E-H4(ELEV-22)MS  | 24042      | 4.0921  | -7471B   | PPB    | 25 Apr 2023 16:59:27 | U                    | SMPL |      |
| O2477-01MSD - 1  | E-H4(ELEV-22)MSD | 24538      | 4.1779  | -7471B   | PPB    | 25 Apr 2023 17:01:43 | U                    | SMPL |      |
| CCV57 - 1        | CCV57            | 28812      | 4.9164  | -7471B   | PPB    | 25 Apr 2023 17:04:00 | U                    | SMPL |      |
| CCB57 - 1        | CCB57            | -97        | -0.0792 | -7471B   | PPB    | 25 Apr 2023 17:06:17 | U                    | SMPL |      |
| O2477-02 - 1     | MID-H4(ELEV-22)  | 2074       | 0.2960  | -7471B   | PPB    | 25 Apr 2023 17:08:34 | U                    | SMPL |      |
| O2477-03 - 1     | W-H4(ELEV-22)    | 4602       | 0.7328  | -7471B   | PPB    | 25 Apr 2023 17:10:50 | U                    | SMPL |      |
| O2483-01 - 1     | S2S-TP1          | 805        | 0.0767  | -7471B   | PPB    | 25 Apr 2023 17:22:41 | U                    | SMPL |      |
| O2483-03 - 1     | S2S-TP2          | 1274       | 0.1577  | -7471B   | PPB    | 25 Apr 2023 17:24:56 | U                    | SMPL |      |
| O2484-01 - 1     | CRCA-1           | 4534       | 0.7211  | -7471B   | PPB    | 25 Apr 2023 17:27:12 | U                    | SMPL |      |
| O2477-01LX5 - 1  |                  | 82         | -0.0483 | -7471B   | PPB    | 25 Apr 2023 17:29:29 | U                    | SMPL |      |
| O2477-01A - 1    |                  | 23448      | 3.9895  | -7471B   | PPB    | 25 Apr 2023 17:31:45 | U                    | SMPL |      |
| O2448-08DLX2 - 1 |                  | 38741      | 6.6322  | -7471B   | PPB    | 25 Apr 2023 17:34:02 | U                    | SMPL |      |
| CCV58 - 1        | CCV58            | 29060      | 4.9593  | -7471B   | PPB    | 25 Apr 2023 17:36:20 | U                    | SMPL |      |
| CCB58 - 1        | CCB58            | -72        | -0.0749 | -7471B   | PPB    | 25 Apr 2023 17:38:39 | U                    | SMPL |      |