



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB125158

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/26/2023	10:50
2	CAL2	1	Water	NA	NA	20.7	7.00	04/26/2023	10:51
3	CAL3	1	Water	NA	NA	20.3	10.02	04/26/2023	10:52
4	ICV	1	Water	NA	NA	20.8	6.99	04/26/2023	11:00
5	CCV1	1	Water	NA	NA	20.6	2.01	04/26/2023	11:01
6	O2485-01	1	Water	NA	NA	20.5	6.31	04/26/2023	11:05
7	O2485-02	1	Water	NA	NA	20.3	5.61	04/26/2023	11:06
8	O2485-03	1	Water	NA	NA	20.1	6.26	04/26/2023	11:10
9	O2495-01	1	Water	NA	NA	20.3	6.66	04/26/2023	11:15
10	O2495-03	1	Water	NA	NA	20.1	6.74	04/26/2023	11:20
11	O2495-05	1	Water	NA	NA	20.4	9.43	04/26/2023	11:25
12	O2495-07	1	Water	NA	NA	20.6	9.49	04/26/2023	11:30
13	O2495-09	1	Water	NA	NA	20.1	7.87	04/26/2023	11:32
14	O2496-01	1	Water	NA	NA	20.7	7.37	04/26/2023	11:35
15	O2496-01DUP	1	Water	NA	NA	20.8	7.38	04/26/2023	11:36
16	CCV2	1	Water	NA	NA	20.8	12.03	04/26/2023	11:40