


LB125159

245.1

INSTRUMENT ID: CV1

μ Abs.:

39695



Linear

A= 0.0000e+000

B= 2.5047e-004

C= -5.4993e-003

Rho= 0.9998757

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/A
0.0	0.000	0.012	0.012	70	0.000	70					0/1
0.05	0.050	0.046	-0.004	205	0.0 %	205	0				1/1
0.2	0.200	0.204	0.004	835	0.0 %	835					1/1
2.5	2.500	2.487	-0.013	9950	0.0 %	9950					1/1
5.0	5.000	4.939	-0.061	19740	0.0 %	19740					1/1
7.5	7.500	7.626	0.126	30469	0.0 %	30469					1/1
10.0	10.000	9.937	-0.063	39695	0.0 %	39695					1/1

LB125159 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 26 Apr 2023 11:19:28

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	59 50	70	-	0.0000	245.1	PPB	26 Apr 2023 11:24:50	S	Std
0.05 - 1	50.2 50.05	205	-	0.0500	245.1	PPB	26 Apr 2023 11:27:07	S	Std
0.2 - 1	52.5 50.2	835	-	0.2000	245.1	PPB	26 Apr 2023 11:29:25	S	Std
2.5 - 1	55 52.5	9950	-	2.5000	245.1	PPB	26 Apr 2023 11:31:42	S	Std
5.0 - 1	55 55	19740	-	5.0000	245.1	PPB	26 Apr 2023 11:33:59	S	Std
7.5 - 1	55 55	30469	-	7.5000	245.1	PPB	26 Apr 2023 11:36:15	S	Std
10.0 - 1	55 55	39695	-	10.0000	245.1	PPB	26 Apr 2023 11:38:33	S	Std
ICV45 - 1	ICV45	16067	4.0188	-245.1		PPB	26 Apr 2023 11:45:52	U	SMPL
ICB45 - 1	ICB45	-113	-0.0338	-245.1		PPB	26 Apr 2023 11:48:07	U	SMPL
CCV59 - 1	CCV59	20126	5.0355	-245.1		PPB	26 Apr 2023 11:50:24	U	SMPL
CCB59 - 1	CCB59	-122	-0.0361	-245.1		PPB	26 Apr 2023 11:52:39	U	SMPL
CRA - 1	CRA	187	0.0413	-245.1		PPB	26 Apr 2023 11:54:56	U	SMPL
HighStd - 1	HighStd	39590	9.9106	-245.1		PPB	26 Apr 2023 11:57:11	U	SMPL
ChkStd - 1	ChkStd	27927	6.9894	-245.1		PPB	26 Apr 2023 11:59:27	U	SMPL
PB152411BL - 1	PBW	30	0.0020	-245.1		PPB	26 Apr 2023 12:04:21	U	SMPL
PB152411BS - 1	LCSW	14286	3.5727	-245.1		PPB	26 Apr 2023 12:06:37	U	SMPL
O2213-10 - 1	LOD-MDL-WATER-04-QT2-2023	217	0.0489	-245.1		PPB	26 Apr 2023 12:16:44	U	SMPL
O2213-11 - 1	LOQ-WATER-05-QT2-2023	795	0.1936	-245.1		PPB	26 Apr 2023 12:19:00	U	SMPL
O2443-02 - 1	EFFLUENT-METALS	193	0.0428	-245.1		PPB	26 Apr 2023 12:21:18	U	SMPL
O2443-04 - 1	EFFLUENT-METALS	106	0.0211	-245.1		PPB	26 Apr 2023 12:23:35	U	SMPL
O2468-01 - 1	SUMP-PUMP-GW-SAMPLE	131	0.0273	-245.1		PPB	26 Apr 2023 12:25:51	U	SMPL
CCV60 - 1	CCV60	19497	4.8779	-245.1		PPB	26 Apr 2023 12:28:06	U	SMPL
CCB60 - 1	CCB60	-90	-0.0280	-245.1		PPB	26 Apr 2023 12:30:22	U	SMPL
O2469-01 - 1	TWP-02	244	0.0556	-245.1		PPB	26 Apr 2023 12:37:31	U	SMPL
O2469-01DUP - 1	TWP-02DUP	197	0.0438	-245.1		PPB	26 Apr 2023 12:39:46	U	SMPL
O2469-01MS - 1	TWP-02MS	14559	3.6411	-245.1		PPB	26 Apr 2023 12:42:01	U	SMPL
O2469-01MSD - 1	TWP-02MSD	14780	3.6965	-245.1		PPB	26 Apr 2023 12:44:17	U	SMPL
O2469-02 - 1	TWP-15	16	-0.0015	-245.1		PPB	26 Apr 2023 12:46:34	U	SMPL
O2469-03 - 1	TWP-24	873	0.2132	-245.1		PPB	26 Apr 2023 12:48:53	U	SMPL
O2469-04 - 1	TWP-02	53	0.0078	-245.1		PPB	26 Apr 2023 12:51:09	U	SMPL
O2469-05 - 1	TWP-15	115	0.0233	-245.1		PPB	26 Apr 2023 12:53:26	U	SMPL
O2469-06 - 1	TWP-24	153	0.0328	-245.1		PPB	26 Apr 2023 12:55:43	U	SMPL
CCV61 - 1	CCV61	19290	4.8261	-245.1		PPB	26 Apr 2023 12:57:59	U	SMPL
CCB61 - 1	CCB61	-63	-0.0213	-245.1		PPB	26 Apr 2023 13:00:15	U	SMPL
O2469-01LX5 - 1		10	-0.0030	-245.1		PPB	26 Apr 2023 13:02:32	U	SMPL
O2469-01A - 1		14846	3.7130	-245.1		PPB	26 Apr 2023 13:04:48	U	SMPL
CCV62 - 1	CCV62	19183	4.7993	-245.1		PPB	26 Apr 2023 13:07:03	U	SMPL
CCB62 - 1	CCB62	-114	-0.0341	-245.1		PPB	26 Apr 2023 13:09:21	U	SMPL