



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB125160

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	04/26/2023	08:40
2	O2485-01		1	14	100.00	>212.0	>212.0	04/26/2023	09:10
3	O2485-01DUP		1	13	100.00	>212.0	>212.0	04/26/2023	09:40
4	O2485-02		1	14	100.00	>212.0	>212.0	04/26/2023	10:10
5	O2485-03		1	12	100.00	>212.0	>212.0	04/26/2023	10:40
6	O2496-01		1	14	100.00	>212.0	>212.0	04/26/2023	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6/25/60

WorkList Name : fp-4-25 WorkList ID : 169404 Department : Wet-Chemistry Date : 04-25-2023 16:45:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/02/2023	Water	O2485-01	☒ Flash Point	Cool 4 deg C	PSEG03	M11	33992	04/25/2023	1010B
05/02/2023	Water	O2485-02	☒ Flash Point	Cool 4 deg C	PSEG03	M11	33994	04/25/2023	1010B
05/02/2023	Water	O2485-03	☒ Flash Point	Cool 4 deg C	PSEG03	M11	33993	04/25/2023	1010B
04/28/2023	Water	O2496-01	☒ Flash Point	Cool 4 deg C	PSEG03	M11	TOTE-4-25-2023	04/25/2023	1010B

Date/Time 04/26/2023 08:30
 Raw Sample Received by: RPT (GWT)
 Raw Sample Relinquished by: 1010B

Date/Time 04/26/2023 11:50
 Raw Sample Received by: 1010B
 Raw Sample Relinquished by: RPT Case