



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/27/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/26/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 04/27/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125163

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2497-01	34506	34	1.16	8.67	9.83	8.63	86.2	
O2497-02	32498	35	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O2500-01	PT-AN-SOIL	1	1.00	1.00	2.00	2.00	100.0	
O2500-02	PT-CORR-SOIL	2	1.00	1.00	2.00	2.00	100.0	
O2500-03	PT-CORR-SOIL	3	1.00	1.00	2.00	2.00	100.0	
O2500-04	PT-CN-SOIL	4	1.00	1.00	2.00	2.00	100.0	
O2500-05	PT-CN-SOIL	5	1.00	1.00	2.00	2.00	100.0	
O2500-06	PT-FP-SOIL	6	1.00	1.00	2.00	2.00	100.0	
O2500-07	PT-CR6-SOIL	7	1.00	1.00	2.00	2.00	100.0	
O2500-08	PT-NUT-SOIL	8	1.00	1.00	2.00	2.00	100.0	
O2500-09	PT-NUT-SOIL	9	1.00	1.00	2.00	2.00	100.0	
O2500-10	PT-OGR-SOIL	10	1.00	1.00	2.00	2.00	100.0	
O2500-11	PT-MET-SOIL	11	1.00	1.00	2.00	2.00	100.0	
O2500-12	PT-BNA-SOIL	12	1.00	1.00	2.00	2.00	100.0	
O2500-13	PT-TRIAZINE-SOIL	13	1.00	1.00	2.00	2.00	100.0	
O2500-14	PT-PAH-SOIL	14	1.00	1.00	2.00	2.00	100.0	
O2500-15	PT-DIES-SOIL	15	1.00	1.00	2.00	2.00	100.0	
O2500-16	PT-GAS-SOIL	16	1.00	1.00	2.00	2.00	100.0	
O2500-17	PT-NJEPH-SOIL	17	1.00	1.00	2.00	2.00	100.0	
O2500-18	PT-HERB-SOIL	18	1.00	1.00	2.00	2.00	100.0	
O2500-19	PT-PCB-SOIL	19	1.00	1.00	2.00	2.00	100.0	
O2500-20	PT-PCBO-SOIL	20	1.00	1.00	2.00	2.00	100.0	
O2500-21	PT-PEST-SOIL	21	1.00	1.00	2.00	2.00	100.0	
O2500-22	PT-CHLR-SOIL	22	1.00	1.00	2.00	2.00	100.0	
O2500-23	PT-TXP-SOIL	23	1.00	1.00	2.00	2.00	100.0	
O2500-24	PT-VOA-SOIL	24	1.00	1.00	2.00	2.00	100.0	
O2500-25	PT-SOL-SOIL	25	0.91	8.65	9.56	7.57	77.0	
O2501-01	RPV-ETGI-352	36	1.00	1.00	2.00	2.00	100.0	plastic sample



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/27/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/26/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 04/27/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125163

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2502-01	BU-511-COMP-01	44	1.16	8.72	9.88	9.19	92.1	
O2502-03	BU-511-VOC-01	45	1.16	8.72	9.88	9.19	92.1	
O2502-05	BU-511-01	46	1.16	8.71	9.87	9.29	93.3	
O2502-06	BU-511-02	47	1.13	8.43	9.56	8.87	91.8	
O2502-07	BU-511-03	48	1.12	8.84	9.96	9.18	91.2	
O2502-08	BU-511-COMP-02	49	1.14	8.84	9.98	9.22	91.4	
O2502-10	BU-511-VOC-02	50	1.14	8.84	9.98	9.22	91.4	
O2502-12	BU-511-04	51	1.14	8.82	9.96	9.44	94.1	
O2502-13	BU-511-05	52	1.15	8.81	9.96	9.4	93.6	
O2502-14	BU-511-06	53	1.14	8.57	9.71	9.16	93.6	
O2503-01	TR-305A-COMP-01	54	1.14	8.61	9.75	8.51	85.6	
O2503-03	TR-305A-VOC-01	55	1.15	8.83	9.98	8.51	83.4	
O2503-05	TR-305A-01	56	1.14	8.84	9.98	8.47	82.9	
O2503-06	TR-305A-02	57	1.14	8.83	9.97	8.69	85.5	
O2503-07	TR-305A-03	58	1.17	8.80	9.97	8.55	83.9	
O2503-08	TR-305A-COMP-02	59	1.16	8.70	9.86	8.53	84.7	
O2503-10	TR-305A-VOC-02	60	1.16	8.70	9.86	8.53	84.7	
O2503-12	TR-305A-04	61	1.16	8.53	9.69	8.2	82.5	
O2503-13	TR-305A-05	62	1.15	8.83	9.98	9.12	90.3	
O2503-14	TR-305A-06	63	1.14	8.50	9.64	8.68	88.7	
O2503-15	TR-305A-COMP-03	64	1.15	8.43	9.58	8.47	86.8	
O2503-17	TR-305A-VOC-03	65	1.15	8.43	9.58	8.47	86.8	
O2503-19	TR-305A-07	66	1.15	8.71	9.86	8.64	86.0	
O2503-20	TR-305A-08	67	1.13	8.83	9.96	8.38	82.1	
O2503-21	TR-305A-09	68	1.14	8.66	9.8	8.29	82.6	
O2504-01	BU-512-COMP-01	69	1.14	8.84	9.98	9.55	95.1	
O2504-03	BU-512-VOC-01	70	1.14	8.84	9.98	9.55	95.1	
O2504-05	BU-512-01	71	1.13	8.51	9.64	9.23	95.2	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/27/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/26/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 04/27/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB125163

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O2504-06	BU-512-02	72	1.14	8.56	9.7	8.9	90.7	
O2504-07	BU-512-03	73	1.15	8.43	9.58	8.94	92.4	
O2504-08	BU-512-COMP-02	74	1.15	8.49	9.64	9.22	95.1	
O2504-10	BU-512-VOC-02	75	1.15	8.49	9.64	9.22	95.1	
O2504-12	BU-512-04	76	1.17	8.51	9.68	8.76	89.2	
O2504-13	BU-512-05	77	1.12	8.42	9.54	9.07	94.4	
O2504-14	BU-512-06	78	1.13	8.77	9.9	9.4	94.3	
O2504-15	BU-512-COMP-03	79	1.15	8.69	9.84	8.73	87.2	
O2504-17	BU-512-VOC-03	80	1.15	8.69	9.84	8.73	87.2	
O2504-19	BU-512-07	81	1.14	8.62	9.76	9.22	93.7	
O2504-20	BU-512-08	82	1.15	8.80	9.95	9.58	95.8	
O2504-21	BU-512-09	83	1.15	8.83	9.98	9.5	94.6	
O2506-01	20230384-COMP-A-1	26	1.14	8.81	9.95	5.68	51.5	
O2506-02	20230385-COMP-A-5	27	1.15	8.81	9.96	5.57	50.2	
O2506-03	20230386-COMP-B-8	28	1.19	8.57	9.76	5.58	51.2	
O2506-04	20230387-COMP-B-14	29	1.17	8.78	9.95	5.43	48.5	
O2506-05	20230388-COMP-C-1	30	1.14	8.64	9.78	4.55	39.5	SLUDGE SAMPLE
O2506-06	20230389-COMP-C-8	31	1.15	8.78	9.93	3.38	25.4	SLUDGE SAMPLE
O2506-07	20230390-COMP-C-16	32	1.14	8.83	9.97	3.95	31.8	SLUDGE SAMPLE
O2506-08	20230391-COMP-C-24	33	1.12	8.50	9.62	4.22	36.5	SLUDGE SAMPLE
O2507-01	HD-01-042623	37	1.12	8.45	9.57	8.66	89.2	
O2507-02	HD-01-042623-EPH-2	38	1.12	8.45	9.57	8.66	89.2	
O2508-01	32514	39	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O2519-01	34323	40	1.18	8.37	9.55	9.4	98.2	
O2519-02	34324	41	1.17	8.43	9.6	9.36	97.2	
O2519-03	32323-24	42	1.13	8.62	9.75	9.55	97.7	
O2519-04	18827	43	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042623

WorkList ID : 169405

Department : Wet-Chemistry

Date : 04-26-2023 08:16:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/03/2023	Solid	O2497-01	Percent Solids	Cool 4 deg C	PSEG03	M11	34506	04/26/2023	Chemtech -SO
05/03/2023	Solid	O2497-02	Percent Solids	Cool 4 deg C	PSEG03	M11	32498	04/26/2023	Chemtech -SO
05/04/2023	Solid	O2500-01	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-AN-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-02	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-CORR-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-03	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-CORR-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-04	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-CN-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-05	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-CN-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-06	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-CN-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-07	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-FP-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-08	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-CR6-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-09	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-NUT-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-10	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-NUT-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-11	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-OGR-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-12	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-MET-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-13	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-BNA-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-14	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-TRIAZINE-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-15	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-PAH-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-16	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-DIES-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-17	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-GAS-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-18	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-NJEPH-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-19	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-HERB-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-19	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-PCB-SOIL	04/24/2023	Chemtech -SO

Date/Time 04/26/23 15:20

Raw Sample Received by: 70 (GRC)

Raw Sample Relinquished by: OR SM

Date/Time 04/26/23 17:10

Raw Sample Received by: CR SM

Raw Sample Relinquished by: 20 (GRC)

✓ 125163

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042623 WorkList ID : 169405 Department : Wet-Chemistry Date : 04-26-2023 08:16:53

Due Date		Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location		Customer Sample	Collect Date	Method
05/04/2023	Solid	O2500-20	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-PCBO-SOIL	QA Of	PT-PCBO-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-21	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-PEST-SOIL	QA Of	PT-PEST-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-22	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-CHLR-SOIL	QA Of	PT-CHLR-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-23	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-TXP-SOIL	QA Of	PT-TXP-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-24	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-VOA-SOIL	QA Of	PT-VOA-SOIL	04/24/2023	Chemtech -SO
05/04/2023	Solid	O2500-25	Percent Solids	Cool 4 deg C	CHEM02	QA Of	PT-SOL-SOIL	QA Of	PT-SOL-SOIL	04/24/2023	Chemtech -SO
05/02/2023	Solid	O2501-01	Percent Solids	Cool 4 deg C	PSEG03	M11	RPV-ETGI-352	M11	RPV-ETGI-352	04/26/2023	Chemtech -SO
05/01/2023	Solid	O2502-01	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-511-COMP-01	M11	BU-511-COMP-01	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2502-03	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-511-VOC-01	M11	BU-511-VOC-01	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2502-05	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-511-01	M11	BU-511-01	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2502-06	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-511-02	M11	BU-511-02	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2502-07	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-511-03	M11	BU-511-03	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2502-08	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-511-COMP-02	M11	BU-511-COMP-02	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2502-10	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-511-VOC-02	M11	BU-511-VOC-02	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2502-12	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-511-04	M11	BU-511-04	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2502-13	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-511-05	M11	BU-511-05	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2502-14	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-511-06	M11	BU-511-06	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-01	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-COMP-01	N11	TR-305A-COMP-01	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-03	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-VOC-01	N11	TR-305A-VOC-01	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-05	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-01	N11	TR-305A-01	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-06	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-02	N11	TR-305A-02	04/25/2023	Chemtech -SO

Date/Time 04/26/23 15:40 Date/Time 04/26/23 17:10
Raw Sample Received by: 70(WC) Raw Sample Received by: CP SM
Raw Sample Relinquished by: CP SM Raw Sample Relinquished by: 70(WC)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042623

WorkList ID : 169405

Department : Wet-Chemistry

Date : 04-26-2023 08:16:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/01/2023	Solid	O2503-07	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-03	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-08	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-COMP-02	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-10	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-VOC-02	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-12	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-04	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-13	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-05	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-14	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-06	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-15	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-COMP-03	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-17	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-VOC-03	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-19	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-07	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-20	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-08	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2503-21	Percent Solids	Cool 4 deg C	PSEG03	N11	TR-305A-09	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-01	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-COMP-01	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-03	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-VOC-01	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-05	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-01	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-06	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-02	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-07	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-03	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-08	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-COMP-02	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-10	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-VOC-02	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-12	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-04	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-13	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-05	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-14	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-06	04/25/2023	Chemtech -SO

Date/Time

04/26/23 15:40

Raw Sample Received by:

20(wc)

Raw Sample Received by:

af sm

Raw Sample Relinquished by:

Raw Sample Relinquished by:

Page 3 of 4

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042623

WorkList ID : 169405

Department : Wet-Chemistry

Date : 04-26-2023 08:16:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/01/2023	Solid	O2504-15	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-COMP-03	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-17	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-VOC-03	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-19	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-07	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-20	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-08	04/25/2023	Chemtech -SO
05/01/2023	Solid	O2504-21	Percent Solids	Cool 4 deg C	PSEG03	M11	BU-512-09	04/25/2023	Chemtech -SO
04/29/2023	Solid	O2506-01	Percent Solids	Cool 4 deg C	AQUA04	M11	20230384-COMP-A-1	04/19/2023	Chemtech -SO
04/29/2023	Solid	O2506-02	Percent Solids	Cool 4 deg C	AQUA04	M11	20230385-COMP-A-5	04/19/2023	Chemtech -SO
04/29/2023	Solid	O2506-03	Percent Solids	Cool 4 deg C	AQUA04	M11	20230386-COMP-B-8	04/19/2023	Chemtech -SO
04/29/2023	Solid	O2506-04	Percent Solids	Cool 4 deg C	AQUA04	M11	20230387-COMP-B-14	04/19/2023	Chemtech -SO
04/29/2023	Solid	O2506-05	Percent Solids	Cool 4 deg C	AQUA04	M11	20230388-COMP-C-1	04/19/2023	Chemtech -SO
04/29/2023	Solid	O2506-06	Percent Solids	Cool 4 deg C	AQUA04	M11	20230389-COMP-C-8	04/19/2023	Chemtech -SO
04/29/2023	Solid	O2506-07	Percent Solids	Cool 4 deg C	AQUA04	M11	20230390-COMP-C-16	04/19/2023	Chemtech -SO
04/29/2023	Solid	O2506-08	Percent Solids	Cool 4 deg C	AQUA04	M11	20230391-COMP-C-24	04/19/2023	Chemtech -SO
05/06/2023	Solid	O2507-01	Percent Solids	Cool 4 deg C	PSEG05	M11	HD-01-042623	04/26/2023	Chemtech -SO
05/06/2023	Solid	O2507-02	Percent Solids	Cool 4 deg C	PSEG05	M11	HD-01-042623-EPH-2	04/26/2023	Chemtech -SO
05/03/2023	Solid	O2508-01	Percent Solids	Cool 4 deg C	PSEG03	M11	32514	04/26/2023	Chemtech -SO
05/03/2023	Solid	O2519-01	Percent Solids	Cool 4 deg C	PSEG03	M11	34323	04/26/2023	Chemtech -SO
05/03/2023	Solid	O2519-02	Percent Solids	Cool 4 deg C	PSEG03	M11	34324	04/26/2023	Chemtech -SO
05/03/2023	Solid	O2519-03	Percent Solids	Cool 4 deg C	PSEG03	M11	32323-24	04/26/2023	Chemtech -SO
05/03/2023	Solid	O2519-04	Percent Solids	Cool 4 deg C	PSEG03	M11	18827	04/26/2023	Chemtech -SO

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: