

LB125164


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

46297



A= 0.0000e+000

B= 2.2089e-004

C= 1.0226e-002

Rho= 0.9988566

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.080	0.080	314	0.000	314	0				2
0.2	0.200	0.205	0.005	884	0.0 %	884					2
2.5	2.500	2.527	0.027	11395	0.0 %	11395					1
5.0	5.000	5.002	0.002	22599	0.0 %	22599					0
7.5	7.500	7.149	-0.351	32316	0.0 %	32316					5
10.0	10.000	10.237	0.237	46297	0.0 %	46297					2

LB125164 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Apr 2023 13:36:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	314	-	0.0000	7470A	PPB	26 Apr 2023 13:38:17	S	Std
0.2 - 1	50.2	884	-	0.2000	7470A	PPB	26 Apr 2023 13:42:35	S	Std
2.5 - 1	52.5	11395	-	2.5000	7470A	PPB	26 Apr 2023 13:44:52	S	Std
5.0 - 1	55	22599	-	5.0000	7470A	PPB	26 Apr 2023 13:47:09	S	Std
7.5 - 1	57.5	32316	-	7.5000	7470A	PPB	26 Apr 2023 13:49:26	S	Std
10.0 - 1	50	46297	-	10.0000	7470A	PPB	26 Apr 2023 13:54:19	S	Std
ICV46 - 1	ICV46	17674	3.9143	-	7470A	PPB	26 Apr 2023 13:59:49	U	SMPL
ICB46 - 1	ICB46	133	0.0396	-	7470A	PPB	26 Apr 2023 14:02:04	U	SMPL
CCV62 - 1	CCV62	22298	4.9357	-	7470A	PPB	26 Apr 2023 14:08:32	U	SMPL
CCB62 - 1	CCB62	130	0.0389	-	7470A	PPB	26 Apr 2023 14:10:47	U	SMPL
CRA - 1	CRA	921	0.2137	-	7470A	PPB	26 Apr 2023 14:15:43	U	SMPL
HighStd - 1	HighStd	45488	10.0582	-	7470A	PPB	26 Apr 2023 14:17:59	U	SMPL
ChkStd - 1	ChkStd	29634	6.5561	-	7470A	PPB	26 Apr 2023 14:20:14	U	SMPL
PB152423BL - 1	PBW	98	0.0319	-	7470A	PPB	26 Apr 2023 14:25:05	U	SMPL
PB152423BS - 1	LCSW	17430	3.8604	-	7470A	PPB	26 Apr 2023 14:27:21	U	SMPL
O2476-02 - 1	M-041923RB02	2319	0.5225	-	7470A	PPB	26 Apr 2023 14:29:37	U	SMPL
O2476-03 - 1	M-041723RB03	1744	0.3955	-	7470A	PPB	26 Apr 2023 14:31:55	U	SMPL
O2476-03DUP - 1	M-041723RB03DUP	1594	0.3623	-	7470A	PPB	26 Apr 2023 14:34:12	U	SMPL
O2476-03MS - 1	M-041723RB03MS	16162	3.5803	-	7470A	PPB	26 Apr 2023 14:41:21	U	SMPL
O2476-03MSD - 1	M-041723RB03MSD	16221	3.5933	-	7470A	PPB	26 Apr 2023 14:43:36	U	SMPL
CCV63 - 1	CCV63	22438	4.9666	-	7470A	PPB	26 Apr 2023 14:45:53	U	SMPL
CCB63 - 1	CCB63	142	0.0416	-	7470A	PPB	26 Apr 2023 14:48:08	U	SMPL
O2476-03LX5 - 1		459	0.1116	-	7470A	PPB	26 Apr 2023 14:54:01	U	SMPL
O2476-03A - 1		18898	4.1846	-	7470A	PPB	26 Apr 2023 14:56:15	U	SMPL
PB152390TB - 1	PB152390TB	27	0.0162	-	7470A	PPB	26 Apr 2023 15:01:02	U	SMPL
CCV64 - 1	CCV64	22796	5.0457	-	7470A	PPB	26 Apr 2023 15:03:17	U	SMPL
CCB64 - 1	CCB64	160	0.0456	-	7470A	PPB	26 Apr 2023 15:05:33	U	SMPL