

LB125167


SFAM01.1

INSTRUMENT ID: CV4

Linear

μ Abs.:

51552



A= 0.0000e+000

B= 1.9559e-004 slope

C= -2.3065e-002 y-intercept

Rho= 0.9998610

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.00	0.000	0.065	0.065	452	0.000	452					14
0.05	0.050					438	0				2
0.20	0.200	0.228	0.028	1284	0.0 %	1284					2
2.50	2.500	2.449	-0.051	12639	0.0 %	12639					2
5.00	5.000	4.893	-0.107	25135	0.0 %	25135					6
7.50	7.500	7.505	0.005	38487	0.0 %	38487					1
10.0	10.000	10.060	0.060	51552	0.0 %	51552					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	452	-		0 SFAM01.1	PPB	4/26/2023 16:04	Std
	0.2 S01	1284	-		0.2 SFAM01.1	PPB	4/26/2023 16:06	Std
	2.5 S02	12639	-		2.5 SFAM01.1	PPB	4/26/2023 16:09	Std
	5 S03	25135	-		5 SFAM01.1	PPB	4/26/2023 16:11	Std
	7.5 S04	38487	-		7.5 SFAM01.1	PPB	4/26/2023 16:13	Std
	10 S05	51552	-		10 SFAM01.1	PPB	4/26/2023 16:16	Std
ICV047	ICV047	21617	4.205 -		SFAM01.1	PPB	4/26/2023 16:18	SMPL
ICB047	ICB047	275	0.0307 -		SFAM01.1	PPB	4/26/2023 16:21	SMPL
CCV065	CCV065	24981	4.8629 -		SFAM01.1	PPB	4/26/2023 16:23	SMPL
CCB065	CCB065	274	0.0305 -		SFAM01.1	PPB	4/26/2023 16:25	SMPL
PB152429BL	PBW429	171	0.0104 -		SFAM01.1	PPB	4/26/2023 16:27	SMPL
O2249-01	EXEN2	549	0.0843 -		SFAM01.1	PPB	4/26/2023 16:30	SMPL
O2249-02	EXEN3	98	-0.0039 -		SFAM01.1	PPB	4/26/2023 16:32	SMPL
O2249-03	EXEN4	191	0.0143 -		SFAM01.1	PPB	4/26/2023 16:34	SMPL
O2249-04	EXEN4D	210	0.018 -		SFAM01.1	PPB	4/26/2023 16:37	SMPL
O2249-05	EXEN4S	4585	0.8737 -		SFAM01.1	PPB	4/26/2023 16:39	SMPL
PB152430BL	PBW430	194	0.0149 -		SFAM01.1	PPB	4/26/2023 16:41	SMPL
O2375-01	MCB9D1	485	0.0718 -		SFAM01.1	PPB	4/26/2023 16:43	SMPL
O2375-02	MCB9D3	458	0.0665 -		SFAM01.1	PPB	4/26/2023 16:46	SMPL
O2375-03	MCB9D7	181	0.0123 -		SFAM01.1	PPB	4/26/2023 16:48	SMPL
O2375-06	MCB9D3D	481	0.071 -		SFAM01.1	PPB	4/26/2023 16:50	SMPL
O2375-07	MCB9D3S	5772	1.1059 -		SFAM01.1	PPB	4/26/2023 16:52	SMPL
CCV066	CCV066	24636	4.7955 -		SFAM01.1	PPB	4/26/2023 16:55	SMPL
CCB066	CCB066	321	0.0397 -		SFAM01.1	PPB	4/26/2023 16:57	SMPL