



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB125170

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/26/2023	17:05
2	CAL2	1	Water	NA	NA	20.7	7.00	04/26/2023	17:06
3	CAL3	1	Water	NA	NA	20.7	10.03	04/26/2023	17:07
4	ICV	1	Water	NA	NA	20.8	6.99	04/26/2023	17:10
5	CCV1	1	Water	NA	NA	20.8	2.01	04/26/2023	17:15
6	O2497-04	1	Water	NA	NA	20.4	7.43	04/26/2023	17:20
7	O2497-04DUP	1	Water	NA	NA	20.5	7.44	04/26/2023	17:21
8	O2497-05	1	Water	NA	NA	20.9	5.73	04/26/2023	17:25
9	O2497-06	1	Water	NA	NA	20.3	7.06	04/26/2023	17:30
10	O2508-03	1	Water	NA	NA	20.7	5.42	04/26/2023	17:33
11	CCV2	1	Water	NA	NA	20.6	12.02	04/26/2023	17:35

WORKLIST(Hardcopy Internal Chain)

VB 125170

WorkList Name : ph o2508

WorkList ID : 169438

Department : Wet-Chemistry

Date : 04-26-2023 16:53:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/26/2023	Water	O2497-04	pH A	Cool 4 deg C	PSEG03	M11	34279	04/26/2023	9040C
04/26/2023	Water	O2497-05	pH A	Cool 4 deg C	PSEG03	M11	34558	04/26/2023	9040C
04/26/2023	Water	O2497-06	pH A	Cool 4 deg C	PSEG03	M11	34564	04/26/2023	9040C
04/26/2023	Water	O2508-03	pH A	Cool 4 deg C	PSEG03	M11	34333	04/26/2023	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

04/26/23 17:00
 JD(CDC)
 JD(CDC)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

04/26/23 18:50
 JD(CDC)
 JD(CDC)