



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB125171

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	04/26/2023	17:05
2	CAL2	1	Water	NA	NA	20.7	7.00	04/26/2023	17:06
3	CAL3	1	Water	NA	NA	20.7	10.03	04/26/2023	17:07
4	ICV	1	Water	NA	NA	20.8	6.99	04/26/2023	17:10
5	CCV1	1	Water	NA	NA	20.3	2.02	04/26/2023	17:40
6	O2497-01	1	Solid	20.02	20	20.8	6.82	04/26/2023	17:44
7	O2519-01	1	Solid	20.01	20	21.7	6.23	04/26/2023	17:45
8	O2519-02	1	Solid	20.03	20	21.9	6.12	04/26/2023	17:50
9	O2519-02DUP	1	Solid	20.04	20	21.8	6.13	04/26/2023	17:51
10	CCV2	1	Water	NA	NA	20.8	12.03	04/26/2023	18:00

WS 12517)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 02519

WorkList ID : 169439

Department : Wet-Chemistry

Date : 04-26-2023 16:55:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
04/26/2023	Solid	O2497-01	pH	✓	Cool 4 deg C	PSEG03 M11	34506	04/26/2023	9045D
04/26/2023	Solid	O2519-01	pH	✓	Cool 4 deg C	PSEG03 M11	34323	04/26/2023	9045D
04/26/2023	Solid	O2519-02	pH	✓	Cool 4 deg C	PSEG03 M11	34324	04/26/2023	9045D

Date/Time 04/26/23 17:00
Raw Sample Received by: 20(CWC)
Raw Sample Relinquished by: 20(CWC)

Date/Time 04/26/23 18:30
Raw Sample Received by: 20(CWC)
Raw Sample Relinquished by: 20(CWC)