

LB125178


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

49860



A= 0.0000e+000

B= 2.0126e-004 slope

C= -8.2644e-003 y-intercept

Rho= 0.9999637

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.007	-0.007	6	0.000	6					5
0.05	0.050	0.053	0.003	302	0.0 %	302	0				25
0.20	0.200	0.249	0.049	1279	0.0 %	1279					11
2.50	2.500	2.478	-0.022	12356	0.0 %	12356					11
5.00	5.000	4.942	-0.058	24599	0.0 %	24599					00
7.50	7.500	7.508	0.008	37347	0.0 %	37347					0
10.0	10.000	10.026	0.026	49860	0.0 %	49860					

LB125178
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	6	-	0	SFAM01.1	PPB	4/27/2023 10:41	Std
	0.05 S01	302	-	0.05	SFAM01.1	PPB	4/27/2023 10:44	Std
	0.2 S02	1279	-	0.2	SFAM01.1	PPB	4/27/2023 10:46	Std
	2.5 S03	12356	-	2.5	SFAM01.1	PPB	4/27/2023 10:48	Std
	5 S04	24599	-	5	SFAM01.1	PPB	4/27/2023 10:51	Std
	7.5 S05	37347	-	7.5	SFAM01.1	PPB	4/27/2023 10:54	Std
	10 S06	49860	-	10	SFAM01.1	PPB	4/27/2023 10:57	Std
ICV048	ICV048	19095	3.8347 -		SFAM01.1	PPB	4/27/2023 11:00	SMPL
ICB048	ICB048	45	0.0008 -		SFAM01.1	PPB	4/27/2023 11:04	SMPL
CCV067	CCV067	24817	4.9863 -		SFAM01.1	PPB	4/27/2023 11:07	SMPL
CCB067	CCB067	108	0.0135 -		SFAM01.1	PPB	4/27/2023 11:09	SMPL
PB152429BL	PBW429	22	-0.0038 -		SFAM01.1	PPB	4/27/2023 11:11	SMPL
O2249-01	EXEN2	265	0.0451 -		SFAM01.1	PPB	4/27/2023 11:13	SMPL
O2249-02	EXEN3	95	0.0109 -		SFAM01.1	PPB	4/27/2023 11:16	SMPL
O2249-03	EXEN4	161	0.0241 -		SFAM01.1	PPB	4/27/2023 11:18	SMPL
O2249-04	EXEN4D	194	0.0308 -		SFAM01.1	PPB	4/27/2023 11:20	SMPL
O2249-05	EXEN4S	4980	0.994 -		SFAM01.1	PPB	4/27/2023 11:23	SMPL
CCV068	CCV068	24738	4.9704 -		SFAM01.1	PPB	4/27/2023 11:25	SMPL
CCB068	CCB068	43	0.0004 -		SFAM01.1	PPB	4/27/2023 11:30	SMPL