

LB125181


7471B

INSTRUMENT ID: EV4

Linear

 μ Abs.:

60999



A= 0.0000e+000

B= 1.6545e-004

C= -1.0349e-001

Rho= 0.9997660

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.024	-0.024	481	0.000	481					
0.2	0.200	0.157	-0.043	1572	0.0 %	1572					-22
2.5	2.500	2.523	0.023	15872	0.0 %	15872					1
5.0	5.000	5.158	0.158	31798	0.0 %	31798					3
7.5	7.500	7.398	-0.102	45342	0.0 %	45342					-1
10.0	10.000	9.989	-0.011	60999	0.0 %	60999					0

LB125181 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Apr 2023 13:00:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	481	-	0.0000	7471B	PPB	27 Apr 2023 13:42:19	S	Std
0.2 - 1	50.2	1572	-	0.2000	7471B	PPB	27 Apr 2023 13:44:38	S	Std
2.5 - 1	52.5	15872	-	2.5000	7471B	PPB	27 Apr 2023 13:46:55	S	Std
5.0 - 1	55	31798	-	5.0000	7471B	PPB	27 Apr 2023 13:49:13	S	Std
7.5 - 1	57.5	45342	-	7.5000	7471B	PPB	27 Apr 2023 13:51:30	S	Std
10.0 - 1	510	60999	-	10.0000	7471B	PPB	27 Apr 2023 13:53:48	S	Std
ICV50 - 1	ICV50	25158	4.0589	-	7471B	PPB	27 Apr 2023 13:57:35	U	SMPL
ICB50 - 1	ICB50	483	-0.0226	-	7471B	PPB	27 Apr 2023 13:59:50	U	SMPL
CCV71 - 1	CCV71	31884	5.1717	-	7471B	PPB	27 Apr 2023 14:02:08	U	SMPL
CCB71 - 1	CCB71	484	-0.0234	-	7471B	PPB	27 Apr 2023 14:04:23	U	SMPL
CRA - 1	CRA	1524	0.1487	-	7471B	PPB	27 Apr 2023 14:06:41	U	SMPL
HighStd - 1	HighStd	60977	9.9852	-	7471B	PPB	27 Apr 2023 14:08:56	U	SMPL
ChkStd - 1	ChkStd	42330	6.9001	-	7471B	PPB	27 Apr 2023 14:11:12	U	SMPL
PB152431BL - 1	PBS	603	-0.0037	-	7471B	PPB	27 Apr 2023 14:13:30	U	SMPL
PB152431BS - 1	LCSS	26834	4.3362	-	7471B	PPB	27 Apr 2023 14:18:38	U	SMPL
O2490-01 - 1	VNJ-201	3510	0.4772	-	7471B	PPB	27 Apr 2023 14:20:54	U	SMPL
O2490-03 - 1	VNJ-208	3729	0.5135	-	7471B	PPB	27 Apr 2023 14:23:12	U	SMPL
O2491-01 - 1	S2S-TP3	1862	0.2046	-	7471B	PPB	27 Apr 2023 14:25:30	U	SMPL
CCV72 - 1	CCV72	31845	5.1653	-	7471B	PPB	27 Apr 2023 14:27:47	U	SMPL
CCB72 - 1	CCB72	450	-0.0290	-	7471B	PPB	27 Apr 2023 14:30:02	U	SMPL
O2491-03 - 1	S2S-TP4	2806	0.3608	-	7471B	PPB	27 Apr 2023 14:32:19	U	SMPL
O2491-03DUP - 1	S2S-TP4DUP	2999	0.3927	-	7471B	PPB	27 Apr 2023 14:34:35	U	SMPL
O2491-03MS - 1	S2S-TP4MS	24944	4.0235	-	7471B	PPB	27 Apr 2023 14:39:22	U	SMPL
O2491-03MSD - 1	S2S-TP4MSD	25283	4.0796	-	7471B	PPB	27 Apr 2023 14:41:37	U	SMPL
O2502-01 - 1	BU-511-COMP-01	2154	0.2529	-	7471B	PPB	27 Apr 2023 14:43:54	U	SMPL
O2502-08 - 1	BU-511-COMP-02	2457	0.3030	-	7471B	PPB	27 Apr 2023 14:46:11	U	SMPL
O2503-01 - 1	TR-305A-COMP-01	18259	2.9175	-	7471B	PPB	27 Apr 2023 14:48:27	U	SMPL
O2503-08 - 1	TR-305A-COMP-02	8933	1.3745	-	7471B	PPB	27 Apr 2023 14:50:44	U	SMPL
O2503-15 - 1	TR-305A-COMP-03	11981	1.8788	-	7471B	PPB	27 Apr 2023 14:53:02	U	SMPL
O2504-01 - 1	BU-512-COMP-01	1290	0.1099	-	7471B	PPB	27 Apr 2023 14:55:19	U	SMPL
CCV73 - 1	CCV73	31882	5.1714	-	7471B	PPB	27 Apr 2023 14:57:38	U	SMPL
CCB73 - 1	CCB73	551	-0.0123	-	7471B	PPB	27 Apr 2023 14:59:54	U	SMPL
O2504-08 - 1	BU-512-COMP-02	2270	0.2721	-	7471B	PPB	27 Apr 2023 15:02:11	U	SMPL
O2504-15 - 1	BU-512-COMP-03	1753	0.1865	-	7471B	PPB	27 Apr 2023 15:04:27	U	SMPL
O2507-01 - 1	HD-01-042623	2338	0.2833	-	7471B	PPB	27 Apr 2023 15:06:42	U	SMPL
O2491-03LX5 - 1		761	0.0224	-	7471B	PPB	27 Apr 2023 15:12:12	U	SMPL
O2491-03A - 1		26941	4.3539	-	7471B	PPB	27 Apr 2023 15:14:28	U	SMPL
CCV74 - 1	CCV74	32235	5.2298	-	7471B	PPB	27 Apr 2023 15:16:44	U	SMPL
CCB74 - 1	CCB74	505	-0.0199	-	7471B	PPB	27 Apr 2023 15:19:01	U	SMPL